

June 7, 1984

LICENSE NO. 44-18427-01
DOCKET NO. 030-15050
CONTROL NO. 17405

Saint Michael's College
Chemistry Department
ATTN: Michael H. Gianni
Acting Chairman
Winooski, Vermont 05404

Gentlemen:

This letter is in response to your letter dated April 10, 1984, for renewal of license number 44-18427-01. Additional information is needed to support your application. To assist you in providing this information, we have enclosed a copy of the guide, "A guide for Preparation of Byproduct Material Applications for the Use of Sealed Sources in Gas Chromatography Devices and X-Ray Fluorescence Analyzers".

If you wish to return the detector cell(s) to the supplier for cleaning and/or exchange, you need only to state this, but provide procedures for removal and exchange of the foil and/or plated source. If you wish to perform in-house cleaning or servicing, the manufacturer's recommended procedures should be followed and a copy of the procedures submitted with the application.

We will continue our review of your application upon receipt of this information. Please reply in duplicate and refer to Control No. 17405. Should you have any questions, please call us at (215) 337-5260.

Sincerely,

Original Signed By:
John E. Glenn

John E. Glenn, Ph.D., Chief
Nuclear Materials Section B
Division of Engineering and
Technical Programs

Technical Programs	
OFFICE	
SURNAME	8408130112 840727
DATE	NMS LIC30 44-18427-01 PDR

ML10

Enclosures:

1. A guide for the Preparation of Byproduct Material Applications for the Use of Sealed Sources in Gas Chromatography Devices and X-Ray Fluorescence Analyzers.

OFFICE						
REGION						
SUBMITTER						
DATE	06/7/84	JJ Miller	06/7/84			