

The Schiffler Cancer Center Gregory S. Merrick, M.D., Director Phone: (304) 243-3490 Fax: (304) 243-5047 / (304) 243-7110

SchifflerCancerCenter@wheelinghospital.org

2 April 2020

Wheeling Office 1 Medical Park Wheeling, WV 26003 (304) 243-3490

Medical Staff

Gregory S. Merrick, M.D. Jondavid Pollock, M.D., Ph.D.

Medical Physics

Wayne M. Butler, Ph.D. Amber L. Hines, M.S., CMD Jonathan H. Lief, Ph.D. Robert W. Galbreath, Ph.D. Brian S. Kurko, M.S., CMD Richard L. Anderson, RTT, CMD Ernest G. Butler, M.S., CMD Abbey Bennett, B.S. Ryan M. Fiano, M.P.H.

<u>Administrative Coordinator</u> Michele R. Braley

<u>Physical Rehabilitation</u> John DeBlasis, M.S., PT, ATC

<u>SELECT Research Nurse</u> Jayme Nardo, R.N. U.S. NRC Region I 2100 Renaissance Blvd., Suite 100 King of Prussia, PA 19406-2713

Re: Amendment Request to License No. 47-05322-02

Dear US NRC:

This letter is in response to your email dated 2 Apr 2020 requesting additional information regarding our license amendment request.

- 1. Although the formal training for Mr. Kurko and Mr. Butler was not recent, I thought the application should acknowledge and document their formal training. Since that information is unnecessary and counterproductive, I would like to use the pathway of 10 CFR 35.50(c) to support their applications as ARSOs.
- 2. Wheeling Hospital's RSO, Wayne Butler, PhD, maintains responsibility for the radiation safety program and may delegate duties, but not the responsibility, to Mr. Kurko and Mr. Butler.
- 3. Please authorize Mr. Kurko and Mr. Butler for calibrations with the Cesium-137 sources listed in in item 6. F. of NRC License No. 47-05322-02. These sources provide a reference value for ion chambers and well chambers before and after their round trip to a calibration laboratory. The reference measurement does not calibrate the device but simply ensures that the device was not damaged in transit. A summary of the training for Brian Kurko and Ernest Butler is below:

Description of Training	Training Provided by	Dates of Training
Use of radiation from Cs-137 tubes to check ion chamber response	Wayne Butler, PhD	Jan, 2019 — Feb, 2020

4. The email address of our administrator, David Lucey, is JDLucey@wheelinghospital.org

Best regards,

Buth Wayne M. Butler. Ph.D