

November 7, 1995

Fernando G. Salazar, M.D.
P. O. Box 1169
Mayaguez, Puerto Rico 00709

SUBJECT: NOTICE OF VIOLATION
(NRC INSPECTION REPORT NO. 52-21149-01/95-01)

Dear Dr. Salazar:

This refers to the inspection conducted by José M. Díaz Vélez of this office on October 19, 1995. The inspection included a review of activities authorized for your Mayagüez, Puerto Rico facility. At the conclusion of the inspection, the findings were discussed with you.

The inspection was an examination of activities conducted under your license with respect to radiation safety and compliance with NRC regulations and the conditions of your license. It included selective examinations of procedures and representative records, interviews with personnel, and direct observations by the inspector.

Based on the results of this inspection, certain of your activities appeared to be in violation of NRC requirements, as specified in the enclosed Notice of Violation (Notice). The violation is of concern because it involves the failure to secure licensed material from unauthorized removal or access, or to maintain the licensed material under constant surveillance and immediate control. The NRC expects the licensee to recognize the potential hazards of storing radioactive material in a manner such that it could result in the unauthorized removal or access to the material.

You are required to respond to this letter and should follow the instructions specified in the enclosed Notice when preparing your response. In your response, you should document the specific actions taken and any additional actions you plan to prevent recurrence. After reviewing your response to this Notice, including your proposed corrective actions and the results of future inspections, the NRC will determine whether further NRC enforcement action is necessary to ensure compliance with NRC regulatory requirements.

In accordance with 10 CFR 2.790 of the NRC's "Rules of Practice," a copy of this letter, its enclosure, and any reply will be placed in the NRC Public Document Room.

The responses directed by this letter and the enclosed Notice are not subject to the clearance procedures of the Office of Management and Budget as required by the Paperwork Reduction Act of 1980, Pub. L. No. 96-511.

150059
9511150380 951107
PDR ADOCK 03019845
C PDR

7E07

Fernando G. Salazar, M.D.

2

Should you have any questions concerning this letter, please contact us.

Sincerely,

(original signed by
J. P. Potter)



John P. Potter, Chief
Materials Licensing/Inspection
Branch 2
Division of Nuclear Materials Safety

Docket No. 030-19845
License No. 52-21149-01

Enclosure: Notice of Violation

cc w/encl: Commonwealth of Puerto Rico

Distribution w/encl:
PUBLIC
RII Docket File, DNMS

SEND TO PUBLIC DOCUMENT ROOM?		YES	NO
OFFICE	RII-DNMS	RII-DNMS	
SIGNATURE			
NAME	J. Diaz Velez	J. Potter	
DATE	11 / 6 / 95	11 / 7 / 95	11 / / 95 11 / / 95 11 / / 95 11 / / 95
COPY?	YES NO	YES NO	YES NO YES NO YES NO YES NO

OFFICIAL RECORD COPY

DOCUMENT NAME: G:\DRSS\NMIS\SALAZAR.NOV