

*file*

APR 20 1984

Family Foot Care  
ATTN: Gary Ascher  
Administrator  
16611 Chagrin Blvd.  
Shaker Heights, OH 44120

Gentlemen:

We are in receipt of your application dated March 5, 1984 for an NRC License to use iodine-125 in a Lixiscope. Additional information is needed, however, in support of your request. Please respond to the following:

1. Your application indicates that Drs. Aronoff, Cohen, Goss, Khouri, Warshaw, Kodish, Moore and Gall received instructions from James A. Gilmartin on the operation of the Lixiscope and associated radiation safety procedures. We do not have in our files, however, documentation of the additional training and experience that Mr. Gilmartin has received which qualifies him to provide training to potential users of the Lixiscope. We will need an outline of the additional training received, the duration of the training, a description of Mr. Gilmartin's experience using and instructing others on the use of the Lixiscope and a statement from his instructor certifying Mr. Gilmartin's qualifications to train potential users and describing the method used to determine his competency. Please contact Mr. Gilmartin and submit this information.
2. Your application does not specify the location, dates and duration of the didactic training that Dr. Frank Gall has received in Radiation Physics and Instrumentation and Radiation Biology. Please submit this information.
3. If you are requesting authorization to possess and use more than one Lixiscope, describe the procedures that you will implement to maintain accountability of all iodine-125 sealed sources received by Family Foot Care Clinics and transferred to Lixi, Inc. for disposal.
4. What provisions have you made for obtaining access to a properly calibrated survey meter in the event of suspected source leakage.
5. Please outline the duties and responsibilities of your Radiation Protection Officer.

We will continue the review of your application upon receipt of this information. Please reply in duplicate within 30 days and reference Control No. 76535.

Sincerely,

B408210597 B40716  
NMS LIC30  
34-24310-01 PDR

B. J. Holt  
Materials Licensing Section

RIII *BJH*  
Holt/bm  
04/17/84 4/19/84