

ENCLOSURE 1

NOTICE OF DEVIATION

South Carolina Electric and Gas Company
V. C. Summer

Docket No. 50-395
License No. NPF-12

The following deviation was identified during an inspection conducted on May 1-31, 1984.

Final Safety Analysis Report (FSAR) Section 12.1.4.2 states that the Area Radiation Monitors are calibrated on a routine basis. This constitutes a commitment to the NRC regarding routine calibration of the Area Radiation Monitors.

Contrary to this commitment, as of May 24, 1984, the Area Radiation Monitors had not been included in the Licensee's Preventative Maintenance Program for routine calibration and as a consequence, the following Area Radiation Monitor calibrations had expired.

<u>Monitor</u>	<u>Area Served</u>
RMG-2	Radiochemistry Lab Area
RMG-3	Sampling Room Area
RMG-12	Auxiliary Building Waste Hold-up Tank Area

Please provide, in writing within 30 days of the date of this Notice, a description of corrective actions regarding these deviations, actions taken to avoid further deviations, and the dates when these actions were or will be completed.

Security or safeguards information should be submitted as an enclosure to facilitate withholding it from public disclosure as required by 10 CFR 2.790(d) or 10 CFR 73.21.

Date: JUN 29 1984