

TABLE II

U S NUCLEAR REGULATORY COMMISSION
 OFFICE OF INSPECTION AND ENFORCEMENT
 CONFIRMATORY MEASUREMENTS PROGRAM
 FACILITY: D.C. COOK
 FOR THE 4 QUARTER OF 1983

SAMPLE	ISOTOPE	-----NRC-----		----LICENSEE----		---LICENSEE:NRC----		
		RESULT	ERROR	RESULT	ERROR	RATIO	RES	T
L WASTE	H-3	2.0E-03	2.0E-05	2.3E-03	9.6E-04	1.1E 00	1.0E 02	A
	SR-89	9.0E-08	7.2E-08	1.5E-07	1.8E-08	1.6E 00	1.3E 00	N
	SR-90	5.1E-08	3.1E-08	2.6E-08	3.7E-09	5.1E-01	1.7E 00	N
	BETA	4.5E-04	2.0E-05	4.5E-04	1.4E-04	1.0E 00	2.3E 01	A

T TEST RESULTS:

- A=AGREEMENT
- D=DISAGREEMENT
- P=POSSIBLE AGREEMENT
- N=NO COMPARISON