

## UNITED STATES NUCLEAR REGULATORY COMMISSION

REGION I 2100 RENAISSANCE BLVD. KING OF PRUSSIA, PA 19406-2713

April 3, 2020

David Giuffrida, P.E. Vice President, Operations Middlesex Health 28 Crescent Street Middletown, CT 06457-3650

SUBJECT: MIDDLESEX HEALTH, AMENDMENT NOT NEEDED, MAIL CONTROL NO.

617937

Dear Mr. Giuffrida:

This is in reference to your letter dated January 9, 2020, requesting the use of Ga-68 for imaging and localization studies permitted by 10 CFR 35.200. After a telephone discussion with Jason Marsden, your Radiation Safety Officer, on March 11, 2020, it is our understanding that you will be obtaining unit dosages of Ga-68 from a licensed radiopharmacy. An amendment is only required if you would be receiving a Ge-68/Ga-68 generator, which is regulated under 10 CFR 35.1000. Unit dosages of Ga-68 are regulated by 10 CFR 35.200, which you are already licensed for. Therefore, an amendment is not required.

Thank you for your cooperation.

Sincerely,

Janice Nguyen, Senior Health Physicist Medical and Licensing Assistance Branch Division of Nuclear Materials Safety Region I

License No. 06-00649-03 Docket No. 030-01242 Mail Control No. 617937

cc: Jason Marsden, Radiation Safety Officer

MIDDLESEX HEALTH, AMENDMENT NOT NEEDED, MAIL CONTROL NO. 617937 DATED APRIL 3, 2020

DOCUMENT NAME: G:\WBL Documents\WBL License Cover Letter\L06-00649-03.617937.docx **SUNSI Review Complete: JNguyen** 

After declaring this document "An Official Agency Record" it <u>will</u> be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	RI:DNMS	N			
NAME	JNguyen (JEN)				
DATE	4/3/2020				

OFFICIAL RECORD COPY