

1 Q.1 Please state your name, address, present occupation
and employer.

2 A.1 My name is Stephen A. Browne. My business address
3 is: Shearon Harris Energy & Environmental Center, Route 1, Box
4 327, New Hill, North Carolina 27562. I am employed by Carolina
5 Power & Light Company ("CP&L") as Project Specialist - Health
6 Physics. In this position I am responsible for the technical
7 direction of the personnel dosimetry program for all CP&L nu-
8 clear plants.

9 Q.2 State your educational background and professional
10 work experience.

11 A.2 I have a Bachelor of Science degree in Physics from
12 Union College. I received a Masters of Science degree in Envi-
13 ronmental Health Engineering from Northwestern University. A
14 summary of my professional experience and qualifications is
15 contained in Attachment A to this testimony. For the last
16 eight years I have been directly involved with the supervision
17 and direction of dosimetry programs using thermoluminescent
18 dosimeter ("TLD") systems manufactured by Harshaw, Teledyne and
19 Panasonic, the major manufacturers of TLD systems used in the
20 United States today. Recently I have been a consultant to the
21 National Bureau of Standards ("NBS") as a technical expert in
22 assessing and evaluating personnel radiation dosimetry proces-
23 sors under the National Voluntary Laboratory Accreditation Pro-
24 gram ("NVLAP").

1 Q.3 Describe the major responsibilities of your position
2 at CP&L.

3 A.3 In my current position, I am responsible for techni-
4 cal and quality assurance management of the CP&L dosimetry pro-
5 gram. The scope of these responsibilities includes supervision
6 of the following types of activities: preparation and pro-
7 cessing of TLD badges, data processing, establishment of meth-
8 ods and procedures for all phases of dosimetry operations,
9 calibration and maintenance of dosimetry equipment, establish-
10 ment of quality assurance standards, performance of quality
11 control inspections and checks, training and qualification of
12 dosimetry staff, performance of tests, development of new meth-
13 ods and systems, and maintenance of official personnel exposure
14 history records.

15 Q.4 What is the purpose of your testimony?

16 A.4 The purpose of my testimony is to address the re-
17 maining issue of Joint Contention IV which can be stated as
18 follows: Can the TLDs and measuring equipment and processes to
19 be used at the Harris Plant measure occupational radiation
20 doses with sufficient accuracy to comply with NRC regulations?
(Tr. 2218, August 10, 1984 Conference Call).

21 My testimony demonstrates that the TLDs to be used at the
22 Harris Plant and CP&L's processing techniques are sufficiently
23 accurate to comply with the American National Standards Insti-
24 tute ("ANSI") standard (which has been proposed as an NRC rule
25 on TLD accuracy), the International Commission on Radiological

1 Protection ("ICRP") standard, as well as the interpretation of
2 the current regulatory standard suggested by the Board. There-
3 fore, there is no merit to the remaining issue in Joint
4 Contention IV.

5 Q.5 How is your testimony organized?

6 A.5 First I will describe the manner in which TLDs are to
7 be used and processed at the Harris Plant. I then will explain
8 the standards for TLD accuracy which have been proposed by or-
9 ganizations involved with radiation protection and which have
10 been suggested by the Board and demonstrate that CP&L's program
11 is consistent with all such standards. Finally, I will explain
12 CP&L's program for maintaining the accuracy of its TLD pro-
13 cessing system and mitigating the most frequently observed
14 causes of inaccuracy.

15 Q.6 What are TLDs?

16 A.6 "TLD" stands for thermoluminescent dosimeter, which
17 is a device used for measuring exposure to radiation. When a
18 TLD is irradiated by ionizing radiation, some energy is ab-
19 sorbed and stored. If the TLD subsequently is heated, some of
20 the stored energy is released as light which can be detected
21 and measured. The quantity of light released is proportional
22 to the dose received by the individual wearing the TLD.

23 Q.7 What type of TLDs will be used at the Harris Plant?

24 A.7 We plan to use Model UD-802 AQ TLDs manufactured by
25 Panasonic. The manufacturer's technical specifications of
those TLDs are set forth in Applicants' Exhibit _____.

Q.8 How will TLDs be used at the Harris Plant?

1 A.8 TLDs will be used to perform routine monitoring of
2 personnel. The objective of routine monitoring is to assess
3 the cumulative dose to individuals for official exposure
4 record-keeping purposes.

5 Q.9 Why are TLDs the chosen method for routine moni-
6 toring?

7 A.9 TLDs are nearly ideal for routine monitoring because
8 they are rugged, reliable, accurate, and sensitive. TLDs are
9 capable of measuring the dose from the types and energies of
10 radiation which represent significant external exposure hazards
11 in nuclear power plants. For instance, for beta radiation, the
12 TLDs proposed for the Harris Plant are capable of measuring
13 dose over the energy range from about 0.1 to 2.3 MeV. Betas
14 below 0.1 MeV are too weak to be a significant external hazard,
15 while betas above 2.3 MeV are very rare. For gamma radiation,
16 the TLDs have a usable energy range from about 40 keV to 7 Mev.
17 Gammas above this range are rare and gammas below this range
18 contribute relatively little to the total dose in nuclear power
19 plants. With appropriate calibration, the TLD also can be used
20 for neutron monitoring in the event that such monitoring is
21 necessary.

22 Q.10 When are TLDs worn?

23 A.10 TLDs are worn continuously by individuals while work-
24 ing in the radiologically controlled areas of nuclear power
25 plants.

Q.11 How frequently are TLDs processed?

1 A.11 TLDs are processed to obtain official dose readings
2 at least monthly. Some TLDs are processed more frequently for
3 exposure control purposes.

4 Q.12 Are there relevant standards for the accuracy re-
5 quired in routine monitoring?

6 A.12 Yes. The International Commission on Radiation Units
7 and Measurements ("ICRU"), the ICRP, and the National Council
8 for Radiation Protection and Measurements ("NCRP") have
9 addressed the issue of accuracy for individual monitoring and
10 published recommendations in ICRU Report 20, ICRP Report 35
11 (which superceded ICRP Report 12) and NCRP Report 57. These
12 organizations are considered to be authorities in the field of
13 radiation protection and measurements, and their recommenda-
14 tions are the basis for radiation protection practice and reg-
15 ulations in countries throughout the world.

16 Q.13 What accuracy level do these institutions recommend?

17 A.13 NCRP Report 57 (1978) states:

18 The measurement accuracy desirable in person-
19 nel dosimetry depends on the radiation level to be
20 measured. At the level of the MPD [Maximum Per-
21 missible Dose] a measurement accuracy of ± 30 per-
22 cent should be achieved. If the dose equivalent
to critical organs is less than 1/4 of the MPD,
personnel monitoring is not required and a lower
level of accuracy (e.g., a factor of 2) is accept-
able.

23 "Instrumentation and Monitoring Methods for Radiation Protec-
24 tion," NCRP Report 57 (1978) at 63.

1 ICRU Report 20 (1971) contains the following language con-
cerning accuracy:

2 It is suggested that when the MADE [Maximum
3 Dose Equivalent] is comparable to the maximum per-
missible dose, an accuracy of $\pm 30\%$ be achieved.
4 When the MADE is considerably less than the MPD,
less accuracy is acceptable (e.g., at a level
5 equal to 0.1 of the MPD an uncertainty of as much
as a factor of three seems acceptable).

6 "Radiation Protection Instrumentation and Its Application,"
7 ICRP Report 20 (1971) at 7.

8 ICRP Publication 12 (1968), which was referenced by the
9 NRC Staff ("Affidavit of Seymour Block in Support of Summary
10 Disposition of Joint Contention IV," January 25, 1984, at 2)
11 and by the Board ("Memorandum and Order (Ruling on Motions for
12 Summary Disposition") April 13, 1984, at 11-12) (hereinafter
13 "April 13 Memorandum") for their interpretation of the accuracy
14 requirements, has been superceded officially by ICRP
15 Publication 35 (1982), which states the following concerning
16 the measurement of shallow and deep dose equivalent:

17 If these quantities are of the order of the
18 relevant annual limits, the uncertainties should
not exceed a factor of 1.5 at the 95% confidence
19 level. Where they amount to less than 10 mSv [1
rem] an uncertainty of a factor of 2 at the 95%
20 confidence level is acceptable.

21 "General Principles of Monitoring for Radiation Protection of
22 Workers," ICRP Publication 35 at 25. In general, all of these
23 organizations recommend accuracy standards in the range of ± 30
24 to 50% at dose levels approximating the annual dose limit of 5
25 rem. At lower dose levels (1 rem), uncertainties on the order

1 of 100% or more are considered acceptable. The standards do
2 not even propose that individual monitoring be required at dose
3 levels less than 25 to 30% of the annual dose limits.

4 Although the above standards are in general agreement on
5 the subject of accuracy, they are not specific enough to be
6 used as the basis for evaluating the performance of dosimetry
7 systems. For this reason, ANSI N13.11-1983 was established to
8 provide specific criteria for testing the performance of
9 dosimetry processors.

10 Q.14 What is the ANSI standard?

11 A.14 ANSI N13.11 (1983) is the current standard used for
12 testing the performance of dosimetry processors under the NBS
13 dosimetry accreditation program, and as such it is the most
14 widely used standard for the accuracy of dosimetry in the nu-
15 clear industry.

16 The ANSI standard sets the tolerance level at 50% for
17 doses from 0.03 to 10 rem and at 30% for doses from 10 to 500
18 rem. The ANSI performance criterion specifies that for a se-
19 ries of dosimeter measurements the sum of the average absolute
20 bias (P) plus the standard deviation (S) must be less than the
21 specified tolerance level (L). The mathematical expression of
22 the criterion is:

$$P + S < L$$

23 The average absolute bias is a measure of the deviation of the
24 average measured dose from the true dose, while the standard
25 deviation is a measure of the variation or spread of the

individual dosimeter measurements about the average measured
1 dose.

2 Q.15 Do the TLDs to be used at the Harris Plant satisfy
3 the ANSI standard?

4 A.15 Yes. The Panasonic TLDs to be used at the Harris
5 Plant were tested in 1982 and 1984 and found to meet the per-
6 formance specifications of ANSI N13.11-1983. In 1982, testing
7 was conducted at the University of Michigan by Dr. Phil Plato
8 as part of a study sponsored by the NRC. During the study the
9 TLDs were irradiated to a variety of radiation sources whose
10 calibrations were verified by NBS. The methods used during
11 this study are documented in Performance Testing of Personnel
12 Dosimetry Services - Revised Procedures Manual, NUREG/CR-2892
13 (February 1983), and the results of testing are documented in
14 Performance Testing of Personnel Dosimetry Study - Final Report
15 Test 3, NUREG/CR-2891, page no. 18 (February 1983)
16 (Attachment B hereto). In these reports, CP&L is listed as
17 processor number 187. In 1984, similar testing was conducted
18 to meet the NVLAP accreditation performance test requirements.
19 Again, the testing was conducted at the University of Michigan,
20 which was selected by NBS as the official testing lab for the
21 NVLAP. The results of this testing have been forwarded to NBS
22 and CP&L, but have not been published.

23 As the table shows, using Panasonic TLD's, CP&L achieved
24 the following performance for categories I through VIII in
25 these two tests:

1	<u>Category</u>	<u>Radiation Type</u>	<u>1982 CP&L</u> <u>Performance</u> <u>(P+S)</u>	<u>1984 CP&L</u> <u>Performance</u> <u>(P+S)</u>	<u>ANSI</u> <u>Limit</u>
2					
3	I	X-ray Accident	.24	.18	.3
4	II	Gamma Accident	.10	.15	.3
5	III	X-ray Shallow	.11	.18	.5
6		X-ray Deep	.12	.16	.5
7	IV	Gamma	.06	.10	.5
8	V	Beta	.30	.28	.5
9	VI	Gamma & X-ray			
10		Shallow	.06	.19	.5
11		Gamma & X-ray Deep	.16	.18	.5
12	VII	Gamma & Beta			
13		Shallow	.16	.29	.5
14		Gamma & Beta Deep	.11	.10	.5
15	VIII	Gamma & Neutron	*	.09	.5

*CP&L did not participate in this test category in 1982.

The tolerance limit is 0.3 for Categories I and II and 0.5 for all other categories. Thus CP&L passed the tolerance limit for all categories in which it participated during each test. These tests demonstrate that the Panasonic TLDs proposed for the Harris Plant meet the accuracy requirements which have been endorsed by the national consensus standard of ANSI.

Q.16 As the Board noted, in the development of the ANSI standard the criterion was relaxed by changing the passing formula from $P+2S < L$ to $P+S < L$. How would the performance of the TLDs to be used at the Harris Plant compare to the more restrictive criterion?

A.16 First, it should be noted that while the criterion was relaxed from 2S to S, the maximum value for L was tightened from 2.0 to 0.5 for low doses. However, the following table shows CP&L results compared to the criterion of $P+2S < L$. As

can be seen, CP&L passed the more restrictive criterion for all categories in which it participated during each test except the accident x-ray category in 1982. In my opinion, it is not realistic to expect that an individual could receive accident level exposures to x-rays in a nuclear power plant.

Category	Radiation Type	1982 CP&L Performance (P+2S)	1984 CP&L Performance (P+2S)	Limit
I	X-ray, Accident	.37	.29	.3
II	Gamma, Accident	.14	.21	.3
III	X-ray Shallow	.16	.26	.5
	X-ray Deep	.22	.25	.5
IV	Gamma	.09	.17	.5
V	Beta	.36	.37	.5
VI	Gamma & X-ray Shallow	.12	.26	.5
	Gamma & X-ray Deep	.23	.28	.5
VII	Gamma & Beta Shallow	.22	.41	.5
	Gamma & Beta Deep	.17	.18	.5
VIII	Gamma & Neutron Deep	*	.15	.5

*CP&L did not participate in this category in 1982.

Q.17 is the ANSI standard compatible with other nationally recognized standards?

A.17 In general, yes. Only ICRP 35, which defines the acceptable uncertainty in routine monitoring at a specific confidence level can be effectively compared to the ANSI standard. For doses on the order of the annual limit (5 rem), ICRP 35 states that the uncertainties should not exceed 50% at the 95% confidence level. This can be expressed in mathematical terms (in an equation similar to the expression of the ANSI standard) as:

$$P + 2S \leq 0.5$$

1 For doses below 1 rem, ICRP 35 recommends that the uncer-
2 tainties not exceed 100% at the 95% confidence level. Mathe-
3 matically, this is expressed as:

$$P + 2S \leq 1.0$$

5 The following table shows the relationship between the
6 ANSI and ICRP standards for the simple case where $P=0$ (i.e.
7 systematic bias is zero):

<u>Dose Range (rem)</u>	<u>ANSI</u>	<u>ICRP</u>
0-1.0	$S < 0.5$	$S \leq 0.5$
1.0-5.0	$S < 0.5$	*
5.0-10.0	$S < 0.5$	$S \leq 0.25$
10-500	$S < 0.3$	$S \leq 0.25$

11 *ICRP does not clearly address this dose range.

12 As the table shows, there is relatively good overall
13 agreement between ANSI and ICRP under the simple case when $P=0$,
14 especially at the dose levels which are most common in practice
15 (doses less than 1 rem).

16 When $P = 0$, as is usually the case, the comparison of the
17 ANSI and ICRP standards becomes more complex. For the most
18 common dose range (less than one 1 rem), ANSI becomes more re-
19 strictive than ICRP. For doses between 5 and 10 rem ICRP is
20 more restrictive. For doses above 10 rem, the standard which
21 is more restrictive depends on the actual values for P and S .

22 The following examples illustrate that ANSI is more re-
23 strictive than ICRP for doses less than 1 rem.

Example 1

1 Assume P=0.2 and calculate the maximum permissible
2 S according to the ANSI and ICRP formulas.

	<u>Formula</u>	<u>P</u>	<u>Maximum S</u>
3			
4	<u>ANSI</u> P+S < 0.5	0.2	0.3
5	<u>ICRP</u> P+2S < 1.0	0.2	0.4

6 As can be seen ICRP allows a larger S for a given P.

7
Example 2

8 Assume S=0.2 and calculate the maximum permissible
9 value for P according to ANSI and ICRP formulas:

	<u>Formula</u>	<u>S</u>	<u>Maximum P</u>
10			
11	<u>ANSI</u> P+S < 0.5	0.2	0.3
12	<u>ICRP</u> P+2S < 1.0	0.2	0.6

13 Again, ICRP allows a larger P for a given S.

14 Thus, overall, the ANSI standard and ICRP recommendations are
15 relatively compatible, especially at the dose levels which are
16 most common in practice (doses less than one rem).

17 Q.18 Is CP&L committed to maintaining the ANSI standard at
18 the Harris Plant?

19 A.18 Yes. A quarterly TLD intercomparison program has
20 been established with the University of Michigan. This program
21 follows the format of the ANSI performance test, except that
22 CP&L has added two additional radiation categories which are
23 applicable to the radiation types and energies found in its nu-
24 clear plants, and has dropped the accident categories which

1 differ from other categories only in the dose level. This pro-
2 vides a regular, independent test of the TLD system against the
3 ANSI criterion.

4 Q.19 Which two categories have been added by CP&L?

5 A.19 CP&L has added new categories for low energy beta and
6 for mixtures of low energy beta with high energy photons.
7 These two categories are designed to test more fully the TLD
8 system capability under exposure conditions which are similar
9 to those which can occur under certain working conditions in
10 the CP&L plants.

11 Q.20 What results have been obtained in those categories?

12 A.20 The results for these two categories in tests con-
13 ducted during the first quarter of 1984 are shown below:

<u>Radiation Type</u>		<u>CP&L Performance (P+S)</u>	<u>Limit</u>
Beta (low)	shallow	.40	.5
Gamma (high) & Beta (low)	shallow	.26	.5
Gamma & Beta	deep	.22	.5

14 It should be noted that this test represented the first actual
15 exposure of CP&L TLDs to these categories. The calibration
16 factor used during these tests was based on the best available
17 information in the literature prior to the test, rather than on
18 experimental data specific to the CP&L system. Based on these
19 test results and additional beta dosimetry studies performed by
20 CP&L in conjunction with the University of Lowell during 1984,
21 adjustments have been made so that future performance can be
22 expected to improve significantly.
23
24
25

1 Q.21 Are there any existing regulatory requirements that
specify TLD accuracy?

2 A.21 No. I understand that the Board views 10 C.F.R. Part
3 20 as establishing implicit limits for TLD accuracy. In my
4 opinion, Part 20 sets limits on the maximum recorded quarterly
5 and yearly dose to various parts of the body, but does not spe-
6 cifically set forth an accuracy standard for measurement of
7 dose.

8 Q.22 What accuracy standard has been suggested by the
9 Board?

10 A.22 With regard to the NRC regulatory standards for radi-
11 ation dose to individuals, the Board initially noted:

12 There is no indication of any latitude or
13 permissible variance in the application of these
standards that would give the Board guidance con-
14 cerning the accuracy required in dosimetric proce-
dures.

15 From this statement, I believe that the Board is in basic
16 agreement with my earlier statement that the regulations
17 contain no explicit standard for accuracy. However, the Board
18 derived an implied standard of accuracy from 10 C.F.R.
19 § 20.407(b), which provides for a statistical summary report of
20 recorded doses. The Board's interpretation was stated as fol-
21 lows:

22 The Board derives from these regulations an
23 implication that the required radiation exposure
control is in terms of integer values in rem
24 units.

April 13 Memorandum at 9.

The Board has clarified this language, based on the NRC
1 Staff's guidance:

2 The Board accepts the Staff's guidance as
3 cited, and we note that the specification of the
4 limit on the uncertainties as 50 percent is com-
5 patible with our reading of the regulations as
6 specifying integer values at occupational exposure
7 levels.

8 Id. at 12. The Board also stated:

9 As noted above, the Board accepts the Staff's
10 guidance in terms of the ICRP recommendation that
11 at a dose level of 2 rem, 1 rem uncertainty is
12 acceptable.

13 Id. at 16.

14 By these statements, the Board appears to suggest 50% as
15 an accuracy standard without reference to a specific confidence
16 level.

17 After suggesting 50% as the standard, the Board expressed
18 its interpretation of the confidence level requirement as fol-
19 lows:

20 The essential issue from the Board's point of
21 view is that reasonable worker radiation protec-
22 tion and demonstrations of regulatory compliance
are not compatible with the acceptance of perfor-
mance with a standard deviation of 0.5. Conven-
tional interpretation of the 0.5 standard devia-
tion would be that at the 95 percent confidence
level an individual dose estimate would be uncer-
tain by \pm 100 percent. This range or latitude is
not compatible with the Board's reading of the
regulations as calling for controlling radiation
doses to workers with a resolution to integer val-
ues at 1 rem and above.

1 Id. at 16. I must note, however, that this statement overlooks
2 the fact that, according to the ANSI criterion, the sum of the
3 bias plus the standard deviation -- not just the standard
4 deviation -- must be within 0.5. But in any case, the Board
5 appears to consider the 95% confidence level to be appropriate.
6 Combined with previous statements, the standard which the Board
7 seems to have suggested is 50% at the 95% confidence level.

8 The Board also suggested an alternative interpretation for
9 the accuracy standard as follows:

10 As the Board has outlined above, we believe
11 that NRC regulations require that personal
12 dosimetry be carried out in a manner such that the
13 results can be relied upon to be accurate to inte-
14 ger values or one significant figure for doses of
15 a few rem. Such performance could be achieved by
16 limiting acceptable bias to 10 to 20 percent and
17 variability or the standard deviation also to 10
18 to 20 percent.

19 Id. at 19. This proposal is different from the Board's other
20 suggestions in the April 13 Memorandum since no confidence
21 level is specified and ranges, rather than single values, are
22 provided for bias and variability. It could be interpreted as
23 allowing a bias of 20% and standard deviation of 20% at the 95%
confidence level, thus yielding a total uncertainty of 60%
rather than 50% as espoused earlier.

In its ruling on the CP&L motion to reconsider its order
on summary disposition, the Board suggested what appears to be
a third interpretation of the accuracy standard:

1 The issue that we are leaving in [the licens-
2 ing proceeding] reflects our view that the exist-
3 ing regulations do embody a standard of accuracy.
4 They require that the Applicants' dosimetry pro-
5 gram reliably distinguish between doses of 2 and 3
6 rems and between 3 and 4 rems; that is to say, er-
7 rors of larger than half a rem are not permitted.

8 Tr. 2218-19, August 10, 1984 Conference Call.

9 I am unable to reconcile this statement with the Board's
10 previous statements to the effect that an accuracy level of 50%
11 may be required. The standard suggested at the time of the
12 conference call would result in a different accuracy level for
13 each dose level. Only at a dose of 1 rem would an error of
14 0.5 rem represent 50% uncertainty.

15 Q.23 What conclusions have you drawn regarding the Board's
16 suggestion of an implied standard for TLD accuracy?

17 A.23 I have carefully analyzed the Board's statements
18 regarding accuracy and reached the following conclusions:

19 First, as acknowledged by the Board, NRC regulations do
20 not presently contain any explicit guidance on the accuracy re-
21 quired in dose measurements. As noted by the Board in its
22 ruling on Applicants' motion for reconsideration, the NRC's
23 rulemaking proceeding will resolve the question of the accuracy
24 for Applicants and all other licensees. Tr. 2217-18,
25 August 10, 1984 Conference Call.

26 Second, the Board has not articulated a single, unambigu-
27 ous technical standard for accuracy against which CP&L TLD per-
28 formance could be measured.

1 Third, the various statements made about accuracy by the
2 Board are not entirely consistent with the guidance of the pre-
3 vailing standards established by ANSI and ICRP. Applicants are
4 committed, however, to maintaining, at a minimum, the ANSI
5 standard.

6 Q.24 Would you consider a requirement of accuracy within
7 50% at the 95% confidence level for doses of a few rem to be an
8 appropriate standard for TLD accuracy?

9 A.24 No. This standard was suggested by the Board,
10 relying, in part, on ICRP 12 (1968). However, as I have noted,
11 ICRP 12 has been superceded by ICRP 35 (1982) which provides
12 more recent and specific guidance on the subject of accuracy in
13 routine monitoring. For doses on the order of the annual limit
14 (5 rem) ICRP 35 states that the uncertainties should not exceed
15 50% at the 95% confidence level. Therefore the standard sug-
16 gested by the Board is essentially consistent with the recom-
17 mendations of ICRP 35 for doses on the order of 5 rem or
18 greater. However, 5 rem is the annual exposure limit and the
19 actual annual exposure of most workers is below this level. In
20 fact, less than 10% of the workers monitored at the H.B. Rob-
21 inson Plant (CP&L's operating pressurized water reactor
22 ("PWR")) received annual doses during 1983 greater than 1 rem.
23 The Robinson Plant is a single-unit, Westinghouse-designed PWR
24 similar to the Harris Plant. At dose levels below 1 rem, ICRP
25 35 recommends that the uncertainties not exceed a factor of 2
(100%) at the 95% confidence level. This recommendation is

1 essentially equivalent to the ANSI criterion of P+S 0.5, as I
2 previously explained in response to question 17. Although the
3 Board's interpretation that accuracy should be within 50% for
4 doses of a few rem is appropriate for doses of 5 rem or
5 greater, in my opinion it is not appropriate or consistent with
6 ICRP or ANSI recommendations for doses of 1 rem or less, which
7 constitute the majority of actual exposures received.

8 Q.25 Do the TLDs to be used at the Harris Plant nonethe-
9 less comply with an accuracy requirement of 50% at the 95% con-
10 fidence level as suggested by the Board?

11 A.25 Yes. As previously shown, P+2S was less than 0.5 for
12 all categories during the 1982 and 1984 tests.

13 Q.26 The Board also suggested that acceptable performance
14 could be achieved by limiting bias and variability to 10 to
15 20%. Do the TLDs to be issued at the Harris Plant meet this
16 criterion?

17 A.26 Yes. During the 1984 ANSI tests, no individual cate-
18 gory had either bias or standard deviation greater than 20%.
19 During the 1982 ANSI tests, no individual category had a stan-
20 dard deviation greater than 20%, and only one category had a
21 bias greater than 20% (beta--24%). During both the 1982 and
22 1984 tests, the average bias and standard deviation for all
23 categories was less than 10%. A table setting forth the bias
24 and standard deviation as separate values has been prepared and
25 is attached to this testimony as Attachment C. As the table
26 shows, the results achieved by CP&L more than meet the Board's
performance criteria.

1 Q.27 Would you consider a standard of accuracy within 1/2
2 rem as suggested by the Board to be an appropriate standard for
3 TLD accuracy?

4 A.27 No. A standard worded in this way represents a dif-
5 ferent level of accuracy on a percentage basis at every dose
6 level. The following table shows how the percent error allowed
7 would vary as a function of dose. At both high and low doses,
8 the allowed error is unreasonable. A standard is incomplete if
9 it does not cover both high and low doses. Thus, the defini-
10 tion of accuracy stated above is not useful in practice.

<u>Dose</u>	<u>Absolute Error</u>	<u>Percent Error Allowed</u>
0.1	0.5	500%
0.5	0.5	100%
1.0	0.5	50%
2.0	0.5	25%
5.0	0.5	10%
10.0	0.5	5%

11
12
13
14
15 Q.28 What steps will CP&L take to ensure that the TLD sys-
16 tem at the Harris Plant continues to meet the applicable stan-
17 dard for accuracy?

18 A.28 As stated in NUREG-2891, four common reasons for poor
19 performance of dosimetry processors were observed during the
20 1982 tests at the University of Michigan. These causes of in-
21 accuracy were: 1) use of incorrect calibration factors; 2)
22 dosimeter variability; 3) clerical errors; and 4) poor
23 calibration for accident doses. CP&L has taken steps to mini-
mize errors in each of those four areas.

1 Q.29 How does CP&L seek to minimize the use of incorrect
2 calibration factors?

3 A.29 Extensive quality control measures are applied to the
4 processing of TLDs and recording of individual doses to ensure
5 accuracy. With regard to incorrect calibration factors, CP&L
6 has taken the following approach. First, calibration factors
7 have been determined for the TLD system based on irradiation of
8 TLDs to NBS traceable radiation standards for beta, gamma,
9 x-rays, and neutrons. Second, these correction factors have
10 been verified by the tests performed at the University of
11 Michigan in 1982 and 1984, which were discussed previously, and
12 by the on-going quarterly intercomparison program previously
13 mentioned. If calibration factors were not valid, the perfor-
14 mance on these tests would so indicate. Third, a monthly in-
15 house cross-check program is in operation. Under this program,
16 blind audit TLDs irradiated with an NBS traceable dose standard
17 are processed on each TLD reader. The performance limit (P+S)
18 of 0.3 is more restrictive than the ANSI limit of 0.5. Fourth,
19 each TLD reader is calibrated semiannually and following any
20 maintenance which affects calibration. A preventative mainte-
21 nance program is followed for all equipment. Fifth, a daily
22 calibration check using TLDs irradiated to known doses using
23 NBS traceable standards is performed. The performance limit
24 for this check requires the accuracy to be within 15%. Sixth,
25 numerous critical parameters in the TLD reader are monitored
26 automatically prior to reading each TLD badge to ensure
stability of equipment during operation.

1 Q.30 How accurate is CP&L's semi-annual calibration equip-
2 ment?

3 A.30 The acceptance criterion for semi-annual calibration
4 of TLD readers requires that 10 TLDs be read at each of 5 known
5 dose levels ranging from 0.25 to 4.0 rem. At each dose level,
6 the average observed reading must be within $\pm 10\%$ of the actual
7 irradiated dose and the percent standard deviation must not ex-
8 ceed 10%.

9 Q.31 What about the daily calibration check of TLD reading
10 equipment?

11 A.31 For daily TLD reader calibration checks, TLDs are
12 read after being irradiated to the following known doses:
13 0.5 rem and 4.0 rem. Each TLD must read within $\pm 15\%$ of the
14 actual irradiated dose. If a reading within $\pm 15\%$ is not
15 obtained, the check is repeated two more times; if the check
16 fails two out of three times, the TLD reader is removed from
17 service. If the observed dose is outside $\pm 25\%$ of the actual
18 dose, all TLD readings since the last satisfactory check are
19 reviewed to determine whether they are valid.

20 Q.32 What steps are taken to minimize dosimeter
21 variability?

22 A.32 CP&L performs an initial acceptance test on each TLD
23 prior to placing it in service. This test identifies TLDs
24 whose response to known radiation doses falls outside approxi-
25 mately the 99% probability level relative to the mean of a sam-
ple of at least 500 TLDs. For samples containing less than 500

1 TLDs, the acceptance criterion requires the TLD response to be
2 within \pm 15% of the irradiated value. TLDs which fail initial
3 acceptance testing are returned to the manufacturer for re-
4 placement. After being placed in service, all TLDs undergo
5 semiannual quality control checks to detect any change in re-
6 sponse. The same procedures and acceptance criteria are used
7 during these quality control checks, and any TLDs which fail
8 are removed from service.

8 Q.33 Are steps taken to prevent clerical errors?

9 A.33 With regard to clerical errors, CP&L has implemented
10 automatic data processing techniques and detailed verification
11 procedures. Individual dose records are maintained using a
12 computer system. The TLD readers are interfaced to this com-
13 puter system, and most TLD readings are transferred elec-
14 tronically to the assigned individual's dose history record.
15 Of course, no data is transferred to the computer system prior
16 to a technical review of the data, and built-in checks ensure
17 the data is transferred completely and correctly. Through elec-
18 tronic transfer and automated data processing techniques, the
19 opportunity for clerical errors is greatly reduced.

20 Although some data is still entered into the computer sys-
21 tem manually, all manual data entry is independently verified
22 by a different individual. In many cases, the data entry is
23 actually verified twice by two different individuals. In addi-
24 tion, hard copies of all source documents are retained perma-
25 nently on microfilm to backup the computer record system.

1 Q.34 What steps has CP&L taken with regard to poor
2 calibration for accident doses?

3 A.34 With regard to poor calibration for accident doses,
4 CP&L has performed in-house tests which establish the dose re-
5 sponse of the TLDs up to doses of 100 rem. The response is es-
6 sentially linear within approximately $\pm 15\%$. In addition, CP&L
7 has participated in and passed the accident dose categories
8 during ANSI performance tests in 1982 and 1984. The ANSI acci-
9 dent test range is 10 rem to 500 rem. This verifies that poor
10 calibration for accident doses is not a problem at CP&L.

11 Q.35 Has CP&L undertaken other measures to minimize the
12 potential for poor TLD performance?

13 A.35 In addition to the specific quality control measures
14 described above which address the major reasons for poor per-
15 formance noted in NUREG CR-2891, the overall quality assurance
16 program includes: (1) detailed written procedures for the per-
17 formance of all routine dosimetry operations; (2) formal
18 training and qualification for all operating personnel; and (3)
19 formal supervisory review of all quality control records. All
20 of these measures ensure that TLDs will be used and processed
21 correctly at the Harris Plant.

22 In addition, as part of the overall quality control ef-
23 fort, CP&L has applied for accreditation of its dosimetry labo-
24 ratory under the recently announced Dosimetry Processor Labora-
25 tory Accreditation Program administered by the NBS. CP&L has
successfully completed the performance testing requirements for

1 accreditation. The on-site inspection of CP&L by NVLAP assess-
2 sors was completed in late August.

3 Furthermore, while CP&L is confident that its dosimetry
4 program is in compliance with NRC regulations, additional im-
5 provements are always being made in an effort to maintain a
6 state-of-art program. Recently, major studies have been com-
7 pleted in the areas of beta and neutron dosimetry. These
8 studies were designed to characterize the response of the TLDs
9 to the typical radiation spectra found in CP&L's nuclear plants
10 and to use this knowledge to improve the accuracy of measure-
11 ment of personnel doses. CP&L continually seeks to upgrade the
12 quality of its dosimetry program through the better training of
13 personnel, new equipment, and improved procedures.

14 Q.36 Please summarize your opinion of the adequacy of
15 CP&L's TLD system for use at the Harris Plant.

16 A.36 In my opinion, CP&L has a well-established, state-of-
17 the-art dosimetry program which has operated successfully for
18 over ten years. The program employs up-to-date equipment and
19 facilities. Within the last three years CP&L has invested well
20 over one million dollars in TLD equipment and has committed
21 over \$200,000 more this year for additional equipment to sup-
22 port the Harris Plant. A highly trained and experienced staff
23 is maintained to provide dedicated support to the dosimetry
24 program. The central support staff includes four professional
25 personnel, five technicians, and two clerks. Each operating
plant, including the Harris Plant, also maintains or will

1 maintain a separate dosimetry staff headed by a dosimetry fore-
2 man. The program is operated under detailed quality assurance
3 procedures which incorporate elaborate quality control checks
4 on all aspects of dosimetry processing and record keeping.

5 CP&L's use of TLDs satisfies all applicable NRC regula-
6 tions, is in substantial compliance with the standards of the
7 nationally and internationally recognized organizations in the
8 field, and conforms to the ANSI N13.11 standard that is the
9 basis for the NVLAP program and the proposed NRC rule. Thus,
10 the dosimetry program to be established at the Harris Plant,
11 utilizing Panasonic TLDs and CP&L's calibration and quality
12 control program, has already demonstrated a consistent degree
13 of accuracy in radiation doses measurement to ensure compliance
14 with NRC regulations and is adequate to protect worker health
and safety.

ATTACHMENT A

Stephen A. Browne
Harris Energy & Environmental Center
Carolina Power & Light Company
New Hill, North Carolina 27562

Education and Training

B.S. degree in Physics, Union College (1971)

M.S. degree in Environmental Health Engineering, Northwestern University (1974)

Professional Societies

Health Physics Society

Experience

- A. 1972 to 1974 - Radiation Safety Officer, Packard Instrument Company, Downers Grove, Ill.
- B. 1974 to September 1978 - Health Physicist, General Electric Company, Knolls Atomic Power Laboratory, Windsor, Conn.
September 1978 to April 1979 - Lead Engineer, General Electric Company, Knolls Atomic Power Laboratory, Windsor, Conn.
- C. April 1979 to October 1981 - Senior Specialist - Dosimetry, Carolina Power & Light Company, New Hill, N.C.
October 1981 to present - Project Specialist - Health Physics, Carolina Power & Light Company, New Hill, N.C.

Performance Testing of Personnel Dosimetry Services

Final Report of Test #3

Manuscript Completed: November 1982
Date Published: February 1983

Prepared by
P. Plato, J. Miklos

The University of Michigan
School of Public Health
Ann Arbor, MI 48109

Prepared for
Division of Facility Operations
Office of Nuclear Regulatory Research
U.S. Nuclear Regulatory Commission
Washington, D.C. 20555
NRC FIN B1049

Table 3 continued.

Radiation Categories	I	II	III		IV	V	VI		VII		VIII
	X ray Accident	Gamma Accident	X ray Shallow	X ray Deep	Gamma	Beta	Gamma + X ray Shallow	X ray Deep	Gamma + Beta Shallow	Gamma + Beta Deep	Gamma + Neutron
Dosimeter Code Number											
182			0.5024	0.1335	0.1005	0.5300	0.3693	0.0891	0.1902	0.0526	1.7145
183					0.0627	0.8957			0.8006	0.0931	
184					0.5416						1.2909
185	0.1297	0.0919	0.1631	0.1881	0.1196	0.0831	0.1723	0.2052	0.1710	0.0843	0.1908
186*											
187	0.2406	0.1052	0.1041	0.1228	0.0614	0.3053	0.0583	0.1594	0.1640	0.1060	
188	0.9237	0.6027	1.0567	0.6499	0.2434	0.3791	0.8830	0.6737	0.5767	0.3368	0.9632
189	0.2355	0.1608	0.3993	0.6199	0.2607	0.3423	0.3385	0.2740	0.4060	0.3176	0.2812
190			0.5470	0.5344	0.1708	0.5538	0.3569	0.3696	0.4432	0.3985	0.1453
191*											
192	0.3187	0.1155	1.4558	0.5534	0.2076	1.4166	0.6948	0.3518	0.7508	0.4031	0.6550
193	1.9733	0.7837	3.0410	0.0966	0.0499	0.1220	3.3129	0.1167	0.1235	0.1860	0.3201
194	1.0326	0.5363	1.4322	1.0703	0.0652	0.2885	1.6137	0.3667	0.1670	0.0818	0.1844
195*											
196					0.2982	0.2348			0.3628	0.4340	
197	0.3867	0.1307	0.4789	0.4493	0.1022	0.1111	0.3294	0.2351	0.1173	0.1225	0.0939
198*											

ATTACHMENT C

CAROLINA POWER & LIGHT

ANSI-N13.11-1983 PERFORMANCE TESTING COMPARISON

CATEGORY	DESCRIPTION	DEPTH	2nd Qtr 1982			1st Qtr 1984			L
			P	S	P +S	P	S	P +S	
I	Accident, Low-Energy Photon	Deep	.11	.13	.24	.07	.11	.18	.3
II	Accident, High-Energy Photon	Deep	-.06	.04	.10	-.09	.06	.15	.3
III	Low Energy Photons	Shallow	.06	.05	.11	.10	.08	.18	.5
		Deep	.02	.10	.12	.07	.09	.16	.5
IV	High Energy Photons	Deep	-.03	.03	.06	-.03	.07	.10	.5
V	Beta Particles	Shallow	.24	.06	.30	.19	.09	.28	.5
VI	Photon Mixtures	Shallow	.00	.06	.06	.12	.07	.19	.5
		Deep	.09	.07	.16	.08	.10	.18	.5
VII	Photon plus Beta	Shallow	.10	.06	.16	.17	.12	.29	.5
		Deep	.05	.06	.11	-.02	.08	.10	.5
VIII	Photon plus Neutrons	Deep	-	-	-	.01	.07	.09	.5
	Average		.08	.07	.13	.06	.09	.17	

P = Performance quotient = (Reported-Delivered)/Delivered

S = Standard deviation of P

L = Tolerance limit

Specifications of TL Badge and
TL Badge Hanger
(Thermoluminescent Dosimeters)

SPECIFICATION OF
TL BADGE
MODEL SPECIFICATIONS

UD-801A
UD-802A
UD-803A
UD-804A
UD-806A
UD-807A
UD-808A
UD-809A
UD-811A
UD-815A

MATSUSHITA ELECTRIC INDUSTRIAL CO., LTD.
MATSUSHITA INDUSTRIAL EQUIPMENT CO., LTD.
ELECTRONIC EQUIPMENT DEPARTMENT
1-1, 3-CHOME, INAZU-CHO, TOYONAKA
OSAKA, 561 JAPAN

* Specifications subject to change without notice.

MODEL SPECIFICATIONS

TL Badge : UD-802A group

Refer to the GENERAL SPECIFICATIONS for common specifications.

1. Model number : UD-802AQ , UD-802AR , UD-802AS
 2. Use : Personnel monitoring
 3. Applicable reader : UD-710A, UD-702E, UD-720A
 4. Appearance : Fig. 1
 5. Element and shield composition : Table 1
 6. Measurable rays and range : γ -x rays (10keV \sim 10MeV) 1mrem \sim 1000rem
(Rough energy evaluation is possible)
 β rays (0.5MeV \sim 4MeV) 10mrem \sim 1000rem
- (Measurable range is in the case where single kind of rays is measured.)
7. Recommended hanger : UD-875A, UD-885A
 8. ID number : Specified serial numbers (7digits at maximum) are punched.
 9. Label : The labels shown in Fig. 1 are stucked.
 10. Weight : Less than grams.




Fig. 1 Appearance of TL Badge: UD-802AQ
(The figure also applies to UD-802AR and UD-802AS)

Table 1 Element and shield composition of TL Badge
UD-802A group

Element	Phosphor	Shield
E1	${}^n\text{Li}_2{}^n\text{B}_4\text{O}_7(\text{Cu})$	Plastics 14mg/cm ²
E2	${}^n\text{Li}_2{}^n\text{B}_4\text{O}_7(\text{Cu})$	Plastics 160mg/cm ²
E3	$\text{CaSO}_4(\text{Tm})$	Plastics 160mg/cm ²
E4	$\text{CaSO}_4(\text{Tm})$	Lead 0.7mm thick

* The thickness of the hanger is not included. To obtain the total thickness when the badge is placed in the hanger, refer to the specifications of the hanger and add that thickness for each element.

SPEC NO. E-BDG/GS-1

SPECIFICATION OF
TL BADGE

GENERAL SPECIFICATIONS

JANUARY 1983

MATSUSHITA ELECTRIC INDUSTRIAL CO., LTD.
MATSUSHITA INDUSTRIAL EQUIPMENT CO., LTD.
ELECTRONIC EQUIPMENT DEPARTMENT
1-1, 3-CHOME, INAZU-CHO, TOYONAKA
OSAKA, 561 JAPAN

* Specifications subject to change without notice.

CONTENTS

INTRODUCTION	1
1. GENERAL	2
2. APPEARANCE AND CONSTRUCTION	2
3. DETAILED COMMON SPECIFICATIONS	3
3-1 Thermoluminescence phosphor	3
3-2 Element Thickness	4
3-3 Measurement Range	5
3-4 Fading	5
3-5 Tribo-thermoluminescence	5
3-6 Light Response	5
3-7 Sensitivity	5
3-8 Life Time	6
3-9 Working Conditions	6
3-10 Bit Format of Optical Code Hole on the TL Badge .	7
4. GUARANTEE	7
Appendix 1.	7
Fig. 1	9
Fig. 2	10
Fig. 3	11
Fig. 4	12

INTRODUCTION

These specifications are common to all National/Panasonic TL Badges. As for detailed specifications, such as phosphorous materials and shield combinations, refer to the "Specifications for each TL Badge model."

For many measurements and dosage assessment applications, the National/Panasonic TL Badge must be used with a TL Badge Hanger.

As for the National/Panasonic TL Badge Hanger, refer to the "Specifications common to all TL Badge Hangers" and the "Specifications for each TL Badge Hanger model."

1. GENERAL

The TL Badge is a thermoluminescence dosimeter for measurement of integral dose of radiation.

Elements (up to four) are mounted on an element plate.

The element plate is put in the badge holder.

The element is made of a thin layer of thermoluminescence phosphor. A radiation shield is provided on the badge holder.

An identification number (up to seven digits) is provided on the badge and automatically read by a reader.

The badge is measured without the necessity of taking the element plate out from the badge holder. The element plate is locked to the badge holder to prevent deliberate removal of the plate.

2. APPEARANCE AND CONSTRUCTION

Figure 1 (a), (b), (c) shows the appearance of the TL Badge. Depending on the use, three types are available; one-window type, two-window type, and no-window type.

An extremity dosimeter shown in figure 1 (d) is also available. Forty-bit optical code holes can be made in the badge holder to provide the ID number, etc.

Figure 2 shows the sectional view and the dimensions of the TL Badge. (Though it shows the one-window type UD-802A, the figure also applies to other types.)

Figure 3 shows the enlarged sectional view of the element.

Virtually mono-layer of granular phosphor (average granule diameter is 90μ) is formed on the thin resin substrate. The phosphor is covered with a protective, transparent film.

3. DETAILED COMMON SPECIFICATIONS

3-1 Thermoluminescence phosphor

Depending on specific measurement, four kinds of phosphor are mounted on the plate in various combinations.

${}^6\text{Li}; {}^{\text{n}}\text{B}_2\text{O}_7(\text{Cu})$ The phosphor is made of tissue-equivalent $\text{Li}_2\text{B}_2\text{O}_7$ with Cu doped as an activator. Natural abundant material ${}^6\text{Li}$ and ${}^{\text{n}}\text{B}$ is used for Li and B. The phosphor is sensitive to γ .x rays and β rays and somewhat sensitive to low-energy neutron.

${}^7\text{Li}; {}^{10}\text{B}_2\text{O}_7(\text{Cu})$ ${}^{\text{n}}\text{Li}$ and ${}^{\text{n}}\text{B}$ in ${}^{\text{n}}\text{Li}; {}^{\text{n}}\text{B}_2\text{O}_7(\text{Cu})$ mentioned above are replaced by enriched material ${}^7\text{Li}$ and ${}^{10}\text{B}$. The phosphor has little sensitivity to neutron rays. The response to γ .x rays and β rays is equivalent to ${}^{\text{n}}\text{Li}; {}^{\text{n}}\text{B}_2\text{O}_7(\text{Cu})$

${}^6\text{Li}$, ${}^{10}\text{B}_2\text{O}_3$, (Cu) ${}^n\text{Li}$ and ${}^n\text{B}$ in ${}^n\text{Li}$, ${}^n\text{B}_2\text{O}_3$, (Cu) mentioned above are replaced by enriched material ${}^6\text{Li}$ and ${}^{10}\text{B}$. The phosphor has high sensitivity to low-energy neutron rays. The response to γ .x rays and β rays is equivalent to ${}^n\text{Li}$, ${}^n\text{B}_2\text{O}_3$, (Cu).

CaSO_4 , (Tm) The phosphor is made of non-tissue-equivalent CaSO_4 with Tm doped as an activator. Since the phosphor has high sensitivity it is used to evaluate low dose in combination with metal shield. Without metal shield, it shows over response to low-energy γ .x rays, and is used for energy evaluation of γ .x rays. The phosphor has little sensitivity to neutron rays.

3-2 Element Thickness

- a) Phosphor layer: Approx. 15 mg/cm²
- b) Substrate: Approx. 11 mg/cm²
- c) Transparent cover film: Less than 28mg/cm²

3-3 Measurement Range

- a) $\text{Li}_2\text{B}_4\text{O}_7(\text{Cu})$ elements: 10 mrem to 1000 rem
(^{60}Co - γ ray equivalent)
- b) $\text{CaSO}_4(\text{Tm})$ element: 1 mrem to 200 rem
(^{60}Co - γ ray equivalent)

3-4 Fading

- a) $\text{Li}_2\text{B}_4\text{O}_7(\text{Cu})$ elements: Less than 10%/month,
- b) $\text{CaSO}_4(\text{Tm})$ element: 3%/month,
at the room temperature (25°C)

3-5 Tribo-thermoluminescence

- a) $\text{Li}_2\text{B}_4\text{O}_7(\text{Cu})$ elements: Lower than detection limit,
- b) $\text{CaSO}_4(\text{Tm})$ element: Lower than detection limit.

3-6 Light Response (light sensitivity and light fading)

- a) $\text{Li}_2\text{B}_4\text{O}_7(\text{Cu})$ elements: Lower than detection limit,
- b) $\text{CaSO}_4(\text{Tm})$ element: Lower than detection limit,
where mounted in the badge holder.

3-7 Sensitivity uniformity of element

There are three classes in the nominal uniformity of element; Class Q, Class R, and Class S. The class is indicated in the seventh digit of the model number.

(Example: UD-802AQ, UD-802AR, UD-802AS etc.)

Nominal uniformity indicated by the classes are as follows:

Class	Nominal uniformity
Q	Percentage standard deviation \leq 5.0%
R	Percentage standard deviation \leq 7.5%
S	Maximum to minimum deviation \leq \pm 30%

It is assumed that the uniformity is evaluated using the reader with rank correction function.

In evaluating uniformity, Class Q sometimes shows a percentage standard deviation of 5 to 6% and Class R does 7 to 8% depending on the exposing method and sampling method.

3-8 Life Time

Repeated use: 300 times,
Total exposure: Lower than 10 rem
(⁶⁰Co- γ ray equivalent)

3-9 Working Conditions

Temperature: $-10^{\circ}\text{C} \sim 40^{\circ}\text{C}$ (Fading increases above 35°C .)
Humidity: Less than 80% RH (without dripping)

For precautions before measurements, refer to the handling instructions.

3-10 Bit Format of Optical Code Hole on the TL Badge

The bit format of the optical code hole is shown in Fig. 4.

The 40 bits are used as follows;

- 28 bits : Number code (7 digits)
- 4 bits : Model code
- 6 bits : Element sensitivity correction code
- 2 bits : Parity code (Odd parity)

4. GUARANTEE

One year from date of delivery.


Replacement of new badge shall constitute a fulfillment of all obligations to the purchaser. The Panasonic will not be responsible for any damage resulting from improper use.

Please read the handling instructions carefully before use.

Appendix 1.

Model number format

The model number of TL Badges is constituted as follows;



- (i) . A code defining the composition of the element and shield.

01 to 15

- . Codes from 01 to 11 are assigned to standard specification badges and the composition is fixed.
- . Codes from 12 to 14 are assigned to special specification badges and the composition is determined by the purchaser. (Refer to (iii).)

(In this case, the reader parameter must be set according to the badge composition. The badge defining parameter for the UD-702E manual operation reader cannot be varied.)

- . Code 15 is assigned to the reader calibration badge.


- (ii) A code defining the nominal uniformity class.

(Refer to 3-(7).)

- (iii) Special code

When a special specification code (between 12 and 14) is assigned to term (i), many TL Badges containing the same model number will result. To identify these badges, a serial number is assigned to the badges. For example, UD-812AQ has many variations: UD-812AQ1, UD-812AQ2, ... and incorrect readings will result when UD-812AQ1 is read by the reader given the badge defining parameters for UD-812AQ2.

(The special code may be assigned to the standard badges when necessary.)



(d) extremity dosimeter

(c) no-window type

(b) two-window type

(a) one-window type

Fig. 1 Appearance of TL Badge




Fig. 2 Sectional view of the TL Badge
(UD-802A as an example).




Fig. 3 Enlarged sectional view of the element.




Fig. 4 Bit format of optical code hole.

SPECIFICATION OF
TL BADGE HANGER
MODEL SPECIFICATIONS

UD-854A
UD-874A
UD-875A
UD-876A
UD-885A
UD-886A
UD-887A

MATSUSHITA ELECTRIC INDUSTRIAL CO., LTD.
MATSUSHITA INDUSTRIAL EQUIPMENT CO., LTD.
ELECTRONIC EQUIPMENT DEPARTMENT
1-1, 3-CHOME, INAZU-CHO, TOYONAKA
OSAKA, 561 JAPAN

* Specifications subject to change without notice.

MODEL SPECIFICATIONS

TL Badge Hanger : UD-854A(H)

Refer to the GENERAL SPECIFICATIONS for common specifications.

1. Model number : UD-854A UD-854AH
2. Type : Open type for one badge
3. Applicable TL Badge (Examples) : UD-801A, UD-815A
4. Appearance : Fig. 1
5. Body material : ABS resin
6. Front wall thickness : For element 1 open
for element 2 plastics 160 mg/cm²
for element 3 plastics 160 mg/cm²
for element 4 plastics 160 mg/cm²
7. Attaching device : An aluminium clip
(with a nylon strap :
UD-854AH only)
8. Color : Clear with slight smokiness.
9. Weight : Less than grams.




Fig. 1 Structure of hanger: UD-854A.

SPEC NU. E-HGR/GS-1

SPECIFICATION OF
TL BADGE HANGER
GENERAL SPECIFICATIONS

JANUARY 1983

MATSUSHITA ELECTRIC INDUSTRIAL CO., LTD.
MATSUSHITA INDUSTRIAL EQUIPMENT CO., LTD.
ELECTRONIC EQUIPMENT DEPARTMENT
1-1, 3-CHOME, INAZU-CHO, TOYONAKA
OSAKA, 561 JAPAN

* Specifications subject to change without notice.

CONTENTS

INTRODUCTION	1
1. GENERAL	2
2. CONSTRUCTION	2
3. SHIELDING	3
4. GUARANTEE	3
Reference	4
Appendix 1.	5
Appendix 2.	6

INTRODUCTION

These specifications are common to all National/Panasonic TL Badge Hangers. As for detailed specifications regarding each hanger, refer to the "Specifications for each TL Badge Hanger model."

For many measurements and dosage assessment applications, the National/Panasonic TL Badge must be used with the TL Badge Hanger.

As for the National/Panasonic TL Badge, refer to the "Specifications common to all TL Badges" and "Specifications for each TL Badge model."

1. General

The TL Badge Hanger serves as both a wearing case and as an additive shield for the TL Badge. Except for special cases, TL Badges should be placed in the hanger before being used.

There are three kinds of hangers: an open hanger for single badges, a closed hanger for single badges, and a closed hanger for two badges.

2. Construction

The three above hangers are shown in Fig. 1.

A γ - β TL Badge is encased in an open hanger, a closed hanger, and also in the right side of the closed hanger for two badges. The hangers are available in several thickness depending on the γ - β TL Badges used.

The UD-809A neutron TL Badge is placed in the left side of the closed two-badge hanger.

The open hanger has a window for the β badge. The closed hanger's window is sealed with a 3 mg/cm^2 thick plastic film, making its wall thickness slightly larger for the β -rays (skin dose) estimating element. The projection shown in Fig. 1 (b) and (c) provides a wall thickness of 1000 mg/cm^2 for the deep dose estimating element.

Closed hangers are designed so that they cannot be easily opened without special jigs. There is a space for a label on each surface.

3. Shielding

3-1 Shielding for γ - β TL Badges

The front wall of the hanger is made of plastic. The thickness of each model is specified in the specifications for each model.

When a TL Badge is encased in a hanger, the total thickness for the γ .x and β rays is obtained by adding the hanger thickness to the badge shield thickness. Examples are shown in Appendix 1.

3-2 The shield for a neutron badge

An additional cadmium-made shield for the UD-809A neutron badge can be formed on the left side of the closed two-badge hanger.

4. Guarantee

This guarantee is good for three months from the date of delivery. This guarantee does not cover problems caused by misuse, abuse, or negligence.

Panasonic will repair or replace, at no charge, your TL Badge Hanger if any problem develops during the guarantee period due to design or workmanship defects.

(Refer to "TL Badge Handling Instructions" for handling.)


Appendix 1. An example of calculation of total wall thickness

Example : UD-808A TL Badge in UD-887A Hanger

Element	Phosphor	Shield thickness of UD-808A	Wall thickness of UD-887A	Total thickness	Remarks
E1	${}^7\text{Li}_2 {}^{11}\text{B}_4\text{O}_7(\text{Cu})$	Plastics 14 mg/cm ²	Plastics 3 mg/cm ²	Plastics 17 mg/cm ²	For β -ray (skin dose) evaluation including β -ray energy correction.
E2	${}^7\text{Li}_2 {}^{11}\text{B}_4\text{O}_7(\text{Cu})$	Plastics 60 mg/cm ²	Plastics 3 mg/cm ²	Plastics 63 mg/cm ²	
E3	$\text{CaSO}_4(\text{Tm})$	Plastics 160 mg/cm ²	Plastics 840 mg/cm ²	Plastics 1000 mg/cm ²	γ -ray energy evaluation by comparing with E4.
E4	${}^7\text{Li}_2 {}^{11}\text{B}_4\text{O}_7(\text{Cu})$	Plastics 160 mg/cm ²	Plastics 840 mg/cm ²	Plastics 1000 mg/cm ²	Deep dose evaluation

Appendix 2. Model number format

The model number of the TL Badge Hanger is constituted as follows;



(i) Type code

5 ----- Open type

7 ----- One-badge closed type

8 ----- Two-badge closed type

(ii) Front wall thickness code (Refer to the figure below.)

(iii) Nylon strap code


H ----- Nylon strap provided

(iv) Special code

A serial number is assigned to the special specifications hangers.

(v) Color code

T ----- Transparent



(a) One-badge
open type

(b) One-badge
closed type

(c) Two-badge
closed type

Fig. 1 Typical appearance of TL Badge hanger