

JUL 20 1984

FCMLB:FAS
930-02138
(17701)

Veterans Administration Medical Center
ATTN: Mr. R. E. Struble
Medical Center Director
Iron Mountain, Michigan 49801

Gentlemen:

This is in reference to a letter, dated June 5, 1984, signed by James Arnold, M.D., requesting an amendment to Byproduct Material License No. 21-12916-01. In order to continue our review, we need the following additional information:

1. Submit a revised diagram of your new storage area which includes the information requested in Item 11 of Regulatory Guide 10.8 (copy enclosed).
2. Prior to release of your previous storage area for unrestricted use, a comprehensive radiation survey should be made which establishes that contamination is within the limits of Table 1 of the enclosed decontamination guide. Note that the reported action level of 200 dpm/square centimeter is not within the limits of Table 1. A copy of the survey report shall be filed with the Chief, Material Licensing Branch, Division of Fuel Cycle and Material Safety, USNRC, Washington, D.C. 20555, and also the Administrator, USNRC Regional Office III, 799 Roosevelt Road, Glen Ellyn, Illinois 60137.

We will continue our review of your application upon receipt of this information. Please reply in duplicate and refer to Control No. 17701.

Sincerely,

8410040137 840921
NMS LIC30
21-12916-01 PDR

Francis A. St. Mary
Material Licensing Branch
Division of Fuel Cycle and
Material Safety

Enclosures:

1. Regulatory Guide 10.8
2. Decontamination Guide

OFFICE	?	FCML					
SURNAME		FAStMary					
DATE		07/19/84					

JUL 20 1984

FCMLB:FAS
030-02138
(17701)

Veterans Administration Medical Center
ATTN: Mr. R. E. Struble
Medical Center Director
Iron Mountain, Michigan 49801

Gentlemen:

This is in reference to a letter, dated June 5, 1984, signed by James Arnold, M.D., requesting an amendment to Byproduct Material License No. 21-12916-01. In order to continue our review, we need the following additional information:

1. Submit a revised diagram of your new storage area which includes the information requested in Item 11 of Regulatory Guide 10.3 (copy enclosed).
2. Prior to release of your previous storage area for unrestricted use, a comprehensive radiation survey should be made which establishes that contamination is within the limits of Table 1 of the enclosed decontamination guide. Note that the reported action level of 200 dpm/square centimeter is not within the limits of Table 1. A copy of the survey report shall be filed with the Chief, Material Licensing Branch, Division of Fuel Cycle and Material Safety, USNRC, Washington, D.C. 20555, and also the Administrator, USNRC Regional Office III, 799 Roosevelt Road, Glen Ellyn, Illinois 60137.

We will continue our review of your application upon receipt of this information. Please reply in duplicate and refer to Control No. 17701.

Sincerely,

Francis A. St. Mary
Material Licensing Branch
Division of Fuel Cycle and
Material Safety

Enclosures:

1. Regulatory Guide 10.8
2. Decontamination Guide

?

FCML *FL*
FAStMary

07/17/84