



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
1600 E. LAMAR BLVD
ARLINGTON TX 76011-4511

April 7, 2020

E. Jefferson Fairbanks, Ph.D.
Radiation Safety Officer
Teton County Hospital District
dba St. John's Hospital
P.O. Box 428
Jackson, WY 83001

SUBJECT: LICENSE NOTIFICATION

By letter dated January 17, 2020, you requested to add two sealed sources to your material license. A member of your staff indicated that the sealed sources will be used as reference sources and shielding. In accordance with 10 CFR 35.65, the sealed sources you requested are less than 30 millicuries (per sealed source) and are not required to be listed on your license. Therefore, your request does not require a modification to your existing material license. The documentation that you submitted, along with your request, is very important and should be retained as part of your records. Your records may be reviewed during your next routine inspection.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Thank you for your cooperation.

Sincerely,

Michelle R. Simmons, Senior Health Physicist
Materials Licensing and Decommissioning Branch

Docket: 03014770
License: 49-18276-01
Control: 617630