

Honeywell

December 20, 1984

Ms. B. J. Holt
Licensing Section
U.S. Nuclear Regulatory Commission
799 Roosevelt Road
Glen Ellyn, IL 60137

Reference: Control Number 24247

Dear Ms. Holt:


The information below is submitted as a clarification to the amendment request to source material license SUB-1430, dated December 20, 1983 and referenced in our telephone conversation of December 19, 1984.

- o The frequency of calibration referred to in item 4 will be 6 month intervals.
- o Item 5 should be deleted and amended to read: "Section 11.f.2(b) should be deleted as calibration beta/gamma detection will be through an outside vendor noted in 4 above".
- o Item 6 should be deleted and amended to read: "Section 11.f.2(c) should be deleted as calibration of alpha detection equipment will be through an outside vendor noted in 4 above".
- o Item 9 should be deleted and amended to read: The first sentence of Section 12b should be deleted and amended to read: "Procedures and actions covering the special hazards that are associated with the handling of DU at Joliet AAP are detailed in the Joliet AAP Safety Manual".

My apology for any inconvenience these discrepancies may have caused. Please contact me at (612)638-5205 if you have any questions or require additional information.

Sincerely,

HONEYWELL, INC.


James A. Fitzsimmons
Radiation Management Engineer
MN30-2546

RECEIVED
DEC 24 1984
REGION III

cc: Mr. Douglas Thompson, Commander's Representative, Joliet AAP
JAF/jc

8503110511 850115
REQ3 LIC40
SUB-1430

PDR

DEFENSE SYSTEMS DIVISION

NEW BRIGHTON, MINNESOTA 55112, TELEPHONE 612/638-5000

DEC 24 1984

CONVERSATION RECORD

TIME

2:35

DATE

12/19/84

TYPE

☐ VISIT

☐ CONFERENCE

☒ TELEPHONE

☐ INCOMING

☒ OUTGOING

ROUTING

NAME/SYMBOL

INT

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

ORGANIZATION (Office, dept., bureau, etc.)

TELEPHONE NO.

Jean Fitzsimmons

Honeywell

(612) 638-5205

SUBJECT

Amendment Request

Control No. 34247

SUMMARY

I requested the following additional information:

1) Frequency of calibration of radiation detection equipment and instrumentation by Health Physics Assoc.

2) Modification of items 5, 6, and 9 in letter dated December 20, 1983 to specify correct references in application with attached letter dated 2/4/83.

ACTION REQUIRED

30 day response

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

B. J. Shaw

12/19/84

ACTION TAKEN

SIGNATURE

TITLE

DATE