

October 23, 1984

Mr. Harold R. Denton, Director Office of Nuclear Reactor Regulation U.S. Nuclear Regulatory Commission Washington, DC 20555

> Subject: Byron Generating Station Units 1 and 2 Braidwood Generating Station Units 1 and 2 Post Accident Sampling System NRC Docket Nos. 50-454/455/456/457

Reference (a): D. H. Smith letter to H. R. Denton dated October 15, 1984

Dear Mr. Denton:

This is to clarify the intent of Attachment 7 to Reference (a). Regarding the Containment High Range Radiation Monitors at Byron, we are requesting a waiver from the factory calibration requirements in NUREG 0737 on the basis that the calibration described in Attachment 7 is an acceptable alternative. We were not addressing any change to "In Situ" calibration.

Please address further questions regarding this matter to this office. One signed original and fifteen copies of this letter are provided for NRC Review.

Very truly yours,

David H. Smith Nuclear Licensing Administrator

Attachment

cc: Resident Inspector - Byron J. Streeter - RIII

> 8410290151 841023 PDR ADDCK 05000454 PDR ADDCK 05000454

9374N