

SEP 7 1984

file

Hurley Medical Center
Department of Radiology
ATTN: John W. Dagenais
Assistant Director
One Hurley Plaza
Flint, MI 48502

Gentlemen:

Enclosed is Amendment No. 85 which renews your NRC License No. 21-00338-02 in its entirety per your application dated September 23, 1983. Several new conditions have been incorporated into your license; in view of this, please review your license carefully to assure that you understand and are in compliance with the terms and conditions contained therein.

Please note Condition 21 requires you to wipe test the final source containers of all licensed material received. Your application exempted certain packages from safe opening procedures. There are no exemptions from safe opening procedures in 10 CFR 20.205. Basically, 10 CFR 20.205 is divided into four parts:

- a. Paragraph (b) refers to monitoring for surface contamination upon receipt of packages containing licensed material. This paragraph contains exemption in your application.
- b. Paragraph (d) refers to procedures for safety opening packages containing licensed material. There are no exemptions in Paragraph (d).

Be advised that we have conditioned you to the Model ALARA Program contained in Regulatory Guide 10.8, October 1980 (License Condition 24) due to your request for increased action levels for Investigational Level I. If you desire to increase your levels above those listed in the Model Program, the Radiation Safety Committee should review the justification for and approve all revisions on the basis that the revisions are in keeping with good ALARA practices for that individual or group. The justification and approval for such increased levels should be documented and maintained on file for review by an NRC inspector.

In view of several new license conditions, please review your license carefully to assure that you understand and are in compliance with the terms and conditions contained therein. If you have questions, please contact our office by dialing (312) 790-5625.

8410310672 840907
NMS LIC30
21-00338-02 PDR

Please review the enclosed document carefully and be sure that you understand all conditions. You must conduct your program involving radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instruction and Reports to Workers; Inspection," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Possess radioactive material only in the quantity and form indicated in your license.
3. Use radioactive material only for the purpose(s) indicated in your license.
4. Notify NRC in writing of any change in mailing address.
5. Request and obtain appropriate amendment if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. Any amendment request should be accompanied by the appropriate fee specified in 10 CFR Part 170.
6. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.
7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material prior to your expiration date.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions and representations in your license application will result in enforcement action against you in accordance with the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

Hurley Medical Center

3

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed
Patricia J. Whiston
Materials Licensing Section

Enclosures:

1. Amendment No. 85
2. Regulatory Guide 10.8
October 1980

RIII

Whiston/as
08/31/84