



UNITED STATES
 ATOMIC ENERGY COMMISSION
 DIRECTORATE OF REGULATORY OPERATIONS
 REGION I
 631 PARK AVENUE
 KING OF PRUSSIA, PENNSYLVANIA 19406

AUG 4 1975

✓ Jersey Central Power and Light Company
 Attention: Mr. Ivan R. Finfrock, Jr.
 Vice President
 Madison Avenue at Punch Bowl Road
 Morristown, New Jersey 07960

License No. DPR-16
 Insp. No. 50-219/75-09
 Docket No. 50-219

Reference: Your letter dated June 27, 1975
 In response to our letter dated May 27, 1975

Gentlemen:

We note that in your letter and during a telephone call with Mr. Scalsky of your staff, you have indicated that linearity checks will be performed on your survey instruments only once every two years. This item (Deviation A) is being referred to our headquarters for further action.

With respect to Deviation B, based on the telephone conversation with Mr. Scalsky, we understand that a program is in effect for ensuring that survey instruments are calibrated quarterly. This program includes assigning responsibility for instrument calibration to an individual, and providing a system of check-off sheets and a tickler file. This program will be reviewed during a subsequent inspection.

Sincerely,

Paul R. Nelson, Chief
 Radiological and Environmental
 Protection Branch

cc: J. T. Carroll, Station Superintendent
 A. Z. Roisman, Counsel for Citizens Committee for
 Protection of the Environment

bcc:
 IE Chief, FS&EB
 IE:HQ (4)
 IE Files
 Central Mail & Files
 Directorate of Licensing (4)
 RS
 PDR
 Local PDR
 NSIC
 TIC

REG:I Reading Room
 ELD
 State of New Jersey

OFFICE	V.P.C.				
SURNAME	<i>WMP</i> Panzarino/jd	<i>PKP</i> Knapp	<i>Nelson</i> Nelson	<i>Phc for</i> O'Reilly	<i>J.H.</i> Brunner
DATE	7/30/75	7/30/75	7/31/75	8/12	7/3/75