



MINNESOTA VALLEY  
TESTING LABORATORIES, Inc.



PHONE (507) 354-8517

CENTER & GERMAN STREETS, NEW ULM, MINNESOTA 56073

January 7, 1985

RECEIVED BY LFMB

U.S. Nuclear Regulatory Commission  
799 Roosevelt Rd.  
Glen Ellyn, IL 60137

Attn: Materials Licensing Section

RE: Application to Amend License Number 22-17614-01

Date 1/28/85  
Log Jan 18  
By CAG  
Orig. To K/TH  
Action Compl. [initials]

Materials Licensing Section,

We wish to amend our current U.S.N.R.C. License number 22-17614-01; Docket/Reference No. 030-13064. Please amend item 8, page 1 of our current license to read "A. No single cell to exceed 15 millicuries". This will delete the limiting statement "Total 30 millicuries".

As a commercial testing laboratory we periodically need to upgrade our equipment to satisfy the needs of our clients. At this time our work load requires that we possess a minimum of 4 electron capture detectors. If item 8 is thus amended it will eliminate the need to apply for an amendment to our license each time we need to purchase another source.

Also, please amend item 12, page 1 to include the name John A. Hicks. See enclosed resume for education and work experience.

Enclosed please find a check for \$40.00 to cover the amendment fee. If there are any questions or need of further information regarding this application please call, (507) 354-8517.

Sincerely,

Applicant

Check No.

Amount

Type of Fee

Date Check Rec'd

Received By

017287440+  
017541320  
3 Paid  
1/28/85  
2/10/85  
[initials]

Jeffrey L. Henderson  
Laboratory Supervisor

Certifying Official,

Fred E. Day  
Chief Executive Officer

8503280155 850313  
REG LIC30  
22-17614-01 PDR

JLH/FED/s

enclosure

CONTROL NO. 78102

RECEIVED  
JAN 14 1985  
REGION III

John A. Hicks

Education - Graduated August, 1980 from the University of Washington, Seattle, Washington, with a Bachelor of Science Degree in Botany.

Work Experience - Has worked for Minnesota Valley Testing Laboratories, Inc., since July, 1983 in the Organic Chemistry Section. Worked for International Multifoods, Courtland, Minnesota from January, 1981 to July, 1983 as a Pharmaceutical Laboratory Technician.



MINNESOTA VALLEY  
TESTING LABORATORIES, Inc.



PHONE (507) 354-8517

CENTER & GERMAN STREETS, NEW ULM, MINNESOTA 56073

January 7, 1985

U.S. Nuclear Regulatory Commission  
799 Roosevelt Rd.  
Glen Ellyn, IL 60137

Attn: Materials Licensing Section

RE: Application to Amend License Number 22-17614-01

030-13064

Materials Licensing Section,

We wish to amend our current U.S.N.R.C. License number 22-17614-01; Docket/Reference No. 030-13064. Please amend item 8, page 1 of our current license to read "A. No single cell to exceed 15 millicuries". This will delete the limiting statement "Total 30 millicuries".

As a commercial testing laboratory we periodically need to upgrade our equipment to satisfy the needs of our clients. At this time our work load requires that we possess a minimum of 4 electron capture detectors. If item 8 is thus amended it will eliminate the need to apply for an amendment to our license each time we need to purchase another source.

Also, please amend item 12, page 1 to include the name John A. Hicks. See enclosed resume for education and work experience.

Enclosed please find a check for \$40.00 to cover the amendment fee. If there are any questions or need of further information regarding this application please call, (507) 354-8517.

Sincerely,

*Jeffrey L. Henderson*  
Jeffrey L. Henderson  
Laboratory Supervisor

Certifying Official,

*Fred E. Day* RECEIVED  
Fred E. Day  
Chief Executive Officer JAN 28 1985

JLH/FED/slu

enclosure

REGION III

CONTROL NO. 78204

JAN 28 1985

John A. Hicks

Education - Graduated August, 1980 from the University of Washington, Seattle, Washington, with a Bachelor of Science Degree in Botany.

Work Experience - Has worked for Minnesota Valley Testing Laboratories, Inc., since July, 1983 in the Organic Chemistry Section. Worked for International Multifoods, Courtland, Minnesota from January, 1981 to July, 1983 as a Pharmaceutical Laboratory Technician.



Steve Scarabin  
Randy Watts  
(919) 469-0259

DETECTOR SERVICE CENTER, INC.  
P.O. Box 1044  
Cary, NC 27511

CERTIFICATE OF INSPECTION

CUSTOMER Minnesota Valley Trading  
ADDRESS Center & Germania Street  
Germania  
New Ulm, Minn  
ATTENTION: John Hols

DETECTOR TYPE: N163 1st SERIAL NO. 1734A05076  
DATE OF TEST: 11/1/84 11/1/85

To Be Completed By DSC

AMOUNT OF CONTAMINATION:  $0.9 \times 10^{-4}$   $\mu\text{Ci}$   
WITHIN NORMAL LIMIT: YES NO  
NEXT WIPE TEST REQUIRED: 7-85

BY: Steve Scarabin INSPECTOR

CONTROL NO. 78204



Steve Scarabin  
Randy Watts  
(919) 469-0259

DETECTOR SERVICE CENTER, INC.  
P.O. Box 1044  
Cary, NC 27511

CERTIFICATE OF INSPECTION

CUSTOMER Monrovia Valley Testing Lab  
ADDRESS Center and German Streets  
Apex, NC 27501

ATTENTION: Jeff Henderson

DETECTOR TYPE: NI 63

SERIAL NO. 22 2343 A0 2782 #1

DATE OF TEST: 11/11/84

To Be Completed By DSC

AMOUNT OF CONTAMINATION:  $0.83 \times 10^{-4}$  pCi

WITHIN NORMAL LIMIT: YES NO

NEXT WIPE TEST REQUIRED: May 85

BY: Steve Scarabin INSPECTOR

CONTROL NO. 78204





Steve Scarabin  
Randy Watts  
(919) 469-0259

DETECTOR SERVICE CENTER, INC.  
P.O. Box 1044  
Cary, NC 27511

CERTIFICATE OF INSPECTION

CUSTOMER Minister Valley Testing Labs.

ADDRESS Carter & German Street  
New Hope, Mo 64632

ATTENTION: Jeff Henderson

DETECTOR TYPE: N163

SERIAL NO. 18803-60520

DATE OF TEST: 11/14/04

To Be Completed By DSC

AMOUNT OF CONTAMINATION:  $0.87 \times 10^{-4}$  pCi

WITHIN NORMAL LIMIT: (YES) NO

NEXT WIPE TEST REQUIRED: May 85

BY: Steve Scarabin INSPECTOR

CONTROL NO. 78204



Steve Scarabin  
Randy Watts  
(919) 469-0259

DETECTOR SERVICE CENTER, INC.  
P.O. Box 1044  
Cary, NC 27511

CERTIFICATE OF INSPECTION

CUSTOMER Community Valley Testing Labs  
ADDRESS Center & Gorman Streets  
New Village, NC 51073

ATTENTION: Jeff Henderson

DETECTOR TYPE: N163

SERIAL NO. 2343402782

#2

DATE OF TEST: 11/16/84

To Be Completed By DSC

AMOUNT OF CONTAMINATION:  $0.76 \times 10^{-4}$  pci

WITHIN NORMAL LIMIT: (YES) NO

NEXT WIPE TEST REQUIRED: May 85

BY: Steve Scarabin INSPECTOR

CONTROL NO. 78201