


## DOW CHEMICAL U.S.A.

MICHIGAN DIVISION January 14, 1985

J. G. Keppler United States Nuclear Regulatory Commission Region III 799 Roosevelt Road Glen Ellyn, Illinois 60137

MICHIGAN DIVISION MIDLAND, MICHIGAN 48640



MALFUNCTION OF THE WIDE-RANGE LOG CHANNEL DURING OPERATION OF THE DOW TRIGA RESEARCH REACTOR

On Monday, January 7, 1985, at about 3:30 P.M., Senior Reactor Operator T. J. Quinn was taking the TRIGA to a power level of 95 KW. As this was the first operation of the day, Mr. Quinn first took the reactor to a steady-state power level of 5 watts for a measurement of the core excess, which he found to be correct for the current configuration of the TRIGA. Mr. Quinn then initiated the rise of power to the desired level by pulling the shim rod while watching the period meter and the log power and linear power recorder traces. Within a few seconds he observed the period meter indicate that the period was decreasing, while at the same time the log power trace was levelling off and the linear power trace continued to rise. At this point, Mr. Quinn initiated a manual scram and secured the reactor. He discussed this event with Ward Rigot, the Assistant Reactor Supervisor. They concluded that the wide-range log channel had malfunctioned, possibly in the Campbell circuitry. Mr. Rigot reported this to me. I reviewed the experience and the recorder tracings with Mr. Quinn and determined that a malfunction had occurred. I notified D. L. Berry, Research Manager and chairman of the Reactor Operations Committee, and telephoned Mr. Kenneth Ridgeway at Region III, U.S.N.R.C., shortly after 4:30 P.M. that day.

Repairs and testing are in progress.

C. W. Kocher

Reactor Supervisor Analytical Laboratories

(1) Kole

1602 Building

cc: D. L. Berry, Chairman Reactor Operations Committee The Dow Chemical Co. 1602 Building Midland, MI 48674

L. W. Rampy, Chairman Radiation Safety Committee The Dow Chemical Co. 1803 Building Midland, MI 48674

8501290072 850114 PDR ADOCK 0500026

