

MF:M:KSD  
(34-05464-01)

APR 2 1975

Drs. Russell, Berkebile, and  
Associates, Inc.  
221 Eighth Street  
Lorain, Ohio 44052

Gentlemen:

Enclosed is a corrected copy of Amendment No. 08 to Byproduct Material License No. 34-05464-01 to authorize use at St. Joseph Hospital, 205 W. 20th Street, Lorain, Ohio, which was inadvertently omitted when the renewal amendment was issued. We have reviewed the survey report submitted by letter dated October 15, 1974, concerning the AECL Model 780 cobalt teletherapy unit. The report indicates that the requirements of Condition 18 of the license have been fulfilled.

Please note that Condition 18 requires that radiation surveys and associated tests be repeated and a report filed each time your teletherapy source is replaced. Note also that Condition 19 requires a radiation survey and report if you should make any changes in the shielding, location, or use of your teletherapy installation which could affect the radiation levels in surrounding areas.

Sincerely,

Kitty S. Dragonette  
Materials Branch  
Division of Materials and Fuel  
Cycle Facility Licensing

Enclosure:  
Amendment 08 to  
License 34-05464-01

B504260357 B50412  
REQ3 LIC30  
34-04474-02 PDR

CRESS OFFICE ▶	MF:M				
137004 acp					
SURNAME ▶	KDragonette:mlm				
3/75 DATE ▶	4/1/75				