

LICENSE NO: 20-12063-01

DOCKET NO. (s) 030-01945

File

PAGE 1 OF 10

ATTACHED

- ☒ Appendix A
☒ Appendix B
☒ Appendix C
☐ Memo

INSPECTION REPORT NO. 030-01945/BB-001

Mount Auburn Hospital
330 Mount Auburn Street
Cambridge, Massachusetts

LICENSEE CONTACT: Merrill Johnson, M.D. (RSO)

Telephone No: (617) 492-3500

LICENSE NO: 20-12063-01

CATEGORY G PRIORITY: 3

CATEGORY _____ PRIORITY: _____

CATEGORY _____ PRIORITY: _____

INSPECTION DATE (s): April 14, 1988

TYPE OF INSPECTION: ☐ SPECIAL ☐ ANNOUNCED
☒ ROUTINE ☒ UNANNOUNCED
☒ DAYSHIFT
☐ OTHER

SUMMARY OF FINDINGS AND ACTION

- ☐ NO NONCOMPLIANCE, CLEAR 591 ISSUED
☐ NO NONCOMPLIANCE, LETTER
☒ NONCOMPLIANCE, APPENDIX A

- ☒ ACTION ON PREVIOUS NONCOMPLIANCE, APPENDIX B
☐ NONCOMPLIANCE, 591 ISSUED
☒ SUPPLEMENTAL INFO, APPENDIX C

RECOMMENDATIONS
SEE BASIS IN APPENDIX C

- ☐ CHANGE CATEGORY TO: _____
☒ NEXT INSPECTION DATE: 04/91

☐ CHANGE PRIORITY TO: _____

PERSONS CONTACTED

* Dean DeMaster CNMT
Kathy Cardinale NMT
Dorothy Morine NMT
Kathy Sarles NMT
* Merrill Johnson M.D. RSO
Nancy Tava Clinical Specialist

Marna Magri Senior Chemistry Tech
* Gerald Foley Admin
* Francis Lynch President

INSPECTOR: Steven Constantine 4/27/88

APPROVED: John E. Glavin

* Present at Ext Interview

INSPECTION PLAN AND REPORT NUMBER 030-01945/85-001

Page 2 of 10

Plan Approved: _____

Date: 4/14/88

Licensee: Mt Auburn Hospital

License No. _____

20-12063-C1

Inspection Items	Scheduled for Inspection	Post Inspection Status	Module No.
Management Meeting - Entrance and Exit Interviews (Required)			30703B
Program Requirements, MC 2860 (Required)			78710B
Followup on Noncompliance and Deviations			92702B
Independent Inspection Effort (Required)			92706B
Transportation			86740B

(July 82)

INSPECTION REPORT NUMBER 030 - 01945/BB-001Page 3 of 1

78710B - Medical

AREAS INSPECTED AND FINDINGS

Licensee: Ht Auburn HospitalLicense No: 20-12063-01Ann. No. 4 of 7

INSPECTION ITEMS	CRITERIA	FINDING
1. <u>Organization</u>	Lic Cond _____	<u>C</u>
Structure of organization as described in requirements?	Dean DeMaster (CA-MT) reports any problems to Merrill Johnson MC (RSC) so that it may be reported to LSC.	
Scope of Program? Patient load?	4000	
NOTES & REMARKS:	3500-4000 patients / yr (15/M per day); M-F 7:00 AM - 5:00 PM 3 full time techs + CA-MT + 1 part time intern Standard Medical imaging and scheduling tests (unfrequently seen) 10-12 Thyroid uptake 600 Bone; 600 Stress; 100 Lung 350 Thyroid uptake 350 Bone out of 3500	
2. <u>Licensee Internal Audits</u>	Lic Cond _____	<u>NC</u>
Scope and frequency of audits as required?	NC	RSC met 4/85; 11/85; 5/86; 4/87; 10/87; 3/88
Conducted by appropriate persons?		required frequency is quarterly.
Records maintained?		Records maintained of the meetings. They should
Reviewed by management?		cover more material than they show. RSC
Deficiencies identified and corrected?		should be more involved in the program.
NOTES & REMARKS:	Deficiencies in RSC meeting frequency and who should attend was brought to their attention by consultant. Other problems brought up by Dean DeMaster.	
3. <u>Training and Qualification of Personnel</u>	Lic Cond _____	<u>C</u>
Training & retraining conducted as required?		Training of nurses and techs conducted
Written & oral exams conducted?		by consultant at required interval
Examination results reviewed by management?		covering topics necessary for a small
Instructions to workers per 19.12?	19.12	training. No test.
Authorized users? On license? Available in emergency?	Lic Cond _____	Also Chandler, Arons, Keltner and Johnson are associated with the hospital. Dr. Nolan is no longer at Ht. Auburn. Drs. Stranes, McKeen and Bannan are associated with Mass General.
NOTES & REMARKS:	Has used visiting physicians to cover for emergency work during the weekends otherwise covered (see item 15)	
4. <u>Radiation Protection Procedures</u>	Lic Cond _____	<u>C</u>
Procedures available and implemented?		Unit doses or labelling of syringe should used
Identify radiopharmaceutical and dose(s)?		to identify radiopharmaceutical. Do not
Cover handling of patients receiving therapeutic doses? Cover handling of cadavers?		used to determine dose. Hypothyroidism
Close out Surveys on Patients receiving temporary implants?	35.14 (b)(5)(v)	treated with I-123
Emergency procedures for spills, etc? Personnel understand procedures?		Personnel appeared knowledgeable of what to do in case of spills and incidents.
NOTES & REMARKS:		

AREAS INSPECTED AND FINDINGS

Licensee: Mt Auburn Hospital License No: 20-12663-01 Amendment No: 33

INSPECTION ITEM	CRITERIA	FINDING
5. <u>Use of Materials</u>		<u>N/C</u>
Procurement and use as required? Authorized form & route of administration?	35.14(b)	Receives I-123 capsules from Mallinckrodt, but doses and receipt tag from Syneos. Licensed material
Special tests (moly breakthrough, leak tests, etc) required?	35.14(b)(4)	
Inventory of brachytherapy sources? <u>N/A</u>	35.14(b)(5)	used for diagnostic tests and hyperthyroidism.
Dose calibration checks performed?		I-131 has not been used since 1982
Posting & labeling as required? <u>C</u>	20.203	Use of Xe-133, Tc-99m and I-123
NOTES & REMARKS: <u>At</u> <u>Only dose calibrator checks not performed</u> <u>on days of emergency use on Sat/Sun</u> <u>Quarterly accuracy performed C</u> <u>annual accuracy check C</u> <u>Geometry checks performed C.</u>		Sealed sources were last tested 12/80, and 10/87
6. <u>Storage of Materials</u>		<u>C</u>
Material secured in both restricted and unrestricted areas? Adequately?	20.207	Waste storage room is kept locked.
NOTES & REMARKS:		Hot lab was open but personnel had it under surveillance.
7. <u>Facilities</u>	Lic Cond <u>C</u>	
As described in lic cond or application?		
Any changes made? Adequacy?		No changes have been made in the facilities since the last amendment.
NOTES & REMARKS:		
8. <u>Instruments</u>	Lic Cond <u>C</u>	
Survey meters & instruments adequate for program?		Survey meters (2) were operable at time of inspection.
Instruments & meters operable? Calibrated? Calibration adequate?		Meters were calibrated at the required frequency.
NOTES & REMARKS:		

AREAS INSPECTED AND FINDINGS

Licensee: St. Auburn HospitalLicense No: 20-12063-01 Amendment No: 33

INSPECTION ITEMS	CRITERIA	FINDING
9. <u>Receipt and Transfer of Material</u>		<u>C</u>
Written procedures for pickup, receiving, opening packages?	20.205	Orders for material recorded and signed off on. Procedures for receiving after hours is followed.
Survey of packages when received? <u>C</u>	20.205(c)(1)	
Records of survey of packages? <u>C</u>	20.401(b)	
Transfer of materials proper? Transfer records maintained? <u>C</u>	30.41, 30.51	R1A Lab has standing orders for material but used 0-12
Authorized containers used? Shipping papers & package labels proper for packages on hand?	71.5	(~125 uCi/bottle) T314 (~100 uCi) T3 (~25 uCi) T3 (~25 uCi)
NOTES & REMARKS:		Surveys of packages when they are received by nuc measure is done by the tech and recorded. No transfers except used unit dose syringes.

10. Personnel Protection - External

Personnel monitoring controls adequate? Exposures minimized?	20.101, 20.202	<u>NC</u> Records show that badges are exchanged monthly. Those that are damaged or missing are assigned exposures. Appears inadequate.
Exposure records (NRC-4 or 5) maintained? Available for employee review?	20.102(b), 20.401(a)	Badly surveys not done during weekend when emergency and at home. Recorded in normal use days.
Surveys conducted? Adequate?	20.202	
Records of monitoring, surveys?	20.401	
Levels in unrestricted areas within limits? (Particularly around nuclear med. hot lab rooms of brachytherapy patients)	20.1, 20.105	
NOTES & REMARKS:		Background

11. Personnel Protection - Internal

Airborne concentrations in restricted areas? (Xe-133, patients treated with I-131)	20.103	<u>C</u> Airborne concentrations kept down because of bathroom increased they
Exposures to minors? <u>None</u>	20.104	For exchange not sufficient to clear air as per evidence.
Posting of airborne radioactivity areas? <u>C</u>	20.203(d)	No I-131 use.
Survey, monitoring bioassay adequate for airborne radioactivity, surface contamination? Records maintained?	20.201 20.401	Weekly surveys performed and recorded as required.
Procedures for use of Xe-133 followed? <u>C</u>		
NOTES & REMARKS:		

AREAS INSPECTED AND FINDINGS

Licensee: Rt Auburn Hospital License No: 2-12063-01 Amendment No: 33

INSPECTION ITEM	CRITERIA	FINDING
<u>12. Effluent Controls, Waste Disposal</u>		
Release of effluents controlled? (particularly Xe-133, radioiodine where used)	Xe trapped in charcoal filter 20.106, 20.33	RtA Lab maintains 4 SS syringe drains for waste and keeps them 20' apart. Does before surveying then goes to disposal
Waste disposals controlled?	designed when removed from 20.301, 20.303, 20.304, 20.305	to mound waste stream
Procedures, records maintained?	Negative 20.401, L1c Cond	time that dispose gloves and paper covers 1: half hour Records are kept
Surveys made? Adequate?	20.401	Unit dose syringes returned to manufacturer. 1) paper product, gloves to leaded trash can
NOTES & REMARKS: Waste is segregated as per the license 2) Empty vials kept segregated for decay. 3) To Kim, used syringes, Xe-133 in leaded container.		

13. Notifications and Reports

To individuals?	C 19.13	
Overexposures, excessive levels & concentrations, incidents?	20.403, 20.405	None
Personnel exposures and monitoring, termination reports?	20.407, 20.408	Monitoring C Termination report on request.
Theft or loss of licensed material?	20.402	None
Misadministrations?	35.41-35.45	ALC 27-637 reported
NOTES & REMARKS:		

14. Posting of Notices

Part 20, license & documents, procedures, notice of violations posted?	19.11(a)	Posted as required
NRC-3 posted?	19.11(c)	
NOTES & REMARKS:		


15. Other License Conditions

CC 15 Visiting physicians (not from Mass General) covering nuclear medicine scans without prior RSC approval. (Hamm approval, however) Aug 1, 1986, Sept 1, 1986, two instances.

AREAS INSPECTED AND FINDINGS

Licensee: MT Auburn Hospital License No: 20-12063-01 Amendment No: 33

INSPECTION ITEMS	CRITERIA	FINDING
------------------	----------	---------

16. Confirmatory Measurements17. Independent Inspection Effort

Gloves, eye bridges worn properly, lab coats, syringe shield used.

NC Visiting physicians w/o RSC permission are
 Drs. Allan Ashar, Rochelle Powsner of University Hospital of Boston
 Dr. Bernie Anderson St. Elizabeth.

18. Incidents and Events

Any incidents of misadministrations, contamination, etc., not otherwise covered by reports?

35.41 - 35.45
 20.402, 20.403, 20.405

None

INSPECTION REPORT NUMBER C30-01945/82-001

Page 8 of 10

APPENDIX A - DOCUMENTATION OF NONCOMPLIANCE

Licensee: Mount Auburn Hospital

License No: 20-1243-01

Reference	Basis for noncompliance
Report item <u>15</u> 10 CFR _____ Lic Cond <u>15</u> Type n/c <u>IV</u>	Visiting physicians did work involving nuclear medicine during weekends w/o RSC written approval.
Report item <u>5</u> 10 CFR <u>35.50</u> Lic Cond _____ Type n/c <u>IV</u>	Daily dose calibrator constancy checks not performed on days of emergency use on weekends.
Report item <u>10</u> 10 CFR <u>20.201</u> Lic Cond _____ Type n/c <u>IV</u>	Daily surveys not performed in nuclear medicine during days of emergency use on weekends.
Report item <u>2</u> 10 CFR <u>35.22</u> Lic Cond _____ Type n/c <u>IV</u>	RSC did not meet at the required quarterly interval.
Report item <u>5</u> 10 CFR <u>35.54</u> Lic Cond _____ Type n/c <u>IV</u>	Sealed sources for calibration of dose calibrator were not leak tested at the required six month frequency.
Report item _____ 10 CFR _____ Lic Cond _____ Type n/c _____	

APPENDIX B - LICENSEE ACTION ON PREVIOUS INSPECTION FINDINGS

Licensee: Mt. Auburn Hospital

License No: 20-12063-01

Identification and summary of action taken	Status
Report No: <u>HLER-87-137</u> ^{Diagnostic Misadministration} Type n/c: _____ Describe: <u>Syringe put in incorrect radiopharmaceutical on unit</u> Action taken: <u>No recurrence. Reported. Syringe was labelled Osteofit instead of Na pertechnetate</u>	<u>Low</u> OPEN <u>CLOSED</u>
Report No: <u>030-01945/85-001</u> Type n/c: <u>II</u> ^{Opening package surveys not conducted} Describe: <u>4/8, 5, 9, 10, 12/85</u> Action taken: <u>Opening package survey conducted on days packages were received.</u>	OPEN <u>CLOSED</u>
Report No: <u>030-01945/85-001</u> Type n/c: <u>III</u> ^{No daily surveys since 5/8/85} Describe: <u>5/8/85</u> Action taken: <u>Survey meter available (1) Daily surveys performed on weekdays but not on days of emergency use (Sat/Sun)</u>	OPEN <u>CLOSED</u>
Report No: <u>030-01945/85-001</u> Type n/c: <u>IV</u> ^{No weekly survey performed since 5/8/85} Describe: <u>5/8/85</u> Action taken: <u>Weekly ^{type} surveys performed</u>	OPEN <u>CLOSED</u>
Report No: <u>030-01945/85-001</u> Type n/c: <u>IV</u> ^{No surveys of packages returned to manufacturer} Describe: <u>to M.A.H.</u> Action taken: <u>When applicable, packages are surveyed prior to return to the manufacturer.</u>	OPEN <u>CLOSED</u>
Report No: _____ Type n/c: _____ Describe: _____ Action taken: _____	OPEN CLOSED

(July 82)

INSPECTION REPORT NUMBER 650-1945/82-001Page 10 of 10

APPENDIX C - SUPPLEMENTARY INFORMATION

Licensee: Mt Auburn HospitalLicense No: 20-12003-01☐ Uncorrected/repeated noncompliance☐ Unresolved items☐ Unusual occurrence, conditions, etc☒ Inspector's comments☐ Basis for change of Category or Priority

Most of the program is run by the CNMT (Aren 'ReMaster) There are 3 full-time technicians and one part-time student besides the CNMT. Nuclear Medicine is open from 7:00 AM - 5:00 PM but the WMT's are on 24 hour on-call service. There has been work during the weekends with none of the tests to Doctors from insure the proper operation of the instruments. Doctors from other hospitals cover for the hospital in order to read scans. They are authorized for the studies but no documentation of their approval was in the minutes of the R'SC.

13-14 scans are performed every day (^{~3200-4000} ~~3200-4000~~ patient yearly making it a mid-size ^{to large} program)

A breakdown of the studies:

800 Bone scans

600 Stress tests

400 Lung scans

350 Thyroid uptake

350 Thyroid scintiscans

10-12 Supperthyroid therapy but no ablations

Tries to always use unit doses. Keeps instant tech on hand in case of a "rush job" (Sydney)

Mallinckrodt supplies the I-123 capsules.

04/21/88

PAGE 2

U.S. NUCLEAR REGULATORY COMMISSION
REGION I
OPEN ITEMS TRACKING SYSTEM

DOCKET NUMBER: 30-01945

LN: 20 / 2063-01

REPORT NUMBER	STATUS	DATE OPEN	DATE CLOSED	REVIEWER NAME	OPTION	CLOSING REFER
1 87-039	CLOSED	01/19/87	04/14/88	DICK LADUN	LER	88-001
ITEM: DIAGNOSTIC MISADMINISTRATION						
2 85-001	CLOSED	06/13/85	04/14/88	JOHANSEN	VIOL	88-001
ITEM: NO OPENING PACKAGE SURVEYS ON JUNE 3, 4, 9, 10, 12 1985.						
3 85-001	OPEN	06/13/85	/ /	JOHANSEN	VIOL	-
ITEM: NO DAILY SURVEYS SINCE 5/29/85						
4 85-001	CLOSED	06/13/85	04/14/88	JOHANSEN	VIOL	88-001
ITEM: NO WEEKLY WIPE SURVEY SINCE 5/30/85.						
5 85-001	CLOSED	06/13/85	04/14/88	JOHANSEN	VIOL	88-001
ITEM: NO SWIPES OF RETURNED LICENSED MATERIAL TO NPI.						
6 88-001	OPEN	04/14/88	/ /	COURTEMANCHE	VIOL	-
ITEM: VISITING PHYS. DID WORK COVERING FOR THE HOSPITAL DURING THE WEEKENDS W/O RSC WRITTEN APPROVAL.						
7 88-001	OPEN	04/14/88	/ /	COURTEMANCHE	VIOL	-
ITEM: CALIBRATOR DAILY CHECKS NOT PERFORMED ON WEEKENDS WHEN THERE IS EMERGENCY WORK.						
8 88-001	OPEN	04/14/88	/ /	COURTEMANCHE	VIOL	-
ITEM: DAILY SURVEY OF NUCLEAR MEDICINE NOT PERFORMED ON DAYS OF EMERGENCY WORK DURING THE WEEKENDS.						
9 88-001	OPEN	04/14/88	/ /	COURTEMANCHE	VIOL	-
ITEM: RADIATION SAFETY COMMITTEE DID NOT MEET AT THE REQUIRED QUARTERLY FREQUENCY.						
10 88-001	OPEN	04/14/88	/ /	COURTEMANCHE	VIOL	-
ITEM: LEAK TESTS OF THE SEALED CALIBRATION SOURCES WERE NOT PERFORMED AT THE REQUIRED SIX MONTH INTERVALS.						