

AUG 17 1983

MID-ATLANTIC RADIATION PHYSICS, INC.

10408 TRUXTON ROAD • (301) 434-9283 • ADELPHI, MD 20783

1165

83

August 15, 1983

AUG 15 PM 2:27

Material Licensing Branch  
Division of Fuel Cycle and Material Safety  
Office of Nuclear Material Safety and Safeguards  
U. S. Nuclear Regulatory Commission  
Washington, D. C. 20555

ATTN: Mrs. Patricia Vacca

RE: Control #15292

Dear Mrs. Vacca:

With reference to our phone conversation last week in regards to the teletherapy licensing of Dr. Lawrence O. Snead, I'm enclosing the following supplemental information:

1. Floor plan and elevation plan (includes the door to the mechanical equipment room).
2. Photographs of the building which also show the location of the adjacent building. Note that the roof of the area over the mechanical equipment room is at a higher elevation than the roof over the therapy room. Unless you object, I would propose that the fencing over the teletherapy room be on three sides at one level and that the fourth side be at the level of the roof over the mechanical equipment room (signs to be posted on all four sides).
3. A letter from Dr. Snead that clarifies the changing of the name of the licensee and the fact that the new "named" licensee shall abide by all previous submissions. Note also that the letter specifically states that: (a) Emergency procedures similar to Para. 2 of Appendix G of the NRC Draft Regulatory Guide dated March 1982 will be posted and followed, (b) the corridor that includes Point B of my shielding calculations is a restricted area, and (c) keys to the mechanical equipment room will be kept by Dr. Snead and Dr. Burke only.

Should you desire more information, please contact the undersigned as soon as possible. Thank you.

Sincerely yours,

*Dwight W. Glenn*  
Dwight W. Glenn, Ph. D.

cc: Dr. Snead

8505020420 850412  
REG2 LIC30  
45-05897-01 PDR

COPY SENT REGION I

DR. SNEAD & ASSOCIATES, INC.  
1104 WEST FRANKLIN STREET  
RICHMOND, VIRGINIA 23220

August 12, 1983

Dear Dr. Glenn:

To supplement the request for an ammended license to cover use of Cobalt 60 teletherapy recently sent to NRC and assigned control #15292, the following information and statements are given.

Dr. A. W. Burke and I have formed a partnership for ownership and use of the teletherapy equipment and were advised to change the name of the licensee to facilatate operation as a partnership. The partnership agrees to all previous submissions of the old license conditions.

We agree to the emergency procedures outlined in sec.2 app. G of NRC guide (3-82).

The corridor which includes point B on the floor plan previously submitted will be considered a "restricted area".

The outside door to the mechanical equipment room is always locked and cannot be unlocked from outside the building. The inside door to this room is always locked and only the partners will have a key to this door.

Before operating the new unit, a barbed wire fence 3-4 feet high will be erected at the perimeters of the roof of the treatment room and a warning sign placed on each of the four sides of this fence .

Two copies of the floor plan for the entire building are enclosed with pictures of the building exterior including the adjacent building. The blue prints show the east property line 9 feet from the bulding, but the paved driveway measures more than 10 feet wide.

Sincerely,  
1104 West Franklin Street Assoc.  
*L. O. Snead, Jr.*  
L. O. Snead, Jr., Gen. Partner

LOSatt