APPENDIX

NOTICE OF VIOLATION

Community Hospital Torrington, Wyoming 82240 Docket No. 030-20277 License No. 49-23121-01

During an NRC inspection conducted on September 16, 1992, A violation of NRC requirements was identified. In accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C. the violation is listed below:

10 CFR 35.50(e)(3) requires, in part, that a licensee retain records of quarterly dose calibrator linearity tests for 3 years unless directed otherwise, and that the records include, in part, the signature of the Radiation Safety Officer.

Contrary to the above, as of September 16, 1992, the licensee's records of quarterly linearity tests of its dose calibrator performed between January 13, 1991, and May 26, 1992, did not include the signature of the Radiation Safety Officer.

This is a repeat violation.

This is a Severity level V violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, Community Hospital is hereby required to submit a written statement or explanation to the Regional Administrator, Region IV, with a copy to the U.S. Nuclear Regulatory Commission, ATTN: Document Control Desk, Washington, D.C. 20555, within 30 days of the date of the letter transmitting this Notice of Violation (Notice). This reply should be clearly marked as a "Reply to a Notice of Violation" and should include for each violation: (1) the reason for the violation, or, if contested, the basis for disputing the violation, (2) the corrective steps that have been taken and the results achieved, (3) the corrective steps that will be taken to avoid further violations, and (4) the date when full compliance will be achieved. If an adequate reply is not received within the time specified in this Notice, an order or demand for information may be issued as to why the license should not be modified, suspended, or revoked, or why such other action as may be proper should not be taken. Where good cause is shown, consideration will be given to extending the response time.

Dated at Arlington, Texas this 29thday of September 1992