

NRC FORM 313M (9-81) 10 CFR 35	U.S. NUCLEAR REGULATORY COMMISSION APPLICATION FOR MATERIALS LICENSE — MEDICAL	Approved by OMB 3150-0041
---	---	------------------------------

INSTRUCTIONS - Complete Items 1 through 26 if this is an initial application or an application for renewal of a license. Use supplemental sheets where necessary. Item 26 must be completed on all applications and signed. Retain one copy. Submit original and one copy of entire application to: Director, Office of Nuclear Materials Safety and Safeguards, U.S. Nuclear Regulatory Commission, Washington, D.C. 20555. Upon approval of this application, the applicant will receive a Materials License. An NRC Materials License is issued in accordance with the general requirements contained in Title 10, Code of Federal Regulations, Part 30, and the Licensee is subject to Title 10, Code of Federal Regulations, Parts 19, 20 and 35 and the license fee provision of Title 10, Code of Federal Regulations, Part 170. The license fee category should be stated in Item 26 and the appropriate fee enclosed.

1.a. NAME AND MAILING ADDRESS OF APPLICANT (institution, firm, clinic, physician, etc.) INCLUDE ZIP CODE Frankford-Mayfair Diagnostic Services 6200 Medical Building-Suite B 6200 Frankford Ave. Philadelphia, PA 19135 TELEPHONE NO.: AREA CODE (215) 535-1211	1.b. STREET ADDRESS(ES) AT WHICH RADIOACTIVE MATERIAL WILL BE USED (If different from 1.a.) INCLUDE ZIP CODE Same
---	---

2. PERSON TO CONTACT REGARDING THIS APPLICATION Walter L. Robinson, M.S. Lancaster, PA 17601 TELEPHONE NO.: AREA CODE (717) 397-2569	3. THIS IS AN APPLICATION FOR: (Check appropriate item) a. <input type="checkbox"/> NEW LICENSE b. <input type="checkbox"/> AMENDMENT TO LICENSE NO. _____ c. <input checked="" type="checkbox"/> RENEWAL OF LICENSE NO. 37-19272-01
--	--

4. INDIVIDUAL USERS (Name individuals who will use or directly supervise use of radioactive material. Complete Supplements A and B for each individual.) Daniel Pisano, M.D. R. Stokes, M.D. David Sabbar, M.D. Denis G. Zervos, M.D.	5. RADIATION SAFETY OFFICER (RSO) (Name of person designated as radiation safety officer. If other than individual user, complete resume of training and experience as in Supplement A.) Daniel Pisano, M.D., residing Walter L. Robinson, M.S. and Jack O'Sullivan, consulting
--	---

6.a. RADIOACTIVE MATERIAL FOR MEDICAL USE																			
RADIOACTIVE MATERIAL LISTED IN:	ITEMS DESIRED "X"	MAXIMUM POSSESSION LIMITS (In millicuries)	ADDITIONAL ITEMS: <div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> IODINE-131 AS IODIDE FOR TREATMENT OF HYPERTHYROIDISM PHOSPHORUS-32 AS SOLUBLE PHOSPHATE FOR TREATMENT OF POLYCYTHEMIA VERA, LEUKEMIA AND BONE METASTASES PHOSPHORUS-32 AS COLLOIDAL CHROMIC PHOSPHATE FOR INTRACAVITARY TREATMENT OF MALIGNANT EFFUSIONS. GOLD-198 AS COLLOID FOR INTRACAVITARY TREATMENT OF MALIGNANT EFFUSIONS. IODINE-131 AS IODIDE FOR TREATMENT OF THYROID CARCINOMA XENON-133 AS GAS OR GAS IN SALINE FOR BLOOD FLOW STUDIES AND PULMONARY FUNCTION STUDIES. </div> <div style="width: 48%;"> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 10%;">MARK ITEMS DESIRED "X"</th> <th style="width: 10%;">MAXIMUM POSSESSION LIMITS (In millicuries)</th> </tr> </thead> <tbody> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> </tbody> </table> </div> </div>	MARK ITEMS DESIRED "X"	MAXIMUM POSSESSION LIMITS (In millicuries)														
MARK ITEMS DESIRED "X"	MAXIMUM POSSESSION LIMITS (In millicuries)																		
10 CFR 31.11 FOR IN VITRO STUDIES																			
10 CFR 35.100, SCHEDULE A, GROUP I	X	AS NEEDED																	
10 CFR 35.100, SCHEDULE A, GROUP II	X	AS NEEDED																	
10 CFR 35.100, SCHEDULE A, GROUP III	X	2,000																	
10 CFR 35.100, SCHEDULE A, GROUP IV		AS NEEDED																	
10 CFR 35.100, SCHEDULE A, GROUP V		AS NEEDED																	
10 CFR 35.100, SCHEDULE A, GROUP VI																			

6.b. RADIOACTIVE MATERIAL FOR USES NOT LISTED IN ITEM 6.a. (Sealed sources up to 3 mCi used for calibration and reference standards are authorized under Section 35.14(d), 10 CFR Part 35, and NEED NOT BE LISTED.)			
ELEMENT AND MASS NUMBER	CHEMICAL AND/OR PHYSICAL FORM	MAXIMUM NUMBER OF MILLICURIES OF EACH FORM	DESCRIBE PURPOSE OF USE

INFORMATION REQUIRED FOR ITEMS 7 THROUGH 23

For Items 7 through 23, check the appropriate box(es) and submit a detailed description of all the requested information. Begin each item on a separate sheet. Identify the item number and the date of the application in the lower right corner of each page. If you indicate that an appendix to the medical licensing guide will be followed, do not submit the pages, but specify the revision number and date of the referenced guide: Regulatory Guide 10.8 Rev. 1 Date: October, 1980

Appendix O Alara Program confirmed

7. MEDICAL ISOTOPES COMMITTEE		15. GENERAL RULES FOR THE SAFE USE OF RADIOACTIVE MATERIAL (Check One)	
<input type="checkbox"/>	Names and Specialties Attached; and <u>See cover letter</u>	<input checked="" type="checkbox"/>	Appendix G Rules Followed; or
<input checked="" type="checkbox"/>	Duties as in Appendix B; or _____ (Check One)	<input type="checkbox"/>	Equivalent Rules Attached
<input type="checkbox"/>	Equivalent Duties Attached	16. EMERGENCY PROCEDURES (Check One)	
8. TRAINING AND EXPERIENCE		<input checked="" type="checkbox"/>	Appendix H Procedures Followed; or
<input type="checkbox"/>	Supplements A & B Attached for Each Individual User; and	<input type="checkbox"/>	Equivalent Procedures Attached
<input type="checkbox"/>	Supplement A Attached for RSO.	17. AREA SURVEY PROCEDURES (Check One)	
9. INSTRUMENTATION (Check One)		<input checked="" type="checkbox"/>	Appendix I Procedures Followed; or
<input type="checkbox"/>	Appendix C Form Attached; or	<input type="checkbox"/>	Equivalent Procedures Attached
<input checked="" type="checkbox"/>	List by Name and Model Number	18. WASTE DISPOSAL (Check One)	
10. CALIBRATION OF INSTRUMENTS		<input checked="" type="checkbox"/>	Appendix J Form Attached; or
<input checked="" type="checkbox"/>	Appendix D Procedures Followed for Survey Instruments; or <u>with cover letter</u> (Check One)	<input type="checkbox"/>	Equivalent Information Attached
<input type="checkbox"/>	Equivalent Procedures Attached; and	19. THERAPEUTIC USE OF RADIOPHARMACEUTICALS (Check One)	
<input checked="" type="checkbox"/>	Appendix D Procedures Followed for Dose Calibrator; or <u>with cover letter</u> (Check One)	<input checked="" type="checkbox"/>	Appendix K Procedures Followed; or
<input type="checkbox"/>	Equivalent Procedures Attached	<input type="checkbox"/>	Equivalent Procedures Attached
11. FACILITIES AND EQUIPMENT		20. THERAPEUTIC USE OF SEALED SOURCES	
<input checked="" type="checkbox"/>	Description and Diagram Attached	<input type="checkbox"/>	Detailed Information Attached; and
12. PERSONNEL TRAINING PROGRAM		<input type="checkbox"/>	Appendix L Procedures Followed; or _____ (Check One)
<input type="checkbox"/>	Description of Training Attached <u>(See cover letter same)</u>	<input type="checkbox"/>	Equivalent Procedures Attached
13. PROCEDURES FOR ORDERING AND RECEIVING RADIOACTIVE MATERIAL		21. PROCEDURES AND PRECAUTIONS FOR USE OF RADIOACTIVE GASES (e.g., Xenon - 133)	
<input checked="" type="checkbox"/>	Detailed Information Attached <u>as per appendix E</u>	<input checked="" type="checkbox"/>	Detailed Information Attached
14. PROCEDURES FOR SAFELY OPENING PACKAGES CONTAINING RADIOACTIVE MATERIALS (Check One)		22. PROCEDURES AND PRECAUTIONS FOR USE OF RADIOACTIVE MATERIAL IN ANIMALS	
<input checked="" type="checkbox"/>	Appendix F Procedures Followed; or	<input type="checkbox"/>	Detailed Information Attached
<input type="checkbox"/>	Equivalent Procedures Attached	23. PROCEDURES AND PRECAUTIONS FOR USE OF RADIOACTIVE MATERIAL SPECIFIED IN ITEM 6.b	
<input type="checkbox"/>		<input type="checkbox"/>	Detailed Information Attached

24. PERSONNEL MONITORING DEVICES

TYPE (Check appropriate box)		SUPPLIER	EXCHANGE FREQUENCY
a. WHOLE BODY	<input checked="" type="checkbox"/> FILM	R.S. Landauer & Sons	monthly
	<input type="checkbox"/> TLD		
	<input type="checkbox"/> OTHER (Specify)		
b. FINGER	<input type="checkbox"/> FILM		
	<input checked="" type="checkbox"/> TLD	R. S. Landauer & Sons	monthly
	<input type="checkbox"/> OTHER (Specify)		
c. WRIST	<input type="checkbox"/> FILM		
	<input type="checkbox"/> TLD		
	<input type="checkbox"/> OTHER (Specify)		

d. OTHER (Specify)


25. FOR PRIVATE PRACTICE APPLICANTS ONLY

a. HOSPITAL AGREEING TO ACCEPT PATIENTS CONTAINING RADIOACTIVE MATERIAL		b. ATTACH A COPY OF THE AGREEMENT LETTER SIGNED BY THE HOSPITAL ADMINISTRATOR.	
NAME OF HOSPITAL Frankford Hospital (Frankford Division)		c. WHEN REQUESTING THERAPY PROCEDURES, ATTACH A COPY OF RADIATION SAFETY PRECAUTIONS TO BE TAKEN AND LIST AVAILABLE RADIATION DETECTION INSTRUMENTS.	
MAILING ADDRESS Frankford & Wakeling Sts.			
CITY Philadelphia,	STATE Pa.		

26. CERTIFICATE

(This item must be completed by applicant)

The applicant and any official executing this certificate on behalf of the applicant named in Item 1a certify that this application is prepared in conformity with Title 10, Code of Federal Regulations, Parts 30 and 35, and that all information contained herein, including any supplements attached hereto, is true and correct to the best of our knowledge and belief.

a. LICENSE FEE REQUIRED (See Section 170.31, 10 CFR 170)		b. APPLICANT OR CERTIFYING OFFICIAL (Signature) 	
(1) LICENSE FEE CATEGORY: 7 c.		(1) NAME (Type of Print) Daniel Pisano, M.D.	
		(2) TITLE Director of Radiology & Nuclear Medicine	
(2) LICENSE FEE ENCLOSED: \$ 580.00		c. DATE	

APPENDIX C INSTRUMENTATION

1. Survey meters

- a. Manufacturer's name: Picker G.M.
 Manufacturer's model number: 655-186
 Number of instruments available: 1
 Minimum range: 0.01 mR/hr to 0.2 mR/hr
 Maximum range: 10.0 mR/hr to 2000 mR/hr
- b. Manufacturer's name: G.M. Monitor by Picker
 Manufacturer's model number: _____
 Number of instruments available: 1
 Minimum range: _____ mR/hr to _____ mR/hr c/m only
 Maximum range: _____ mR/hr to _____ mR/hr

2. Dose calibrator

Manufacturer's name: Picker Digital
 Manufacturer's model number: N/A
 Number of instruments available: 1

3. Instruments used for diagnostic procedures

Type of Instrument	Manufacturer's Name	Model No.
Gamma Camera	Picker	4/11/37
Thyroid Uptake Probe	Picker	N/A

4. Other (e.g., liquid scintillation counter, area monitor, velometer)

None

CALIBRATION OF SURVEY INSTRUMENTS

Check appropriate items.

- ☒ 1. Survey instruments will be calibrated at least annually and following repair.
- ☒ 2. Calibration will be performed at two points on each scale used for radiation protection purposes, i.e., at least up to 1 R/hr.

The two points will be approximately 1/3 and 2/3 of full scale. A survey instrument may be considered properly calibrated when the instrument readings are within ± 10 percent of the calculated or known values for each point checked. Readings within ± 20 percent are considered acceptable if a calibration chart, graph, or response factor is prepared, attached to the instrument, and used to interpret readings to within ± 10 percent. Also, when higher scales are not checked or calibrated, an appropriate precautionary note will be posted on the instrument.

3. Survey instruments will be calibrated

- ☐ a. By the manufacturer
- ☐ b. At the licensee's facility

(1) Calibration source

Manufacturer's name _____
 Model no. _____
 Activity in millicuries _____
 or
 Exposure rate at a specified distance _____
 Accuracy _____
 Traceability to primary standard _____

- ☒ (2) The calibration procedures in Section I of Appendix D will be used
 or
☐ (3) The step-by-step procedures, including radiation safety procedures, are attached.

☒ c. By a consultant or outside firm

- (1) Name Radiation Management Corp. or Nuclear Pharmacy, Inc.
Philadelphia, PA
- (2) Location Market St., Philadelphia, PA

(3) Procedures and sources

RMC NRC # 37-13129-01

☒ have been approved by NRC and are on file in License No. NPI NRC# - See Below

☐ have been approved by an Agreement State; a copy of the Agreement State license, the procedures, and a description of the sources are attached, and the consultant's report will contain the information on

☐ the attached "Certificate of Instrument Calibration."

☐ the consultant's reporting form as attached.

☐ are described in the attachment, and the consultant's report will contain the information on

☐ the attached "Certificate of Instrument Calibration."

☐ the consultant's reporting form as attached.

Nuclear Pharmacy, Inc.
 Van Nuys, Calif. #3822-70
 Houston, Texas #11-1911
 Approved by amendment to
 Philadelphia NRC #37-1846101MD

CALIBRATION OF DOSE CALIBRATOR

A. Sources Used for Linearity Test

(Check as appropriate)

_____ First elution from new Mo-99/Tc-99m generator

or

X Other* (specify) See S.C.L.C.M. Procedure attached

B. Sources Used for Instrument Accuracy and Constancy Tests

Radionuclide	Suggested Activity (mCi)	Activity (mCi)	Accuracy
Co-57	3-5	<u>1.0-3.0</u>	<u>+/- 1%</u>
Ba-133	0.1-0.5	<u>0.2</u>	<u>+/- 1%</u>
Cs-137	0.1-0.2	_____	_____
Ra-226	1-2	_____	_____
_____		_____	_____

C. X The procedures described in Section 2 of Appendix D will be used for calibration of the dose calibrator, except as stated in the cover letter.

or

_____. Equivalent procedures are attached.

*For licensees who are not authorized for Mo-99/Tc-99m generators, activity must be equivalent to the highest activity used.

Please amend our NRC License # _____ to include the following procedure for dose calibrator activity linearity check in lieu of our current procedure (that implied in NRC Reg. Guide 10.8 Appendix D., Part E, p. 10.8 - 27 to 28).

The new procedure is as follows:

The procedure is similar to the commercially available "Calcorp" sequential lead cylinder method. The variance in our method is that it is a Sequential Concentric Lead Cylinder Method (SCLCM). A lead disk is placed in the bottom of the radioisotope calibrator "dipper". Then sequentially 5 lead cylinders are concentrically placed over top of each other for 5 successive reductions in "apparent" assay readings. Then a lead disk is placed on top of the cylinders column, and a 6th reading is made. These values are reported on the enclosed form.

The lead effects a contrived degradation of radiation intensity measured over the useful range of assay values from total eluate activity to micro curie quantities. The effective "decay-equivalent" hours evolved empirically and averaged over 20 trials are as follows:

shield 1 = 18.93 hrs.	+ 3.9%
shield 2 = 27.88 hrs.	+ 2.4%
shield 3 = 35.87 hrs.	+ 3.1%
shield 4 = 43.05 hrs.	+ 1.3%
shield 5 = 47.88 hrs.	+ 2.8%
Top lid = 53.90 hrs.	+ 2.6%

The assayed values are divided by the original unshielded total eluate assay for a fraction (decimal).

The natural log of this decimal is divided by $0.1155 \times 0.693/6\text{hr.}$ to equal the "calculated effective time elapsed". This calculated value is compared with the empirically-determined values of "shielded-effective hours". A percent variance is calculated. The mean of the 6 variances is tabulated and compared for quarterly changes. Notation is made if the unit is autoranging or not. A plot is then made of assayed shielded values versus calculated elapsed time to assure acceptable linearity and $\pm 5\%$ precision.

The summation of error between run variability, between instrument variability, geometric mispositioning, and high to low original activity all add to $\pm 3.9\%$ for each measurement. Since this is within $\pm 5\%$, and consistent with our previous method, we submit this as our alternative method to comply with the intent of the NRC rules governing this check.

We are able to do the check quarterly or more frequently now with total eluate without loss of the use of that eluate clinically. We can do the check in less than 5 minutes, thus saving technological and professional time which is cost-effective and within A.L.A.R.A. constraints in our institution.

The method was refined by our consultant radiation physicist, Walter L. Robinson, M.S. A.B.S.N.M. (phone-(717) 397-2569), who can be contacted for further details on the method.

It is hoped that this information will be adequate for you to assess the equivalency of our proposed procedure to our existing one.

DOSE CALIBRATION ACTIVITY LINEARITY CHECK

Tc-99m

MILLICURIES OR % OF ORIGINAL ACTIVITY

100
90
80
70
60
50
40
30
20
10
9
8
7
6
5
4
3
2
1.0
0.9
0.8
0.7
0.6
0.5
0.4
0.3
0.2
0.1

Walter L. Robinson & Associates
Consultant Radiation Physicists
Diagnostic Imaging Specialists
2624 Spring Valley Rd.
Lancaster, PA 17601

	Assay	Calc. Hrs.	Shield Effect Hrs.	% Variance
Shield #0			0	
Shield # 1			18.93	
Shield # 2			27.88	
Shield # 3			35.87	
Shield # 4			43.05	
Shield # 5			47.88	
Top Lid			53.90	

Mean Variance: \bar{X} =

This $\leq + 5\%$ Therefore

Measurement Error: $\pm 3.9\%$

Mean Variance
is % of
last Quarter's
check. Therefore:

This unit is Autoranging
Yes No

Measured By: _____
Date: _____

0 3 6 9 12 15 18 21 24 30 36 42 48 54 60 66 72 84

APPENDIX J
WASTE DISPOSAL

Note: In view of the recent problems with shallow-land burial sites used by commercial waste disposal firms, NRC is encouraging its licensees to reduce the volume of wastes sent to these facilities. Important steps in volume reduction are to segregate radioactive from nonradioactive waste, to hold short-lived radioactive waste for decay in storage, and to release certain materials in the sanitary sewer in accordance with § 20.303 of 10 CFR Part 20.

1. Liquid waste will be disposed of (check as appropriate)

☐ In the sanitary sewer system in accordance with § 20.303 of 10 CFR Part 20.

☐ By commercial waste disposal service (see also Item 4 below).

☒ Other (specify): decayed to back-ground and returned to the radiopharmacy

2. Mo-99/Tc-99m generators will be (check as appropriate)

☒ Returned to the manufacturer for disposal.

☐ Held for decay* until radiation levels, as measured in a low background area with a low-level survey meter and with all shielding removed, have reached background levels. All radiation labels will be removed or obliterated, and the generators will be disposed of as normal trash.**

* Be sure that waste storage areas were described in Item 11 and that they are surveyed periodically (Item 17).

** These generators may contain long-lived radioisotopic contaminants. Therefore, the generator columns will be segregated so that they may be monitored separately to ensure decay to background levels prior to disposal.

☐ Disposed of by commercial waste disposal service (see also Item 4 below).

☐ Other (specify): _____

3. Other solid waste will be (check as appropriate)

☒ Held for decay* until radiation levels, as measured in a low background area with a low-level survey meter and with all shielding removed, have reached background levels. All radiation labels will be removed or obliterated, and the waste will be disposed of in normal trash.

☐ Disposed of by commercial waste disposal service (see also Item 4 below).

☒ Other (specify): decayed to background and returned to the radiopharmacy

4. The commercial waste disposal service used will be

(Name)

(City, State)

NRC/Agreement State License No. _____

BETWEEN: William O. Miller, Chief
License Fee Management Branch
Office of Administration

John E. Glenn, Chief
Nuclear Materials Section B
Division of Engineering and
Technical Programs

Address change

LICENSE FEE TRANSMITTAL

A. REGION I

1. APPLICATION ATTACHED

Applicant/Licensee: Frankford Mayfair Diagnostic Services

Application Dated: 2/1/85

Control No.: 03426

License No.: 37-19272-CU

2. FEE ATTACHED

Amount: \$ 580.00

Check No.: 2584

3. COMMENTS

Signed _____

Date _____

B. LICENSE FEE MANAGEMENT BRANCH

1. Fee Category and Amount: 7C \$580

2. Correct Fee Paid. Application may be processed for:

Amendment _____

Renewal ✓

License ✓

Signed Frances Brown

Date 2/19/85 409 2/20/85