NRC Form \$137 (2-52) 10 CFR 35	APPLIC	U.S. NUCL	ATERIALS	LICENSE -	- TELETHERAPY		Approved by OMB 3150-0081 Expres 1-31-85
INSTRUCTIONS - Com be or Safet Licer Licer Part	plete items 1 through 22 impleted on sil application y and Safeguards, U.S. N se. An NRC Materials Li see is subject to Title 10 170. The license fee cate	If this is an initial appli ins and signed. Retain o uclear Regulatory Com oense is issued in accor I, Code of Federal Regu gory should be stated in	lication or an appi one copy. Submit imission, Washingt dance with the ge ulations, Parts 19, in Item 22 and the	ication for renew original and one on, D.C. 20655. neral requirement 20, 21, and 35 a a appropriate fee	ai of a license. Use supplement copy of entire application to: (Upon approval of this applicati is contained in Title 10, Code of nd the license fee provision of enclosed.	al sheets where nec Director, Office of on, the applicant w of Federal Regulatic Title 1D, Code of I	essery, Item 22 must Nuclear Materials ill receive a Materials ons, Part 30, and the Federal Regulations,
LA NAME AND MAILING AD	DRESS OF APPLICANT fin	stitution, firm, clinic, phys	úcien, etc.)	1.5. STREET ADD	RESSIESI, ACTUAL LOCATION OF	TELETHERAPY SOU	RCE, INCLUDING
St. John H Department 7911 Detro Cleveland,	ospital of Radiolo it Avenue Ohio 4410	оду 12		Cobalt Radiat St. Jo 7911 D	t Room tion Therapy D ohn Hospital Detroit Avenue	epartmen (0 ³⁰	t 340
TELEPHONE AREA CODE 016) NUMBER 651-7000				Clevel	land, Ohio 44	102 0	0-
2 PERSON TO CONTACT RE Shirley Z.	Jucius, M.	S.		3. THIS IS AN AP	PLICATION FOR (Check appropr	iata itam)	
To source Lansa o	and 216 Jacob	se 651-7000	1 X1009	X c. RENEWA	L OF LICENSE NO34-008	69-04	
A INDIVIDUAL USERS (Nem materie/, Complete Supplem Paul S. La Hae K. Hon	e individuals who will use or ents A and 8 for sech indivis vik, M.D. g, M.D.	directly supervise use of r duel.)	adioactive	5. RADIATIONS If other then in The Ra Paul S from S	AFETY OFFICER (RSO) (Name of dividual user, complete resume of t adiation Safet S. Lavik, M.D. Shirley Z. Juc	y Office with co tius, M.S	rediction safety officer, a in Supplement Al r will be nsultation
6. SEALED SOURCES TO BE	USED IN TELETHERAPY	UNITS (Attach supplemen	tel pages il necessary	r)			
BYPRODUCT MATE (Element and Mass	RIAL NAME No./ MANU	OF SOURCE	SOU MODEL	RCE NUMBER	MAXIMUM ACTIVITY PER SOURCE	NUM	BER OF SOURCES
A Cobalt-60	balt-60 AECL C-I		46	6000		2	
8.							
C.							
7. TELETHERAPY UNITS A	tach supplemental pages, if i	necessary)			1		
NA	ME OF MANUFACTURER	(include description, if un	it is custom macle)		MC	DOEL NUMBER	
A. AECL					Theratron 7	80	
8.					RECEIVED BY	LEMB	
c	1	Applicant	Kon you		Day Blant	cep-	
B. USE (Artach supplementary	Davas, il nacassary	Check No.	20/04	1 11 1. 1.1	Jac. april.	Tott	
ABC	1	Amouni fee Cei	STON #3	1011	or file		
X HUMAN USE	ONLY 1	vee of iee X	in.		By the J.D.	Inter	
HUMAN AND (Specify on a	DOTHER USE	laie Check Rech	16/2A	184	Ong To	5.	
	P	leceived By	¢		Action Complety	·····	
		9. PE	RSONNEL MONITO	RING DEVICES	T		
(Check and/or complete as appropriate) (S (1) FILM BADGE - WHOLE BODY R. S. Landau)			SUPPLIER (Service Company)		EXCHANGE FREQUENCY		
			er Jr. a	and Co.	monthly		
(2) THERMOLUMINESC WHOLE BODY	ENT DOSIMETER (TLD)-						
(3) OTHER (Specify): 85071904	98 850626			1		JUN 1	8 1984
REG3 LIC 34-0086	30 7-01 PDR	1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	Cont	vol No.	76961		

For Items 10 through 21, check the appropriate box(es) and submit a detailed description of all the requested information. Begin each item on a separate sheet. Identify the item number and the date of the application in the lower right corner of each page. If you indicate that an appendix to the teletherapy licensing guide will be followed, do not submit the pages, but specify the revision number and date of the referenced guide. Regulatory Guide 10.

Rev. Date March 1982

1

-							
10	MEDICAL ISOTOPE COMMITTEE	15.	BEAM STOPS				
x	Names and specialties attuched, and (check one)		Description of stops used to restrict beam orientation attached				
x	a. Duties as in Appendix A, or	16.	SHIELDING EVALUATION				
	b. Equivalent duties attached.		Evaluation of proposed shielding attached.				
11	TRAINING AND EXPERIENCE	17.	OPERATING AND EMERGENCY PROCEDURES				
	a. Supplements A & B attached for each individual user, and	X	a. Description of operating procedures attached, and				
	b. Supplement A attached for RSO.	X	b. Copy of emergency procedures attached.				
12	INSTRUMENTATION (check one)	18.	INSTRUCTION OF PERSONNEL (check one)				
х	a. Appendix C form attached, or	X	a. Training program and schedule in Appendix H-followed, or				
	b. List manufacturer's name and model number.		b. Description of instruction program for employees attached.				
13	CALIBRATION OF INSTRUMENTS (check one)	19.	LEAK TESTS OF SEALED SOURCES				
	a. Appendix D, Part 2 procedures followed for instrumentation calibration, or	x	Description of leak-test procedures attached				
х	 Description of sources, calibration frequency and equivalent procedures attached. 	20	QUALIFIED EXPERT (Use only if the individual fails to meet 10 CFR 35.24 requirements.)				
14	FACILITIES AND EQUIPMENT		Statement of qualifications of the expert who will perform teletherapy calibrations attached.				
	a. Description and drawing of facilities attached; and	21.	ALARA PROGRAM (check one)				
x	b. Description of patient viewing and communicating systems attached; and	x	ALARA Program as in Appendix I, or				
	c. Description of area safeguards attached.		Equivalent ALARA Program attached.				
			Check (La Ade-100 - 20 Ade-100 - 20 Addate				
	22. CER (This item must be con	TIFIC mplete	CATE In applicant (
The	applicant and any official executing this certificate on behalf of the applicant name de of Federal Regulations, Parts 30 and 35, and that all information contained hereir pwiedge and belief.	ed in I n, incl	Item 1a certifies that this application is prepared in conformity with Title 10. ūding supplements attached hereto, is true and correct to the best of our				
	a. LICENSE FEE REQUIRED (See section 170.31, 10 CFA 120)	P Si	APPLICANT OR CERTIFYING OFFICIAL (Signature) Restandight Com Karam, C.S.H. WAME (Type or prov) Ister Judith Ann Karam, C.S.A.				
0	Teletherapy President						
12) LICENSE FEE ENCLOSED							
\$	270.00	Ju	ine [시, 1984				
WA	RNING: 18 U.S.C. Section 1001; Act of June 25, 1948, 62 Stat. 749, makes it a or agency of the United States as to any matter within its jurisdiction.	crimin	al offense to make a willfully false statement or representation to any department				

10. Radiation Safety Committee Paul S. Lavik, M.D. William J. Fayen, M.D. Shirley Z. Jucius, M.S. John Wysocki Mary Lou Vogelberger, R.N. Nursing Rep.

1.0

Ø.

1

Radiotherapist Nuclear Medicine Radiation Physicist Administrative Rep.

F

Duties as in Appendix A and Sect. 35.11 dated August 25, 1982

Item #10 June 14, 1984

Control N- 76961

11. Training and Experience

Paul S. Lavik, M.D. Hae-Kyung Hong, M.D.

See Application dated Dec. 15, 1981 April 8, 1982 June 14, 1982

Paul S. Lavik, M.D. is the Radiation Safety Officer with consultation from Shirley Z. Jucius, M.S. (Credentials submitted December 15, 1981).

Item # 11 June , 1984

APPENDIX C INSTRUMENTATION

1. Survey meters

2.

3.

a.	Manufacturer's 1	dine.		and the second se			and the second se	
	Manufacturer's r	node1	number:	4	70A		In a second	gering all a
	Number of instru	uments	availat	ole:		1	$(x,y) \in B_{1}^{1,1}(Y)$	14 A (154) 23.
	Minimum range:	(0	mr/hr	to	3	and Filmer	mr/h
	Maximum range:	0)	mr/hr	to	1000	1. 1. 1. 1. 1.	mr/hi
×						STALL.		
b.	Manufacturer's	name:			-	а		
	Manufacturer's	model	number:			15		
1.	Number of instr	uments	availa	ble:				
	Ranges:				-	1.1.1.1	<u></u>	
	Minimum range:			mr/hr	to.			mr/hi
	Maximum range:			mr/hr	to .			mr/hi
Manu Manu Numb	ufacturer's name: ufacturer's model per of instrument	Nuc numbe s avai	er: Mod	ssociat el 05-4 1	es 37	Prima S/N	lert 35 N 21092	1941 - 20 1941 - 20 1941 - 20 1944 - 20 1944 - 20
Manu Manu Numb Bacl	ufacturer's name: ufacturer's model ber of instrument kup Battery Power	Nuc numbe s avai Suppl	er: <u>Mod</u> lable: ly: Y	<u>ssociat</u> el 05-4 / 1 es	es 37	Prima S/I No X	lert 35 N 21092	
Manu Manu Numb Bacl	ufacturer's name: ufacturer's model ber of instrument kup Battery Power imetry System	Nuc numbe s avai Suppl	er: <u>Mod</u> Hable: Hy: Y	<u>ssociat</u> el 05-4 / 1 es	<u>es</u> 37	Prima S/I No X	lert 35 N 21092	
Manı Manı Numt Bacl Dos	ufacturer's name: ufacturer's model ber of instrument kup Battery Power imetry System Electrometer	Nuc numbe s avai Suppl	er: <u>Mod</u> Nable: Ny: Y	<u>ssociat</u> el 05-4 1 es	<u>es</u> 37	Prima S/I No X	lert 35 N 21092	
Manu Manu Numb Bacl Dos a.	ufacturer's name: ufacturer's model ber of instrument kup Battery Power imetry System Electrometer Manufacturer's	Nuc numbe s avai Suppl	lear A er: Mod llable: ly: Y	ssociat el 05-4 / 1 es	<u>es</u> 37	Prima S/I No X	lert 35 N 21092	pintec
Manu Manu Numb Bacl Dost a.	ufacturer's name: ufacturer's model ber of instrument kup Battery Power imetry System Electrometer Manufacturer's Manufacturer's	Nuc numbe s avai Suppl name: model	vinumber:	ssociat el 05-4 l es s 	<u>es</u> 37	Prima S/N No X	lert 35 N 21092	pintec 192
Manu Manu Numb Bacl Dos a.	ufacturer's name: ufacturer's model ber of instrument kup Battery Power imetry System Electrometer Manufacturer's Manufacturer's Probes	Nuc numbe s avai Suppl name: model	<u>lear A</u> er: <u>Mod</u> ilable: ly: Y <u>Vi</u> number:	ssociat el 05-4 / 1 es 	<u>es</u> 37	Prima S/N No X	lert_35 N_21092	pintec 192
Manu Manu Numb Bacl Dost a.	ufacturer's name: ufacturer's model ber of instrument kup Battery Power imetry System Electrometer Manufacturer's Manufacturer's Probes Manufacturer's	Nuc numbe s avai Suppl name: model	lear A er: Mod lable: ly: Y number:	ssociat el 05-4 / 1 es 	es 37	Prima S/I No X	Ca	pintec 192
Manu Manu Numb Bacl Dost a.	ufacturer's name: ufacturer's model ber of instrument kup Battery Power imetry System Electrometer Manufacturer's Probes Manufacturer's Manufacturer's	Nuc numbe s avai Suppl name: model	number:	ssociat el 05-4 les ctoreen 570 Victore 131	<u>es</u> 37	Prima S/I No X	Lert 35 N 21092 Ca Vict	pintec 192 oreen 0-5
Manu Manu Numb Bacl Dos a.	ufacturer's name: ufacturer's model ber of instrument kup Battery Power imetry System Electrometer Manufacturer's Probes Manufacturer's Manufacturer's Number of probe	Nuc numbe s avai Suppl name: model name: model	ilable: Ny: Y Number:	ssociat el 05-4 les 	es 37	Prima S/N No X	Lert 35 N 21092 Ca Vict 7	pintec 192 oreen 0-5 1

4. Other (use additional pages)

Control No. 76961

the state of same shart the

24 Item #12 June | 4, 1984 b. Probes

· · · · · ·

1

Capintec PR-06C	Capintec PS-033	Capintec PR-16C	Capintec PM-30	
1	1	1	1	
2000R	2000R	2000R	20R	

Item #12 June 14, 1984

Control No. 76961

13. Calibration of Instruments

·

Survey Meter calibrated annually by:

Nuclear Medicine Associates, Inc. 9700 Garfield Blvd. Cleveland, Ohio 44125

Last Calibration May 1984

Electrometers and Probes calibrated bi-annually by:

University of Wisconsin Radiation Calibration Services Department of Medical Physics 1530 Medical Sciences Center 1300 University Avenue Madison, Wisconsin 53706

Last Calibration November 1983

Item #13 June 14, 1984

14. Facilities and Equipment

· · · · ·

*

b. Patient Viewing Closed Circuit TV system Window in door and mirror

> Communicating System Talk-A-Phone intercom

> > Item #14 June 14, 1984

15. Beam Stops

. . .

.

*

Refer to survey submitted January 11, 1984 and supplementary information submitted April 4, 1984

Item #15 June 14, 1984

17. Operating Procedures

1. Safety Device Checks

Safety devices should be checked periodically to ensure that they are operating properly. The checks will be made monthly by the Radiation Physicist as part of the monthly spot check measurements. The results of the measurements will be recorded in the notebook entitled "Cobalt Data Book". Any malfunctions or defects will be promptly corrected. Malfunctions of the timer, collimating system or source transport system will result in cessation of patient treatments until the malfunction is repaired. The Primalert, beam-on monitor and all control panel lights will be checked daily for proper functioning and recorded on a check list. Any malfunction will be promptly corrected. If the Primalert should malfunction, personnel will be required to enter the room with the Victoreen Panoramic Survey Meter.

2. Personnel Dosimetry

All teletherapy personnel will wear film badges, to be placed between the shoulders and the hips. In the event that a person receives or suspects that he or she has received a high exposure, the Radiation Physicist should be notified immediately. The film badge of the affected r person will be processed immediately. When not being worn, the film badges will be stored in an area which does not receive any radiation.

3. Procedures for Securing Teletherapy Unit

When the teletherapy unit is unattended, the control panel will be locked. Doors to the treatment area will be locked outside of normal working hours.

4. Instrument Calibration

The Victoreen Panoramic Survey Meter will be calibrated annually by Nuclear Medicine Associates, 9700 Garfield Blvd., Cleveland, Ohio 44125. The meter will be hand delivered and picked up after calibration. The Victoreen 570 and Capintec 192 dosimetry systems will be calibrated biannually by an approved calibration laboratory. The instruments will be secured in their storage cases using the specially designed packing material. The instruments will be shipped via UPS in their specially designed shipping containers. The instruments will be picked up after calibration and hand carried back to avoid the possibility of damage to the instruments.

> Item #17 June /4, 1984

5. Full Calibration of Teletherapy Units

The annual teletherapy source calibration will be performed by the Radiation Physicist. The procedures for full calibration of teletherapy units specified in Section 35.21 of 10 CFR Part 35 will be followed.

6. Monthly Spot-Check Measurements of Teletherapy Units

The monthly spot check of the teletherapy unit will be performed by the Radiation Physicist. The procedures for monthly spot check measurements of teletherapy units specified in Section 35.22 of 10CFR Part 35 will be followed.

7. Leak Testing

The leak test of the Cobalt source will be performed by the Radiation Physicist utilizing a Mark V leak test kit supplied by Applied Health Physics, Inc., 2986 Industrial Blvd, Box 197, Bethel Park, Pa. 15102. Instructions for use of the Mark V leak test kit supplied by Applied Health Physics, Inc. will be followed.

8. Record keeping

The daily treatment records will be maintained by personnel during the course of their work. Film badge records will be maintained by the Chief Technologist in the Radiology Department. Records of the annual source calibrations, monthly spot check measurements and leak test results will be maintained by the Radiation Physicist.

9. Emergency Procedures

The emergency procedures to be followed in the event of an emergency or other unusual occurrence are posted at the control console of the teletherapy unit. They are as follows:

IN THE EVENT OF EQUIPMENT FAILURE RESULTING IN THE SOURCE REMAINING "ON"

IMMEDIATELY ENTER THE TREATMENT ROOM AND:

- 1. CLOSE ADJUSTABLE FIELD DEFINER
- 2. REMOVE PATIENT FROM TREATMENT ROOM
- 3. MANUALLY RETURN SOURCE TO "OFF" POSITION FOLLOWING POSTED PROCEDURE

Item #17 June /4, 1984 4. TELEPHONE:

* *

PAUL S. LAVIK, M.D. 473-4161 651-7000 Ext. 1007

AND

AECL MEDICAL ELK GROVE, ILLINOIS (312) 593-3242

CAUTION: STAY OUT OF THE DIRECT BEAM AT ALL TIMES

IN THE EVENT OF A POWER FAILURE, THE ROOM SHOULD BE ENTERED USING THE VICTOREEN PANORAMIC SURVEY METER TO ENSURE THAT THE SOURCE HAS RETURNED TO THE "OFF" POSITION

IF THE DRAWER FAILS TO CLOSE, PROCEED AS FOLLOWS:

- 1. Remove the patient from the treatment room.
- 2. The drawer return emergency T-bar, which is supplied with the unit and located at the control station, inside the door, should be placed over the beam condition indicating rod. Forward pressure on the source drawer with the T-bar will push the drawer backwards and into the safe position.

NOTE:

 The amber colored portion of the emergency T-bar must be entirely inside the front head cover before the source is in the fully safe position. This will reduce external radiation fields to normal levels and allow repairs to be made to the drawer. The front portion of the T-bar is painted red and the source can be considered relatively safe if

no red marking appears outside the front cover.

FOR EMERGENCY SERVICE TELEPHONE:

(312) 593-3242 AECL MEDICAL ELK GROVE, ILL.

> Item #17 June 14, 1984

10. Procedures for Notifying Proper Persons in the Event of an Accident or Unusual Occurrence

The emergency prodecures shown above specify the appropriate persons to be notified in the event of an accident or unusual occurrence. In the case of a misadministration of dose, that is the treatment dose differing from the final prescribed total treatment dose by more than 10 percent, the procedures in Section 35.42 of 10 CFR Part 35 should be followed.

> Item #17 June /4, 1984

> > Control No. 76961

19. Laak Tests of Sealed Sources

.

. . . .

The Cobalt source is leak tested utilizing a Mark V leak test kit supplied by Applied Health Physics Inc. The attached instructions are followed.

> Item #19 June 14, 1984

pplied HEALTH PHYSICS inc.

2986 Industrial Blvd. 2 Box 197 · Bethel Park, Pa. 15102 · Phone 412 · 563-2242

INSTRUCTIONS FOR USE OF MARK V LEAR TEST KIT ON SEALED SOURCES

These procedures are to be followed by the individual trained and authorized by the licensee's RSO to employ the Mark V Leak Test Kit for leak testing sealed sources of radioisotopes. Should any question arise concerning proper use of the Kit, contact Applied Health Physics, Inc.

1. Pre-Test Procedures

In preparing the Mark V Leak Test Kit, follow these simple procedures:

a. Remove the plastic cap with its cotton swab insert from the plastic test tube. Add a few drops of water to dissolve the powdered wetting agent in the tube. Slightly dampen the swab's cotton tip with the wetting agent solution, and discard any unused solution that may remain in the tube.

Return the prepared swab to the test tube.

- b. Complete the information required on the self-sticking, circular leak test label which is included in the Kit, and securely attach to the midsection of the test tube.
- c. Obtain a remote handling device, such as an AHP Protecta-Holder, which will be used in manipulating the swab stick during the actual testing procedure.
- d. Carry out final preparation for all radiation protection measures that must be employed.

2. General Testing Procedures

The following general testing procedures should be used on sealed sources:

- a. Grasp the cap or the bare end of its swab stick with the AHP Protecta-Holder or other suitable remote handling device.
- b. Carefully, but firmly, wipe the dampened tip of the cotton swab over surface areas of the sealed source. With certain sealed sources that are located or permanently used in special equipment, it is only necessary to wipe surfaces of the mounting or storage device on which radioisotope contamination might be expected to accumulate.
- c. Immediately following the wiping procedure, dispense with the remote handling device and securely replace the cap and its swab insert in the labeled plastic test tube. Avoid touching the cotton tip to the body or other objects.

AHP:130A Rev. 9/81

Control No. 7696 1

Item #19 June /4, 1984

CONSULTATION . SERVICES PRODUCTS RADIATION APPLICATIONS

J. Post-Test Procedures

Pursuant to completion of the leak test, these steps must be taken:

- a. Complete the Mark V Leak Test Data Form in a legible fashion. This form must be signed by the individual who performed the sealed source leak test.
- b. Enclose the Data Form and the sealed plastic test tube in the mailing box, and seal the box. Fill in the proper return address on the Applied Health Physics, Inc.'s shipping label and securely attach to the box.
- c. Monitor all external surfaces of the mailing box with a calibrated survey meter, such as a Geiger-Muller meter with an end-window probe detector. Post Office Department regulations require that radiation levels at any surface of the box must be less than 10 milliroentgens for 24 hours; i.e., an average of approximately 0.4 milliroentgens per hour.

If results of the survey meet these requirements, proceed with mailing the Mark V Leak Test Kit to Applied Health Physics, Inc. Should the survey indicate that any surface of the box has a dose-rate greater than 0.4 milliroentgens per hour, immediately notify Applied Health Physics, Inc. by telephone.

AHP: 130A Rev. 9/81

Item #19 June /4 , 1984