



COMMUNITY  
HEALTH  
CENTER

Wright Street  
Palmer, Ma. 01069  
413/283-7651

WITH SATELLITE  
HEALTH CENTERS  
AT

10 South Main Street  
Belchertown, Ma. 01007  
413/323-5118

121 Main Street  
Monson, Ma. 01057  
413/267-5181

3 Crane Park Drive  
Wilbraham, Ma. 01095  
413/596-3455

# WING MEMORIAL HOSPITAL

MS 16  
KO

September 9, 1985

License No. 20-15280-01

Mail Control No.  
104278

United States  
Nuclear Regulatory Commission  
Region 1  
631 Park Avenue  
King of Prussia, Penn. 19406

Dear Sirs;

In response to our telephone conversation on August 29, 1985, find the enclosed information on procedures, standards and frequency of calibration of survey meters.

We are continuing action on the preceptor statements, supplements A & B for Dr. Nicholas Spencer.

Respectfully submitted:

*Richard A. Scheffer*  
Mr. Richard Scheffer  
President

Wing Memorial Hospital Corp.  
Wright Street  
Palmer, Mass. 01069

8511220111 851011  
REQ1 LIC30  
20-15280-01 PDR

SEP 24 PM 2:55

RECEIVED-REGION 1

AFFILIATED WITH THE UNIVERSITY OF MASSACHUSETTS MEDICAL CENTER

04278  
"OFFICIAL RECORD COPY"

ML10

SEP 24 1985



# WING MEMORIAL HOSPITAL

CODE NO.	PAGE 1 OF 3
EFFECTIVE DATE SEPTEMBER 9, 1985	
SUPERSEDES	
PERSONNEL APPROVAL	
HOSPITAL ADMINISTRATOR	

☐ POLICY ☒ PROCEDURES ☐ JOB DESCRIPTION

SUBJECT: METHOD FOR CALIBRATION OF SURVEY METERS

## PROCEDURES, STANDARDS, FREQUENCY

- A. Calibration of survey meters shall be performed with radionuclide sources.
1. The source shall be approximate point sources.
  2. The source activities or exposure rates at given distances shall be traceable by documented measurements to a standard source certified within 5% accuracy to the U.S. National Bureau of Standards (NBS) calibrations.
  3. The frequency shall be at least annually and after servicing.
  4. Each scale of the instrument shall be calibrated at least at two points located at approximately 1/3 and 2/3 of full scale.
  5. The exposure rate measured by the instrument shall differ from the true exposure rate by less than 10 % at the two points on each scale (read appropriate section of the instrument manual to determine how to make necessary adjustments to bring instrument into calibration). Readings within +/- 20% will be considered acceptable if a calibration chart, graph, or response factor is prepared, attached to the instrument, and used to interpret meter readings to within 10% for radiation protection purposes.
  6. A Cesium 137 source will be used. ( 41.14 mr/hr at one meter on 1/5/84, +/- 3.0%).
- B. A reference check source of long half-life, e.g. Cs137 or Ra D and E. shall also be read at the time of the above calibration or as soon as the instrument is received from the calibration laboratory. The reading shall be taken with the check source placed in specific geometry relative to the detector. A reading of this reference check source should be taken:
1. Before each use and after each survey to ensure that the instrument was operating during the survey.

2. After each maintenance and/or battery change.
3. At least quarterly.

If any reading with the same geometry is not within  $\pm 20\%$  of the reading measured immediately after calibration, the instrument should be recalibrated (see item A).

- C. The instrument will be calibrated at lower energies if its response is energy dependent and if the instrument is to be used for quantitative measurements in the Xe-133 or Tc-99m energy ranges.

The calibration will be done either:

1. As a relative intercomparison with a energy independent instrument and uncalibrated radionuclides.
  2. Alternatively, the manufacturer's energy response curve(s) may be used to correct instrument readings appropriately when lower-energy radiation is monitored.
- D. Records of the above items A, B-2, B-3, and C will be maintained.
  - E. Use of Inverse Square Law and the Radioactive Decay Law
    1. A calibrated source will have a calibration certificate giving its exposure rate at a given distance, or its activity, measured, on a specific date by the manufacturer or NBS.
      - a. The inverse square law may be used with any point source to calculate the exposure rate at other distances.
      - b. The radioactive decay law may be used to calculate the exposure rates or source activities at times other than the calibration date.
    2. INVERSE SQUARE LAW

Consider a "point" source of radiation at position S, as shown in Figure D-1. Then, the relationship between the exposure rates  $R_1$  and  $R_2$  at detector positions P1 and P2 which are at distances  $D_1$  and  $D_2$  from S respectively, is given by the following equation:

$$R_2 = \frac{D_1^2}{D_2^2} \times R_1$$

where  $R_1$  and  $R_2$  are exposure rates in the same units (eg. mR/hr, R/hr), and  $D_1$  and  $D_2$  are the distances in figure D-1 in the same units (eg. m, cm, ft.).




FIGURE D-1

### 3. RADIOACTIVE DECAY LAW

Exposure rate  $t$  units of time after specified calibration date

$$R_t = R_o \times e^{-\left[\frac{0.693}{T_{1/2}} \times t\right]}$$

WHERE:

$R_o$ and $R_t$	are in the same units (eg mR/hr or R/hr).
$R_o$	is exposure rate on the specific calibration date.
$R_t$	is exposure rate $t$ units of time later.
$T_{1/2}$ and $t$	are in the same units ( years months, days, etc.).
$T_{1/2}$	is radionuclide half-life
$t$	is number of unit of time elapsed between calibration and present time.

# CERTIFICATE OF INSTRUMENT CALIBRATION

For: WING MEMORIAL HOSPITAL / NUCLEAR MEDICINE  
LIC. NO. 20-15280-01

**Instrument:**

Manufacturer \_\_\_\_\_

Type \_\_\_\_\_

Model No. \_\_\_\_\_

Serial No. \_\_\_\_\_

Calibration Data:

[illegible]

Comments:

<u>Nuclide</u>	Activity or	<u>Calibration Accuracy</u>
	<u>Exposure Rate at Specified Distance</u>	

**Calibration Source:**

Calibrated by \_\_\_\_\_ Date \_\_\_\_\_



DATE 8/28/85

TELEPHONE OR VERBAL CONVERSATION RECORD

TIME 4:10 ☐ A.M. ☒ P.M.

☐ INCOMING CALL

☐ OUTGOING CALL

☐ VISIT

PERSON CALLING

OFFICE/ADDRESS

PHONE NUMBER

EXTENSION

Piccone

NRC - Region I

PERSON CALLED

OFFICE/ADDRESS Wing Memorial Hosp

PHONE NUMBER

EXTENSION

Dr. Kenneth McEwen  
Kathleen LaFleur, Chief  
Tech.

Wright Street  
Palmer, MA 01069

(413) 283-7651

CONVERSATION

SUBJECT

Amendment Request

SUMMARY Call returned by Mr. LaFleur on 8/29/85 @ 9:00 A.M.

- problems with NRC 313M A+B for  
Nicholas Spencer, M.D.

① need 500 hrs. total clinical training

② need 5 generator elutions and  
5 kit preps for Group III

③ We require a copy of the  
signed preceptor's statement

- submit copy of the survey meter calibration  
procedure so it can be on record with your  
license.

"OFFICIAL RECORD COPY" ML18

REFERRED TO:

ACTION REQUESTED

A letter detailing the above will be  
sent.

☐ ADVISE ME OF  
ACTION TAKEN.

INITIALS J.M.P.

DATE 8/29/85

ACTION TAKEN

INITIALS

DATE



COMMUNITY  
HEALTH  
CENTER

Wright Street  
Palmer, Ma. 01069  
413/283-7651

WITH SATELLITE  
HEALTH CENTERS  
AT

10 South Main Street  
Belchertown, Ma. 01007  
413/323-5118

101 Main Street  
Monson, Ma. 01057  
413/267-5181

3 Crane Park Drive  
Wilbraham, Ma. 01095  
413/596-3455

# WING MEMORIAL HOSPITAL

July 30, 1985

License No. 20-15280-01

Docket No. 030-08859

RECEIVED

United States  
Nuclear Regulatory Commission  
Region 1  
King of Prussia, Penn. 19406

'85 SEP -3 A10:21

U.S. N.R.C.  
LIC. FEE MGMT. BRANCH

Dear Dr. Glenn:

In response to the routine safety inspection conducted by Jenny Johansen and to your letter dated July 17, 1985, the following actions have been taken.

## Responses to Appendix A ; Notice of Violations:

- A) Please find enclosed the NRC-313M- Supplements A and B for Nicholas Spencer M.D., We would like him to be added to our license for Groups I, II, and III.

Prior written approval by the hospital administrator and the Medical Isotopes Committee will be maintained for inspection for all visiting physicians.

- B) 1) We will strictly enforce our commitment to use syringe shields in the preparation and injection of patient doses.
- 2) We would like to amend our license to allow us to possess 165 millicuries of Cesium 137 as a sealed source for use in the calibration of our survey meters. The source is owned by Mercy Hospital in Springfield, Mass. ( license No. 20-00096-02). The source will be transported to our facility by private carrier ( Ronald P. Hanc, who is listed on the Mercy license as the authorized user). Mr. Hanc will perform the survey meter calibrations as detailed in the Mercy Hospital license. Current D.O.T. regulations will be followed when the source is transported. Upon arrival of the source at our

8509190436 7PP.

AFFILIATED WITH THE UNIVERSITY OF MASSACHUSETTS MEDICAL CENTER

"OFFICIAL RECORD COPY"

ML10

04278

AUG 19 1985



COMMUNITY  
HEALTH  
CENTER

Wright Street  
Palmer, Ma. 01069  
413/283-7651

WITH SATELLITE  
HEALTH CENTERS  
AT

10 South Main Street  
Belchertown, Ma. 01007  
413/323-5118

121 Main Street  
Monson, Ma. 01057  
413/267-5181

3 Crane Park Drive  
Wilbraham, Ma. 01095  
413/596-3455

# WING MEMORIAL HOSPITAL

facility a Radioactive Shipment Receipt Report ( appendix F of Guide 10.8) will be completed. The source will be returned to Mercy Hospital later the same day by Mr. Hanc, after the calibrations of our survey meters have been completed.

- 3) We would like to amend our license to require accuracy testing of our dose calibrator on an annual basis rather than quarterly. Please find enclosed a copy of the test performed on 7/26/85. The dose claibrator linearity test will continue to be performed quarterly. We would, however, like to perform the linearity test using the CALICHECK Dose Calibrator Activity Linearity Test Kit. Manufacturers instructions and procedures will be carefully followed. Find enclosed the manufacturers literature on this test kit.
- 4) As per our discussion with Jenny Johansen we initiated ( 7/1/85 ) a study to evaluate the exposure to both hands of our ambidextrous technologist to determine which hand most adequately reflects the extremity exposure received by this individual. The study will be evaluated after three months of finger badge readings.

Respectfully submitted,

Mr. Richard Scheffer  
President  
Wing Memorial Hospital  
Wright Street  
Palmer, Mass. 01069

Sept - 1 - I

Applicant	006748
Check No.	006748
Amount	100.00
Type of	Amc
Date Check Rec'd	9/3/85
Received By	Jacques

refund 5



**TRAINING AND EXPERIENCE**  
**AUTHORIZED USER ~~OR RADIATION SAFETY OFFICER~~**  
(Completed for RSC Record 6/19/85)

1. NAME OF AUTHORIZED USER OR RADIATION SAFETY OFFICER  NICHOLAS SPENCER, M.D.	2. STATE OR TERRITORY IN WHICH LICENSED TO PRACTICE MEDICINE Massachusetts
--	---

3. CERTIFICATION		
SPECIALTY BOARD A	CATEGORY B	MONTH AND YEAR CERTIFIED C
American Board of Radiology	Diagnostic	June, 1981

4. TRAINING RECEIVED IN BASIC RADIOISOTOPE HANDLING TECHNIQUES			
FIELD OF TRAINING A	LOCATION AND DATE(S) OF TRAINING B	TYPE AND LENGTH OF TRAINING	
		LECTURE/ LABORATORY COURSES (Hours) C	SUPERVISED LABORATORY EXPERIENCE (Hours) D
a. RADIATION PHYSICS AND INSTRUMENTATION	Boston City Hospital - 1980	150	20
b. RADIATION PROTECTION	Boston City Hospital - 1980	30	--
c. MATHEMATICS PERTAINING TO THE USE AND MEASUREMENT OF RADIOACTIVITY	Boston City Hospital - 1980	50	15
d. RADIATION BIOLOGY	Boston City Hospital - 1980	25	--
e. RADIOPHARMACEUTICAL CHEMISTRY	Boston City Hospital - 1980	15	10

5. EXPERIENCE WITH RADIATION. (Actual use of Radioisotopes or Equivalent Experience)				
ISOTOPE	MAXIMUM AMOUNT	WHERE EXPERIENCE WAS GAINED	DURATION OF EXPERIENCE	TYPE OF USE
As listed above	100 mCi	Boston City Hospital	9 months	clinical

PRECEPTOR STATEMENT

Supplement B must be completed by the applicant physician's preceptor. If more than one preceptor is necessary to document experience, obtain a separate statement from each.

1. APPLICANT PHYSICIAN'S NAME AND ADDRESS			<b>KEY TO COLUMN C</b> <b>PERSONAL PARTICIPATION SHOULD CONSIST OF:</b> 1-Supervised examination of patients to determine the suitability for radioisotope diagnosis and/or treatment and recommendation for prescribed dosage. 2-Collaboration in dose calibration and actual administration of dose to the patient including calculation of the radiation dose, related measurements and plotting of data. 3-Adequate period of training to enable physician to manage radioactive patients and follow patients through diagnosis and/or course of treatment.
FULL NAME			
NICHOLAS SPENCER, M.D.			
STREET ADDRESS			
130 Maple Street			
CITY	STATE	ZIP CODE	
Springfield	MA	01105	

2. CLINICAL TRAINING AND EXPERIENCE OF ABOVE NAMED PHYSICIAN

ISOTOPE A	CONDITIONS DIAGNOSED OR TREATED B	NUMBER OF CASES INVOLVING PERSONAL PARTICIPATION C	COMMENTS (Additional information or comments may be submitted in duplicate on separate sheets.) D
I-131 or I-125	DIAGNOSIS OF THYROID FUNCTION	65	
	DETERMINATION OF BLOOD AND BLOOD PLASMA VOLUME	---	
	LIVER FUNCTION STUDIES	---	
	FAT ABSORPTION STUDIES	---	
	KIDNEY FUNCTION STUDIES	---	
	IN VITRO STUDIES	---	
OTHER		---	
I-125	DETECTION OF THROMBOSIS	---	
I-131	THYROID IMAGING	65	
P-32	EYE TUMOR LOCALIZATION	---	
Se-75	PANCREAS IMAGING	---	
Yb-169	CISTERNOGRAPHY	10	
Xe-133	BLOOD FLOW STUDIES AND PULMONARY FUNCTION STUDIES	70	
OTHER		---	
Tc-99m	BRAIN IMAGING	30	
	CARDIAC IMAGING	60	
	THYROID IMAGING	---	
	SALIVARY GLAND IMAGING	---	
	BLOOD POOL IMAGING	---	
	PLACENTA LOCALIZATION	---	
	LIVER AND SPLEEN IMAGING	100	
	LUNG IMAGING	70	
	BONE IMAGING	100	
OTHER		---	

# PRECEPTOR STATEMENT (Continued)

## 2. CLINICAL TRAINING AND EXPERIENCE OF ABOVE NAMED PHYSICIAN (Continued)

ISOTOPE A	CONDITIONS DIAGNOSED OR TREATED B	NUMBER OF CASES INVOLVING PERSONAL PARTICIPATION C	COMMENTS (Additional information or comments may be submitted in duplicate on separate sheets.) D
P-32 (Soluble)	TREATMENT OF POLYCYTHEMIA VERA, LEUKEMIA, AND BONE METASTASES	--	
P-32 (Colloidal)	INTRACAVITARY TREATMENT	--	
I-131	TREATMENT OF THYROID CARCINOMA	15	
	TREATMENT OF HYPERTHYROIDISM	15	
Au-198	INTRACAVITARY TREATMENT	--	
Co-60 or Cs-137	INTERSTITIAL TREATMENT	--	
	INTRACAVITARY TREATMENT	--	
I-125 or Ir-192	INTERSTITIAL TREATMENT	--	
Co-60 or Cs-137	TELETHERAPY TREATMENT	--	
Sr-90	TREATMENT OF EYE DISEASE	--	
	RADIOPHARMACEUTICAL PREPARATION	--	
Mo-99/ Tc-99m	GENERATOR	--	
Sn-113/ In-113m	GENERATOR	--	
Tc-99m	REAGENT KITS	--	
Other		--	

## 3. DATES AND TOTAL NUMBER OF HOURS RECEIVED IN CLINICAL RADIOISOTOPE TRAINING

400 hours

1980 - 1981

## 4. THE TRAINING AND EXPERIENCE INDICATED ABOVE WAS OBTAINED UNDER THE SUPERVISION OF:

a. NAME OF SUPERVISOR

Victor Lee, M.D.

b. NAME OF INSTITUTION

Boston City Hospital

c. MAILING ADDRESS

Harrison Avenue

d. CITY

Boston, MA

## 5. MATERIALS LICENSE NUMBER(S)

## 6. PRECEPTOR'S SIGNATURE

Signature on original forms in  
Medical Staff Office

## 7. PRECEPTOR'S NAME (Please type or print)

Victor Lee, M.D.

## 8. DATE

Completed for our records 6/19/85 by  
Dr. Brahmavar in consultation with  
Dr. Spencer.

Nuclear Medicine Department

ANNUAL ACCURACY TEST OF THE DOSE CALIBRATOR

HOSPITAL Wing Memorial Hosp.  
INSTRUMENT LOCATION Nuclear Med.  
MODEL NUMBER CRC 6A  
TECHNOLOGIST Kathy Lapham RTR

DATE July 26, 1985  
MANUFACTOR Capintec  
SERIAL NUMBER 62630  
PHYSICIST Ronald P. Hare, M.S.

STANDARDS USED:

CALCULATED ACTIVITY DAY OF CALIBRATIONS:

Source	Serial #	Original Act/Date	Activity today	Dose Cal. Reading	% Variation
#1 $^{60}\text{Co}$ 57	200484B-47	4.8mCi 4/26/84	1.493mCi	1.460mCi	-2.2%
#2 $^{137}\text{Cs}$ 137	3561081A-37	207uCi 10/6/81	189.9uCi	190.5uCi	+ .3%
#3 $^{60}\text{Co}$ 60	3540581A-51	53uCi 5-8-81	34.72uCi	33.8uCi	-2.4%

Percent variation = Dose cal reading div by actual act.

THE ACTIVITY SHOULD BE WITHIN + OR - 5.0% OF THE CALCULATED ACTIVITY TO ASSURE ACCURACY OF THE DOSE CALIBRATOR. IF THE READINGS ARE NOT WITHIN THE 5% RANGE, REPAIR OR ADJUSTMENT IS NECESSARY.

THIS SURVEY IS TO BE PERFORMED ANNUALLY AS A REQUIREMNT OF THE NRC. REGULATORY GUIDE 10.8. APPENDIX D, SECTION 2, ITEM G.



# DOSE CALIBRATOR

## Activity Linearity Testing the easy way

### Fast

Now with the newly developed Calicheck™ dose calibrator activity linearity test kit, you can meet N.R.C. Regulatory Guide 10.8, appendix D., Section 2E or your state's equivalent requirement in just 4 minutes — not days. You can complete the test in one short sitting and check for linearity virtually at a glance. Plus you eliminate the frustration of having to start the test all over simply because you forgot to take a reading on time.

### Accurate and Reliable

The new Calicheck kit is designed to attenuate  $^{99m}\text{Tc}$  by known values — accurate using a high yield generator eluant or a unit dose.

A Calicheck kit provides for seven successive measurements simulating the decay of  $^{99m}\text{Tc}$  at approximately 0, 6, 12, 20, 30, 40 and 50 hours from the initial assay.

### Complete Yet Reusable

Your Calicheck kit comes to you complete with its own storage container, a unique arrangement of seven color-coded lead-wrapped tubes, work/record keeping sheets, instructions for use and a license amendment form (if needed.)

Your Calicheck kit is completely reusable for an indefinite period of time. There is nothing to wear out or use up. If damage should cause a tube to malfunction, individual replacements are available.

### Safe

Your use of a Calicheck kit eliminates the need to fractionate eluants or decay the elution for several days while periodically collecting data to determine linearity. Time of potential exposure to radiation is drastically reduced, thereby maintaining exposures ALARA.

### Lowers Department Cost

When you test with a Calicheck kit, both the source activity and

dose calibrator can be returned to active service in just minutes. This savings alone can pay for a Calicheck kit in just three to four linearity tests. A Calicheck kit lets you return to active service too!


### Can Improve Patient Care

A Calicheck kit is so fast, efficient and easy to use, you may wish to check dose calibrator linearity more frequently. Lets you spot trouble before it becomes serious.

### Low Price

A Calicheck dose calibrator activity linearity test kit is just \$375.00 shipping included.


Just call (216) 663-1773 or write: Calcorp, Inc., P.O. Box 25589, Cleveland, Ohio 44125-0589.



### Just four minutes

As simple as 1, 2, 3, 4, 5, 6, 7. Place central tube in the dose calibrator. Place the source in this tube and take a reading. Then sequentially place color-coded tubes over the central tube. Additional readings are taken immediately, converted with a predetermined factor and you can see the degree of linearity virtually at a glance.

May require approval of the Agency issuing your radioactive materials license.



**Calicheck™**

• DOSE CALIBRATOR ACTIVITY LINEARITY TEST KIT •

Patent pending

See us at the SNM Show in Miami Beach at Booth 404

Reprinted from the JOURNAL OF NUCLEAR MEDICINE  
May, 1982, Volume 23, Number 5