

ms 16
P2

Cardiovascular Associates, P.C.
WALLACE B. LEBOWITZ, M. D.

2660 MAIN STREET
BRIDGEPORT, CONNECTICUT 06606
TELEPHONE 334-1755

2068 BRIDGEPORT AVENUE
MILFORD, CONNECTICUT 06460
TELEPHONE 874-1799

October 23, 1985

John E. Glenn, PHD.,
Chief of Nuclear Materials, Safety Section B
Division of Radiation, Safety and Safeguards
United States Nuclear Regulatory Commission, Region 1
631 Park Avenue
King of Prussia, PA 19406


re: Docket # 030-28674
Control # 18953

Dear Dr. Glenn:

Pursuant to our phone conversation of October 17, 1985 this letter is to document our intention to obtain either a hundred to two hundred microcuries of cesium-137 or greater than one millicurie of cobalt-57 for daily constancy check of our dose calibrator.

Thank you for your help in the completion of the documentation necessary for obtaining my material license.

Sincerely,



Theodore F. Ritzer, M.D.

TFR:ks

8601170473 851203
REG1 LIC30
06-23533-01 PDR

RECEIVED-REGION 1
OCT 28 PM 5 18

18953
"OFFICIAL RECORD COPY"

ML10

OCT 28 1985

10/17/85

TELEPHONE OR VERBAL CONVERSATION RECORD

TIME

2:15

☐ A.M.
☒ P.M.

☒ INCOMING CALL

☐ OUTGOING CALL

☐ VISIT

PERSON CALLING

THEODORE RITZER

OFFICE/ADDRESS

PHONE NUMBER

EXTENSION

203 874 1799

PERSON CALLED

JOHN GLENN

OFFICE/ADDRESS

Region I

PHONE NUMBER

EXTENSION

CONVERSATION

SUBJECT

Application, Control # 118953

SUMMARY

Dr. Ritzer called to inquire about the status of his application and as to ways to expedite its approval. As I had a draft deficiency letter on my desk, I informed him of the deficiency.

Dr. ~~Ritzer~~ Ritzer will send a letter confirming that proper sources for the daily constancy check will be purchased and available. Either ~~Co-57~~ Co-137 or Co-57.

"OFFICIAL RECORD COPY"

ML10

REFERRED TO:

File

ACTION REQUESTED

Milestone 15

☐ ADVISE ME OF ACTION TAKEN.

INITIALS

DATE

ACTION TAKEN

INITIALS

DATE