

SEP 23 1985

Docket No. 030-14988

License No. 07-18391-01

Riverside Hospital
ATTN: Mr. Noval Copeland
Executive Director
P. O. Box 345
700 Lea Boulevard
Wilmington, Delaware 19802

Gentlemen:

Subject: Inspection Report No. 030-14988/85-01

This refers to the routine safety inspection conducted by R. H. Ladun of this office on September 4, 1985, of activities authorized by NRC License No. 07-18391-01 and to the discussions of our findings held by Mr. Ladun with Dr. C. Milunsky of your staff at the conclusion of the inspection.

The inspection was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, measurements made by the inspector, and observations by the inspector.

Our inspector also verified the steps you have taken to correct the violations brought to your attention in the enclosure to our letter dated December 15, 1982. We have no further questions regarding your action at this time.

Within the scope of this inspection, no violations were observed.

During the course of the inspection, it was determined that your dose calibrator was taken out for repair and that you were utilizing a calibrator at a nearby hospital to assay radioisotopes. We understand that this was an emergency situation. However, you are advised to amend your license to include such a contingency in the future.

In accordance with Section 2.790 of the NRC's "Rules of Practice," Part 2, Title 10, Code of Federal Regulations, a copy of this letter will be placed in the Public Document Room.

OFFICIAL RECORD COPY

DL RIVERSIDE HOSPITAL - 0001.0.0
09/17/85

8510080590 850923
REG1 LIC30
07-18391-01 PDR

RETURN ORIGINAL TO
REGION I

IE:07

10

SEP 23 1985

Riverside Hospital

2

No reply to this letter is required. Your cooperation with us in this matter is appreciated.

Sincerely,

Original Signed By:

John D. Kinneman


John D. Kinneman, Chief
Nuclear Materials Safety Section A,
Division of Radiation Safety
and Safeguards

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
State of Delaware

bcc w/encl:
Region I Docket Room (w/concurrences)
Senior Operations Officer

RI:DRSS
Ladun/djh

9/16/85
R


RI:DRSS
Kinneman

9/19/85

OFFICIAL RECORD COPY

DL RIVERSIDE HOSPITAL - 0002.0.0
09/17/85