

FEB 5 1986

Basha Radiology
ATTN: Y. M. Basha, M.D.
1401 S. Washington Avenue
Royal Oak, MI 48067

Gentlemen:

Enclosed is Amendment Number 06 to your NRC License Number 21-20416-01 which authorizes the changes you requested.

Please be advised that we cannot authorize you to release your old nuclear medicine space for unrestricted use (even by other members of your staff) until we have received and reviewed a copy of the results of your close-out survey. The survey should consist of exposure rate measurements to show that all sources of radioactive material have been removed, and contamination checks of areas where radioactive materials were used or stored. Average radiation levels associated with surface contamination and removable contamination should not exceed those specified in the enclosed decontamination guide. Please submit the following information with your close-out survey:

- a. A diagram of your old facility with survey and wipe test results keyed to specific locations.
- b. The name of the person performing the survey.
- c. The date the survey was performed.
- d. The instrument(s) used for exposure rate measurements and for analysis of the wipes.
- e. Background readings.
- f. The date that the survey instrument was last calibrated.

Information submitted in response to this letter should state that it is additional information and not a new amendment request. If you have questions, please contact our office by dialing (312) 790-5625.

Sincerely,

8604070043 860205
REG3 LIC30
21-20416-01 PDR

Original Signed By
Evelyn R. Matson
Materials Licensing Section

Enclosures:

1. Amendment No. 06
2. Guidelines for Decontamination
of Facilities

RIII

ERM
Matson/cm

2/4/86