

# UNITED STATES ATOMIC ENERGY COMMISSION

## DIVISION OF COMPLIANCE

*I. C. III, District*

<p>1. LICENSEE</p> <p><b>Monsanto Chemical Company</b>  <b>Organic Chemicals Division</b>  <b>1700 South Second Street</b>  <b>St. Louis, Missouri</b></p>	<p>2. REGIONAL OFFICE</p> <p style="text-align: center;">REGION III, DIV. OF COMPLIANCE  OAKBROOK PROFESSIONAL BLDG.  OAK BROOK, ILLINOIS</p>
<p>3. LICENSE NUMBER</p> <p><b>24-1113-9</b></p>	<p>4. DATE(S) OF INSPECTION</p> <p style="text-align: center;"><b>August 15, 1963</b></p>

5. The following activities under your license (identified in Item No. 3 above) appear to be in noncompliance with AEC regulations or license requirements, as indicated.

- A. The 250 millicurie tritium source possessed under this license is mounted in a Perkin Elmer Model 154-0709 Detector and used in a Perkin Elmer Model 154 gas chromatograph, whereas Items 7 and 9 of this license require that this source be a Radiation Research Corporation Model TR-1 source for use in either a Barber-Colman 61-C or a Wilkens electron capture detector.

Supplementary page None attached.

*Glen W. Roy*  
**Glen W. Roy**  
AEC Compliance Inspector

**9-17-63**  
Date

ORIGINAL: LICENSEE.

COPIES: ☒ CO REGION ☐ CO HEADQUARTERS ☐ L&R HEADQUARTERS.

U.S. GOVERNMENT PRINTING OFFICE: 1962-O-606811

9702060374 970127  
PDR FOIA  
NEITZEL96-314 PDR