NRC FORM 366
(7-77)

U. S. NUCLEAR REGULATORY COMMISSION

LICENSEE EVENT REPORT

CONTROL BLOCK: (1) (PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

01 | NCBEP1 (2) | 00-0000-00 (3) | 4111 1 (4) | (5)
LICENSEE CODE — LICENSE NUMBER — LICENSE TYPE — CAT

CON'T
01 | REPORT SOURCE L (6) | 0500-0325 (7) | 122579 (8) | 011880 (9)
DOCKET NUMBER — EVENT DATE — REPORT DATE

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10)

02 During normal plant operation, the control operator noticed that the containment
03 atmospheric monitors CAC-AQH-1261 and 1262, channels 2 & 3, had failed downscale.
04
05
06
07
08 Technical Specification 3.3.5.3, 6.9.1.9b

09 | SYSTEM CODE SC (11) | CAUSE CODE E (12) | CAUSE SUBCODE G (13) | COMPONENT CODE INSTRU (14) | COMP SUBCODE E (15) | VALVE SUBCODE Z (16)

(17) LER RO REPORT NUMBER | EVENT YEAR 79 | — | SEQUENTIAL REPORT NO. 111 | — | OCCURRENCE CODE 03 | REPORT TYPE L | — | REVISION NO. 0

ACTION TAKEN B (18) | FUTURE ACTION X (19) | EFFECT ON PLANT Z (20) | SHUTDOWN METHOD Z (21) | HOURS (22) 0000 | ATTACHMENT SUBMITTED N (23) | NPRD-4 FORM SUB. Y (24) | PRIME COMP. SUPPLIER N (25) | COMPONENT MANUFACTURER N305

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)

10 The problem was determined to be condensation in the detector cabinets causing the
11 electronics to fail. The detectors were repaired, the detector chambers dried out,
12 and the system returned to service. An Engineering Work Request has been submitted
13 to replace these monitors with more dependable equipment as condensation has been
14 a problem with these instruments.

FACILITY STATUS E (28) | POWER 098 (29) | OTHER STATUS (30) NA | METHOD OF DISCOVERY A (31) | DISCOVERY DESCRIPTION (32) Operator Surveillance

ACTIVITY RELEASED Z (33) | CONTENT OF RELEASE Z (34) | AMOUNT OF ACTIVITY (35) NA | NA | LOCATION OF RELEASE (36)

PERSONNEL EXPOSURES NUMBER 000 | TYPE Z (38) | DESCRIPTION (39) NA 90030260

PERSONNEL INJURIES NUMBER 000 (40) | DESCRIPTION (41) NA

LOSS OF OR DAMAGE TO FACILITY TYPE Z (42) | DESCRIPTION (43) NA 8001240437

PUBLICITY ISSUED N (44) | DESCRIPTION (45) NA | NRC USE ONLY

NAME OF PREPARER A. C. Tollison, Jr. PHONE 919-457-9521