

GUTHRIE RESEARCH INSTITUTE

November 23, 1987

Donald Guthrie Foundation
for Medical Research
Guthrie Square
Sayre, PA 18840-1692
717-888-6666, ext. 4620
717-882-4620



John D. Kinneman, Chief
Nuclear Materials Safety Section A
Division of Radiation Safety
and Safeguards
U.S. Nuclear Regulatory Commission
Region I
631 Park Avenue
King of Prussia, PA 19406

Dear Mr. Kinneman:

I would like to respond to your letter of October 29, 1987 in which you cite violations noted on a recent inspection of our facility (License No. 37-21383-01). We have taken significant steps to correct these deficiencies, a description of which is enclosed.

In addition, I would like to inform you of two other changes with regard to our facility and its use of radioactive materials. The first is that I have recently (August 3) been appointed the new Scientific Director and Radiation Protection Officer of the Guthrie Research Institute and will replace Dr. Litwin who left this institution on September 30th. We have recently filed an amendment to our license to indicate this change. Secondly, the Guthrie Medical Center has recently recruited a new Health Physicist, Dr. T.K. Nair, who will be inspecting the Guthrie Research Institute on a regular basis and will be available for consultation. After receiving your letter of October 29, Dr. Nair conducted several detailed inspections of the facility and advised and assisted us in implementing the changes outlined in the enclosed documents.

I hope that the steps taken will be judged appropriate by your office in order to bring our institution into compliance with NRC rules and regulations. However, if there are questions or if you feel that additional measures are necessary, please do not hesitate to get in touch with me immediately. Thank you.

Yours truly,

Robert E. Hall, M.D., Ph.D.
Scientific Director

8802090113 880201
REG1 LIC30
37-21383-01 PDR

John M. Thomas, M.D.
President
Robert E. Hall, M.D., Ph.D.
Scientific Director

SCIENTIFIC ADVISORY BOARD

D. Bernard Amos, M.D.
*James B. Duke Professor of Immunology
and Experimental Surgery*
Duke University
Douglas D. McGregor, M.D., D. Phil.
Director
James A. Baker Institute for Animal Health
Cornell University
Matthew D. Scharff, M.D.
*Professor of the Department of
Cell Biology*
Albert Einstein College of Medicine
of Yeshiva University
Henry J. Winn, Ph.D.
*Chief of Immunology and Professor
Department of Microbiology &
Molecular Genetics*
Massachusetts General Hospital

PRESIDENTIAL ADVISORY COUNCIL

William C. Beck, M.D.
President Emeritus
Guthrie Foundation
Willibald H. Conzen
Former Chairman
Schering-Plough Corporation
Richard Day, M.D.
Former Chairman, Board of Trustees
Union College
Francis R. Grebe
Investment Counselor
James M. Davidson and Company
Nicholas Niles
Publisher
"Changing Times"

RESPONSE TO NRC LETTER OF OCTOBER 29, 1987

REGARDING LICENSE NO. 37-21383-01

- A. The following survey instruments used in the laboratory for bench monitoring and I-125 contamination measurement were calibrated and updated as detailed below.

1. Survey meter Model 5-10E, S/N 022676 on August 12, 1987.
2. Survey meter Model 5-10E, S/N 022697 on August 12, 1987.
3. Survey meter Model Series 900, S/N 024212 on August 12, 1987, by Technical Associates and
4. Survey meter Model PUG-1, S/N 083131 on November 19, 1987, by Applied Health Physics.

Steps have been taken in consultation with the acting Radiation Safety Officer at Robert Packer Hospital, to insure that the survey instruments are calibrated annually.

- B. The contamination detected near the sink of Bench #5 was an unusual occurrence and the area has been decontaminated on August 4, 1987, by Dr. Kestler, Research Associate at the lab. The wipetest showed less than 200 cpm after decontamination. The survey and wipetest conducted on November 17, 1987, showed no removable contamination detected over the working benches, floor and sinks. To prevent undetected contamination, the following steps have been implemented.

1. Daily survey of the work benches and sinks at the end of work.
2. Wipetest of working benches sinks, floor and radiation protection survey of the laboratory once a week.

Further acting radiation safety officer of Robert Packer Hospital will be consulted for quarterly review of the radiation safety program.

CERTIFICATE OF INSTRUMENT CALIBRATION

AND CERTIFICATE OF COMPLIANCE

This is to certify that Survey Meter (Model) 5.105
(Serial Number) 022676 was calibrated on 8/12/82 (Date).

Except as noted below, this instrument now meets the manufacturer's tolerance of \pm 15 % F.S. Source used for calibration is traceable to NBS. The isotope used was Cs-137. The instrument was calibrated at an ambient temperature of 60-80° F and includes a background reading of 0.02 mR/hr.

Calibration is performed in conformance with recommendations of the U.S. Nuclear Regulatory Commission, and agreement State regulations and of the International Commission of Radiation Protection. Ref: 1. U.S. Nuclear Regulatory Comm. RG8.21 Section 1.12, and, 2. California ADM. Code, Title 17, Section 30332(c) (1), also N-323-1978.

Two points were checked at each scale; one point is the lower 25% of the scale and one point in the upper 25%. Adjustments have been made so that the instrument reads within the above accuracy on all ranges when compared to true dose rate, unless noted below. Analog instruments reading in counts/minutes are calibrated against a pulse generator traceable to the NBS in addition to checking against a calibrated radiation source.

Comments:

Calibration Technician: JK

Approved by: BJ

PURCHASE ORDER: R89037

Property of: D. Guthrie Foundation



TECHNICAL ASSOCIATES

7051 ETON AVENUE • CANOGA PARK, CA 91303 • (818) 883-7043

CERTIFICATE OF INSTRUMENT CALIBRATION

AND CERTIFICATE OF COMPLIANCE

This is to certify that Survey Meter (Model) 5-10E
(Serial Number) 022697 was calibrated on 8/12/87 (Date).

Except as noted below, this instrument now meets the manufacturer's tolerance of \pm 15 % F.S. Source used for calibration is traceable to NBS. The isotope used was Cs-137. The instrument was calibrated at an ambient temperature of 60-80° F and includes a background reading of 0.02 mR/hr.

Calibration is performed in conformance with recommendations of the U.S. Nuclear Regulatory Commission, and agreement State regulations and of the International Commission of Radiation Protection. Ref: 1. U.S. Nuclear Regulatory Comm. RG8 21 Section 1.12, and, 2. California ADM. Code, Title 17, Section 30332(c) (1), also N-323-1978.

Two points were checked at each scale; one point is the lower 25% of the scale and one point in the upper 25%. Adjustments have been made so that the instrument reads within the above accuracy on all ranges when compared to true dose rate, unless noted below. Analog instruments reading in counts/minutes are calibrated against a pulse generator traceable to the NBS in addition to checking against a calibrated radiation source.

Comments:

Calibration Technician: VH
Approved by: BG
PURCHASE ORDER: R89037
Property of: A Guthrie Foundation



TECHNICAL ASSOCIATES

7051 ETON AVENUE • CANOGA PARK, CA 91303 • (818) 883-7043

CERTIFICATE OF INSTRUMENT CALIBRATION

AND CERTIFICATE OF COMPLIANCE

This is to certify that Survey Meter (Model) Series 900
(Serial Number) 024212 was calibrated on 8/12/87 (Date).

Except as noted below, this instrument now meets the manufacturer's tolerance of \pm 15 % F.S. Source used for calibration is traceable to NBS. The isotope used was Cs-137. The instrument was calibrated at an ambient temperature of 60-80° F and includes a background reading of 0.02 mR/hr.

Calibration is performed in conformance with recommendations of the U.S. Nuclear Regulatory Commission, and agreement State regulations and of the International Commission of Radiation Protection. Ref: 1. U.S. Nuclear Regulatory Comm. RG8.21 Section 1.12, and, 2. California ADM. Code, Title 17, Section 30332(c) (1), also N-323-1978.

Two points were checked at each scale; one point is the lower 25% of the scale and one point in the upper 25%. Adjustments have been made so that the instrument reads within the above accuracy on all ranges when compared to true dose rate, unless noted below. Analog instruments reading in counts/minutes are calibrated against a pulse generator traceable to the NBS in addition to checking against a calibrated radiation source.

Comments:

Calibration Technician: JK
Approved by: BG
PURCHASE ORDER: R89037
Property of: A. Guthrie Foundation



TECHNICAL ASSOCIATES

7051 ETON AVENUE • CANOGA PARK, CA 91303 • (818) 883-7043



HEALTH PHYSICS Inc.

2966 Industrial Blvd. • Box 197 • Bethel Park, Pa. 15102 • Phone 412 • 563-2242

CALIBRATION CERTIFICATE

SHIPPING ADDRESS

BILLING ADDRESS

Cuthrie Research Institute
Cuthrie Square
Sayre, PA 18740

NAME

CONTACT: R.L.H. Porter PHONE # 517-882-4627 DATE 11/19/87 P.O. # 3205

condition of equipment received Rush Calibration

Mfg. Inst. Tech. Assoc.
Detector IIMod#
Mod#PUE-1
P-8Ser#
Ser#083121
08400

CALIBRATION



REPAIR



SALE



LOAN

By:

J. Langel

scale	source	reading	scale	source	reading	scale	source	reading
cpm	cpm	cpm	MR/MR	MR/MR	cpm			
X1	200	202	X1	1	270			
X10	2000	2050	X10	1	2600			
X100	20000	20000	X100	11.2	25000			

Calibration Source



GAMMA



ELECT



ALPHA



BETA



OTHER

Description



ra-226



cs-137



sr-90




pu-239



MP-1



Response Graph



N/A

Probe Efficiencies

✓ ALPHA

8

✓ BETA

17

Check Source Readings

EW

SW

PW

IC

Maintenance & Comments Replaced 2 D cells, Taped & Tightened VAC Connector.

Tested, Inspected & Calibrated

Calibration

\$50.00/Cal

50.00

DOS CHECK

\$50.00/4ea

LABOR

\$50.00/Hr

MATERIALS

2 D cells

20.00

&

SALES

Shipping

UPS/NDP

10.00

\$24.00

Shipping

DATE

UPS/NDP
11/19/87

QA DEPT

PICK-UP

DATE

X

X

NOTICE: Under Applied Health Physics license #37-09135-01 & in accordance with Federal, Local or State regulations sources are traceable to the National Bureau of Standards. This certificate expires in 12 months and should be recalibrated on or before 11/19/88. The frequency of this recalibration may vary due to governmental requirements. A copy of this record should be maintained for future inspection by the appropriate authorities.

CONSULTATION • SERVICES • PRODUCTS • RADIATION APPLICATIONS

AHP RIS REV 3.87