

SAFETY INSPECTION

1. LICENSEE Royal C. Johnson Veterans Memorial Hospital P.O. Box 5046 Sioux Falls, South Dakota 57117		2. REGIONAL OFFICE U. S. NRC, Region IV 611 Ryan Plaza, Suite 1000 Arlington, Texas 76011	
3. DOCKET NUMBER(S) 30-10840	4. LICENSE NUMBER(S) 40-16336-01	5. DATE OF INSPECTION Feb 25, 1988	

Licensee:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission's (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews, with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

- ☐ 1. Within the scope of this inspection, no violations were observed.
- ☐ 2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.
- ☒ 3. During this inspection certain of your activities, as checked below, were in violation of NRC requirements.
THIS IS A NOTICE OF VIOLATION which is required to be posted in accordance with 10 CFR 19.11.
- ☐ A. _____ was not properly posted to indicate the presence of a _____ 10 CFR 20.203(b), (c), (d), (e) or 34.42.
- ☐ B. Containers located in _____ were not properly labeled to indicate the presence of radioactive material. 10 CFR 20.203(f)(1), or (f)(2).
- ☐ C. _____ of sealed sources were not performed at the proper frequencies. 10 CFR _____ License Condition Number _____
- ☐ D. Records of _____ were not properly maintained. 10 CFR _____ or License Condition Number _____
- ☐ E. Documents were not properly posted or otherwise made available. 10 CFR 19.11.
- ☐ F. Reports or notifications of _____ were not made in accordance with 10 CFR _____ or License Condition Number _____
- ☒ H. Unauthorized Radiation Safety Officer contrary to License Condition 11.
- ☒ I. Incomplete Dose Calibrator daily constancy check procedure Contrary to License Condition 14.
- ☒ J. Dose Calibrator Accuracy Tests determined the ¹³³Ba reading to contain an error greater than $\pm 5\%$ for 1984, 1985, 1986, and 1987. The calibrator was not repaired, adjusted, or a calibration factor calibrated
- ☐ K. Calculated, Contrary to License Condition 14.

I hereby state that within 30 days the actions described by me to the inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

SIGNATURE - LICENSEE <i>Robert D. D.</i> 8803160052 880225 REG 4 LIC 30 40-16336-01	DATE 2/25/88	SIGNATURE - NRC INSPECTOR <i>Wesley L. Kelly</i> 2/25/88	DATE 2/25/88
---	-----------------	--	-----------------