

01 AUG 1988

Docket No. 030-20741

License No. 08-20702-01

Metropolitan Radiology Associates, CHTD
ATTN: Barry G. Brotman, M.D.
Partner
2139 Georgia Avenue, N.W.
Suite 4B
Washington, D.C. 20060

Gentlemen:

Subject: Inspection No. 030-20741/88-01

This refers to your letter dated June 10, 1988, in response to our letter dated May 26, 1988.

Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

From a telephone conversation of July 18, 1988 between you and Waymon Wallace of this office, we understand that daily constancy checks of your dose calibrator will be performed everyday when patient doses are drawn to verify that the measured activity is within $\pm 5\%$ of the predicted activity. We will review this matter during the next inspection of your licensed program.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:
John E. Glenn

John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section A
Division of Radiation Safety
and Safeguards

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
District of Columbia

OFFICIAL RECORD COPY

RL METROPOLITAN - 0001.0.0
07/21/88

8808030134 880801
REG1 LIC30
08-20702-01 PDC

RETURN ORIGINAL TO
REGION I

11
IE:07

01 AUG 1988

Metropolitan Radio'ogy
Associates, CHTD

2

bcc:
Region I Docket Room (w/concurrences)

RI:DRSS
Wallace/tlm
07/26/88

gg RI:DRSS
Glenn
07/26/88

OFFICIAL RECORD COPY

RL METROPOLITAN - 0002.0.0
07/21/88