



walter c. mcrone associates, inc.

CONSULTING: ULTRAMICROANALYSIS • MICROSCOPY • SMALL PARTICLE PROBLEMS • SOLID-STATE CHEMISTRY

5 June 1985

Mr. James W. Patterson
U. S. N. R. C.
Materials License Section
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Ref: BML #12-10632-1

Dear Mr. Patterson:

We have returned the H3 foil from our Varian E.C. detector to Varian Associates for disposal. Currently we hold a Valco Instruments Company Model 140B S/N152 electron capture detector.

I have included pages from their manual and a picture of the notice on the surface of the chamber. These indicate that the source is normally not accessible. We will not attempt to repair or replace the foil should such service be necessary, but will return the detector to Valco for service.

I have included a copy of Record of Source Transfer. Upon receipt of the detector the blue copy was sent to the Region 1 USNRC. At that time I thought that that was all that was necessary. I am now requesting that an amendment be made to our license to delete the Varian source and include this source.

If you have any questions, please write and we will respond promptly.

Sincerely,

Robert Z. Muggli PhD

Robert Z. Muggli, PhD
Senior Research Chemist

RZM:cb
Encls.

8601020439 851125
REG3 LIC30
12-10632-01 PDR

RECEIVED

JUN 07 1985

REGION III

JUN 7 1985

APPENDIX A

RADIATION SAFETY

Important Information

Read before installation of Model 140 detector

General Description

The following comments are not intended to negate or modify the appropriate government regulations covering radiation sources. The discussion herein is provided to acquaint the user with the radiation safety considerations associated with the Valco Model-140 ECD system. The comments also serve to indicate the responsibilities of personnel and organizations which use this instrument.

Within the detector assembly of the Model 140 is a scandium-coated stainless-steel foil which has been impregnated with tritium. This radioactive isotope, ^3H , decays with a half life of 12.26 years by the emission of beta particles to a stable nuclide of helium, ^3He . These beta particles are emitted with a maximum energy of 0.018 MeV and are absorbed by less than 1 mg/cm² of aluminum. Thus, there is no discernible radiation from the tritium external to the detector chamber and no hazard so long as the chamber integrity is not violated.


Temperature Limit

Scandium tritide is chemically stable at temperatures up to at least 325° C, which exceeds the maximum operating temperature of the Model-140 ECD. The instrument's temperature-control circuitry is designed to limit the maximum source temperature to less than 320° C, and a bimetallic thermostat is provided for back-up over-temperature protection.

It is very important that no attempt be made to defeat these temperature limit controls or otherwise heat the detector beyond 320° C. Such modifications would be in violation of applicable government regulations and might permit the release of radioactive material into the ambient air.

Assurance of Source Containment

The Model-140 ECD detector block is designed to resist attempts to open it using ordinary tools and nondestructive methods. Following assembly and initial testing, drive pins are inserted to prevent rotation of the threaded plugs which would otherwise provide access to the source foil. This design feature ensures that destructive, clearly detectable means are necessary to gain access to the scandium-tritide source foil.




Any attempt to open the detector block forcibly is prohibited without prior written and specific approval from the cognizant government agency responsible for radiation safety enforcement.

Venting

To avoid lab contamination with traces of tritium, it is recommended that the detector be vented via the outlet tubes connection to a fume hood or outside building air trough, a window or vent.

Registration is Required

Registration with the State Health Department is required in states which regulate use of radioactive material under agreement with the U.S. Nuclear Regulatory Commission and for your convenience has been accomplished by Valco Instruments Co., Inc. (a copy of this registration form has been provided to you for your information). As a general licensee you are required to comply with radiation control regulations in force in your state. A copy of the regulations currently in effect in Texas and the name, address, and telephone number of your local radiation control agency are appended.



YOUR LOCAL RADIATION CONTROL AGENCY IS:

4.2 Ionizing-Source Type

The purpose of the ionizing source is to produce ionization within the sample gas at a high and stable rate. Although gas ionization may readily be produced by electrical means, the far greater simplicity, stability, and reliability of radioactive sources have led to their use, almost exclusively, in electron-capture detectors. Furthermore, the desire for dense ionization in a small volume, with little or no penetrating radiation component, is best met by low-energy beta emitters; specifically, ^3H (tritium) and ^{63}Ni . Almost all of the EC detectors available use one of these isotopes.

4.2.1 Tritiated-Titanium Source. Tritium holds several major advantages over Nickel-63 as an ECD source material. Its betas have much lower energies (18 keV maximum compared with 63 keV), so that all the ionization is contained within 2 mm of the source-foil surface, and a very small detector volume is possible. Furthermore, tritium sources can be prepared with much higher specific activities than Nickel-63, leading to larger baseline currents and consequently greater sensitivity and wider dynamic range. Finally, the radiological health aspects of tritium are not nearly as severe as for Nickel-63, which tends to be retained indefinitely within the body, if ingested. The ECD source must be in solid form, however, so that tritium source technology is largely concerned with finding a mechanism for binding larger numbers of hydrogen atoms on or very near the surface of a solid. Until very recently, the best solution involved adsorbing the tritium gas onto a hot titanium-plated metal surface. This technique produces high-intensity (0.25 curies/cm²) sources which will maintain the tritium surface bond at temperatures up to approximately 220° C. Unfortunately, it is often desirable to operate GC detectors at higher temperatures than this, particularly for cleaning and bakeout, and tritiated titanium is relegated to those special applications where its restricted temperature limit is not important.

4.2.2 Nickel-63 Sources. Detector operation in excess of 220° C is the one advantage which made Nickel-63 the most widely used ECD source during the time when tritiated titanium was the only other reasonable alternative. Nickel is a metal which can be plated on other metals and maintain source integrity over the full temperature range required for ECD operation and bakeout. As described before, however, the higher-energy betas it produces require a larger detector volume than is desirable, and its low specific activity restricts source activities to a few tens of millicuries. Detectors which use Nickel-63, therefore, achieve high-temperature operation at the expense of baseline current, dynamic range, sensitivity, and response time. Coupled with the more severe restrictions on handling and distributing Nickel-63, these drawbacks have spurred the development of a higher-temperature tritium source foil.

4.2.3 Tritiated-Scandium Sources. Tritiated scandium combines all of the best features of tritium with a maximum temperature limit of 320° C, which is quite adequate for essentially all ECD applications. Source activities are available that approach one curie/cm², producing a higher specific activity and thus a higher baseline current than any other beta source available for ECD use. The Valco Model 140 Wide-Range Electron-Capture Detector employs a 1-curie tritiated-scandium source with a surface area of 1.6 cm², and achieves a baseline current of at least 2×10^{-8} amperes in a detector volume of only 180 microliters.

Valco Model-140 Wide Range Electron-Capture Detector

RECORD OF SOURCE TRANSFER

(License No. 8-2584G)

(Generally Licensed)

Date: 6-24-83

1 (Qty) Model 140B detector assemblies with the following serial numbers

Detector S/N 152

Controller S/N 112

and containing a total of 1000 millicuries of H3 Tritium were

shipped on 6-23-83 via UPS

to:

HNU Systems
160 Charlemont St.
Newton, MA 02161

The cognizant Radiation Control Agency for the recipient is:

Region 1, USNRC
Office of Inspection & Enforcement
631 Park Ave.
King of Prussia, PA 19406

Receipt acknowledged by:

Robert Z. Muggli
(Name)

Lab. Director
(Title/Position)


4 Oct 83
(Date)

Sign and send BLUE COPY to Radiation Control Agency (address given above)

Sign and return YELLOW COPY to: Valco Instruments Co., Inc.

Attn: Radiation Safety Officer
7812 Bobbitt Lane
Houston, Texas. 77055

Retain WHITE COPY for your information



VISI

Vaico Instruments Co. Inc.
Houston, Texas
Model 140B

WIDE RANGE ELECTRON CAPTURE DETECTOR
Serial No. 1

CAUTION - RADIOACTIVE MATERIAL

Incorporated within the housing is a detector cell which contains radioactive material (unpurified form of Hydrogen-3). This detector assembly should not be removed or disposed of except by transfer to a Federal agency, or to a State or local agency licensed by the U.S. Atomic Energy Commission. The receipt, possession, use and transfer of this electron capture detector assembly are subject to a general license or equivalent authorization of the U.S. AEC, or of an Agreement or State, if there is any indication of possible failure of or damage to the containment of the detector cell, its operation shall be suspended until any necessary repairs have been made. Disassembly, maintenance, or repair of the enclosed detector cell shall only be performed by Vaico Instruments Company, Incorporated or by a person holding a specific license to perform such service. Removal of the label is strictly prohibited by the U.S. AEC and all Agreement States. This label shall be replaced by a new one.



CONVERSATION RECORD

TIME

11:30 AM

DATE

5/21/85

TYPE

☐ VISIT

☐ CONFERENCE

☒ TELEPHONE

☐ INCOMING

☒ OUTGOING

ROUTING

NAME/SYMBOL

INT

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

Robert Z. Muggli, PhD

ORGANIZATION (Office, dept., bureau, etc.)

Walter C. McGone Asso. Inc.

TELEPHONE NO.

SUBJECT

Submitting ~~the~~ detector cell information

SUMMARY

Mr Muggli was told that he would have to submit detector cell cleaning information to the Commissioner.

He agreed to forward such information as needed.

10:15 11/08/85 Talked to Dr Muggli - He had no problem in authorizing them for 2 - Curie H³ foils.

ACTION REQUIRED

None

NAME OF PERSON DOCUMENTING CONVERSATION

JAMES W. PATTERSON

SIGNATURE

James W. Patterson

DATE

5/21/85

ACTION TAKEN

SIGNATURE

TITLE

DATE