

*file*

MAY 04 1984

Wright State University  
ATTN: Bruce T. Austin, Ph.D.  
Radiation Safety Officer  
7751 Colonel Glen Highway  
Dayton, OH 45435

License No. 34-11912-03  
Control No. 76674

SUBJECT: LICENSE RENEWAL APPLICATION

Gentlemen:

This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.

Any correspondence regarding the renewal application should reference the control number specified and your license number.

Sincerely,

Material Licensing Section  
Region III

8602110491 851203  
REQ LIC30  
34-11912-03 PDR

030-15160

April 26, 1984

U.S. Nuclear Regulatory Commission  
Materials Licensing Section  
799 Roosevelt Rd.  
Glen Ellyn, IL 60137

Madam or Sir,

This letter is submitted in application for renewal of USNRC License No. 34-11912-03, issued to the University.

The University wishes to continue operation under the provisions of this license and will continue to operate in accordance with the relevant documents, cited below, applicable USNRC regulations and License conditions.

Documents listed in License condition 20 that remain relevant to the current radiation safety and material control program are:

- Application dated August 3, 1978
- Letters dated March 5, 1981
  - October 12, 1982
  - December 1, 1982
  - March 4, 1983
  - April 22, 1983
  - October 7, 1983
- Radiation Safety Manual dated August 1983

Those superseded are:

- Radiation Safety Manual dated 1978
- Letter dated March 20, 1979
- Application dated August 31, 1979
- Letters dated September 17, 1980
  - April 7, 1982
  - July 6, 1982

<b>RECEIVED BY LFMB</b>	
Date	5/4/84
Log	May 4/84
By	cap
Orig. To	R/T
Action Cont.	cap

In addition to the above, the following changes have recently occurred and should be reflected in license documentation:

By Provost's memorandum dated April 3, 1984, Dr. Albert E. Langley, Associate Professor and Chairman of Pharmacology and Toxicology has been appointed as member and chairman of the University Radiation Safety Committee. This

170.1(a)(9) - state  
**FEE EXEMPT**

**RECEIVED**  
APR 30 1984  
REGION III

8508130287

Control No. 7667 <sup>APR 30 1984</sup>

results in partial revision of the letter dated April 22, 1983. A copy of the Provost's memorandum and Dr. Langley's credentials relevant to his appointment are enclosed (enclosures 1 and 2).

Due to replacement of radiation survey and assay equipment that is available for use for radiation safety purposes, item 10 of Appendix D to the application of 1978, last updated by letter dated March 4, 1983, should be updated by substitution of the enclosed listing (enclosure 3). No significant alteration of radiation survey and assay capability is reflected in the updated listing.

Reflecting the changing education and research mission of the University School of Medicine at the V.A. Medical Center, the lease designating University control of space at the center has been amended to add portions of V.A. Building 315 and to delete portions of V.A. Building 307 in which material has not been used. While material has not yet been used in these new areas, the University Radiation Safety Committee will continue to authorize material use in accord with policy previously elaborated for areas under University jurisdiction. Enclosure 4, (4 pages), is enclosed in description of areas under University control.

Enclosure 5 depicts a new laboratory complex for the radiation safety program that is nearing completion.

Besides a contained working environment, the facility will provide a centralized location for improved program efficiency and control of material receipt and distribution in support of the University Radiation Safety Program.

With the availability of this dedicated space, the J.L. Shepherd, model 28-6 calibration source is to be relocated for use and storage from the area specified in the letter dated March 4, 1983. Use and storage procedures, previously specified, remain unchanged. At a source to barrier distance of 28 feet or greater, the unattenuated beam intensity at the proximate surface of the barrier will not exceed 3.67 mR/hr. Typical use for calibration purposes involving attenuation for most procedures, a source use cycle of less than 50% and inherent barrier attenuation, will assure an exposure rate of less than 2 mR in any hour in uncontrolled areas adjacent to the laboratory.

Should any questions arise concerning this application, please do not hesitate to contact the undersigned at 513/873-2215.

Sincerely,



Bruce T. Austin, Ph.D., Consultant  
University Radiation Safety Officer

BTA/sdb

Enclosures

Control No. 76674

WRIGHT STATE UNIVERSITY  
OFFICE OF THE PROVOST

Provost's Memorandum

April 3, 1984

No. 84-1 (Supersedes No. 83-1)

SUBJECT: University Radiation Safety Committee

Nuclear Regulatory Commission guidelines require an official announcement of University Radiation Safety Committee membership. Current members of the Committee are:

Dr. Albert E. Langley, Associate Professor,  
Pharmacology and Toxicology (Chairperson)

Dr. George G. Hess, Associate Professor  
Chemistry

Dr. Lawrence J. Prochaska, Assistant Professor,  
Biological Chemistry

Dr. Adrian Rake, Associate Professor,  
Biological Sciences

Mr. David S. Atwater, Assistant Vice President,  
Facilities and General Services

Dr. Bruce T. Austin, University Radiation Safety  
Officer, (Ex-officio/Non-voting member)

The University Radiation Safety Committee will report to the Provost. This memorandum supersedes Provost's Memorandum 83-1, dated April 11, 1983.

Application for Renewal  
April 26, 1984

Control No. 76674

WRIGHT STATE UNIVERSITY  
RADIATION SAFETY COMMITTEE

Albert E. Langley, Ph.D.  
Associate Professor, Pharmacology and Toxicology

I. Education:

Ph.D. Ohio State University, 1974

II. Training:

Ohio State University  
Isotopic Tracers in Biology, 3 credit hrs

III. Experience:

<u>ISOTOPE</u>	<u>WHERE USED</u>	<u>DURATION</u>	<u>TYPE OF USE</u>
$^{14}\text{C}$	Ohio St. University	3 years	Enzyme Assay
$^{14}\text{C}$	Univ. of Colorado	2 years	Enzyme Assay
$^3\text{H}$	Univ. of Colorado	2 years	Tracer labeling
$^3\text{H}$	Warner-Lambert Co.	1 year	Tracer labeling
$^3\text{H}$	Wright State University	6 1/2 yrs	Tracer labeling
$^{125}\text{I}$	Wright State University	2 years	RIAs
$^{32}\text{P}$	Wright State University	1 year	Enzyme Assay

IV. Previous Radiation Safety Committee Experience:

None

Application for Renewal  
April 26, 1984

Control No. 76674

## APPENDIX D ITEM 10

<u>TYPE INSTRUMENT</u>	<u>RADIATION DETECTED</u>	<u>COUNTING RANGE</u>	<u>WINDOW THICKNESS</u>	<u>USE</u>
<u>SAFETY DEPT.</u>				
Eberline E120	Beta, Gamma	0-500 mR/Hr	1.4-2 mg/cm <sup>2</sup>	GM survey meter
Eberline PAC-4G-3	Alpha	0-500,000 cpm	.85 mg/cm <sup>2</sup>	Alpha survey meter
Ludlum 2200			7 mg/cm <sup>2</sup>	scintillation analyzer
Victoreen 470A	Beta, Gamma	0-500,000 cpm	.093 mg/cm <sup>2</sup>	
	x-ray	0-1000 mr/hr	w/o cap 17 mg/cm <sup>2</sup>	survey meter
Victoreen 541R	Beta, Gamma	0-1000 r/hr	w. cap 5000 mg/cm <sup>2</sup>	pocket dosimeter
Victoreen 855 (3)	Gamma, x-ray	0-200 mr		pocket dosimeter
Beckman LS9000	Gamma, x-ray	0-999 mr		liquid scintillation counter
	Beta	0-5.6x10 <sup>6</sup> cpm		
Victoreen 471	Alpha, Beta			
	Gamma, x-ray	0-3000 mr/hr	1.1 mg/cm <sup>2</sup>	survey meter
Victoreen 808D	Beta, Gamma	.1-1000 mr/hr	680 mg/cm <sup>2</sup>	area monitor
Victoreen 498 (2)	Beta, Gamma	0-1 r/hr	w/o cap	survey meter
			170 mg/cm <sup>2</sup>	
Siersat 08-400	none	n/a	n/a	air sampler
Nuclear-Chicago 2592, 2593, 2594	Beta, Gamma	0-1 r/hr	1 mg/cm <sup>2</sup>	cutiepie survey meter
Victoreen 570 r-meter	x-ray	0-250 roentgens	663-213 mg/cm <sup>2</sup>	precision ion chambers
	Beta, Gamma		705-67 mg/cm <sup>2</sup>	
			621-576 mg/cm <sup>2</sup>	
			651-6.1 mg/cm <sup>2</sup>	
			154-89 mg/cm <sup>2</sup>	
			130-212 mg/cm <sup>2</sup>	
Ludlum Model 3 (4)	Beta, Gamma	0-200 mR/hr		survey meter
D.S. Davidson	Beta, Gamma	0-1,000,000 counts		multi-channel analyzer
<u>FELS INSTITUTE</u>				
Nuclear-Chicago	Beta, Gamma	0-100,000 cpm		planchet counter
Packard 2420A (tri carb)	Beta	0-2,000,000 cpm		liquid scintillation counter
Packard 410 (tri carb) series 314F	Beta, Gamma	0-999,999 cpm		well counter
Victoreen 498	Beta, Gamma	0-1 r/hr	170 mg/cm <sup>2</sup>	survey meter
Packard 5110	Beta	0-999,999 cpm		liquid scintillation counter
<u>BIOLOGICAL SCIENCES</u>				
Packard 3255 (2)	Beta	0-999,999 cpm		liquid scintillation counter
Nuclear-Chicago 6770	Beta	0-499,999 cpm		liquid scintillation counter
Packard Autogamma	Gamma	0-3999,999 cpm		gamma counter
Packard B460CD	Beta	0-999,999 cpm		liquid scintillation counter

Application for Renewal  
April 26, 1984

Control No. 76674

## Appendix D Item 10 Cont.

<u>TYPE INSTRUMENT</u>	<u>RADIATION DETECTED</u>	<u>COUNTING RANGE</u>	<u>WINDOW THICKNESS</u>	<u>USE</u>
<u>COX HEART INSTITUTE</u>				
Searle Delta 300	Beta	0-999,999 cpm		liquid scintillation counter
Packard 800 CD	Beta, Gamma	0-999,999 cpm		Gamma counter
Eberline rm-14	Beta, Gamma	0-50,000 cpm	30 mg/cm <sup>2</sup>	radiation monitor
<u>PHYSICS</u>				
Xetec 501A-1	Beta, Gamma	.1-99.9 mr/hr	internal GM tube	range monitor
Victoreen 808D	Gamma	.1-1000 mr/hr	internal GM tube	area monitor
Nuclear-Chicago Labitron	Beta, Gamma	0-20,000 cpm		counter
Victoreen Fricker Model 495-5	Beta, Gamma	0-500 x 10 <sup>3</sup> cpm		survey meter
<u>MICROBIOLOGY &amp; INNUNOLOGY</u>				
Packard 3255	Beta	0-999,999 cpm		liquid scintillation counter
Abbott Auto Logic (2)	Gamma	0-999,999 cpm		well counter
Eberline E120	Beta, Gamma	0-500 mr/hr	30 mg/cm <sup>2</sup>	survey meter
<u>ENVIRONMENTAL HEALTH</u>				
Eberline E530	Beta, Gamma	0-200 mr/hr	30 mg/cm <sup>2</sup>	survey meter
<u>CENTRAL STORES</u>				
Eberline E120	Beta, Gamma	0-500 mr/hr	30 mg/cm <sup>2</sup>	survey meter
<u>CHEMISTRY</u>				
Nuclear Supplies SA-250	Beta, Gamma	0-999,999 cpm		GM counter
<u>BLDG. 307 V.A. CENTER</u>				
Beckman LS3011	Beta	0-999,999 cpm		liquid scintillation counter
Beckman Gamma Mate	Gamma	0-99,999 cpm		single well counter
<u>GEOLOGY</u>				
Victoreen 498 (2)	Beta, Gamma	0-1 r/hr	170 mg/cm <sup>2</sup>	survey meter
<u>PHARMACOLOGY</u>				
Packard Prias Model P1	Beta	0-9,999,999 cpm		liquid scintillation counter

Application for Renewal  
April 26, 1984



Control No. 76674





241	ROOM NUMBER
(275)	ROOM NO. F.I.

REVISION NO.	DATE	TITLE
		WRIGHT STATE UNIVERSITY
TITLE		
BASIC SCIENCE TEACHING ADDITIC		
GROUND FLOOR PLAN V.A. CENTER		
BLDG # U635 VA # 315		
DWG NO.	BY	APPROVED
1180	C.G.	
	SCALE	DATE
	1" = 20'-0"	7/7/82




SQUARE FOOT AREAS			
GROSS	NET	CIRCULAR	MICHAICAL
	VEHICLE	CIRCULA	
	AREAS	TION	
		TURAL	
19550	3014	4700	2520
	166	3192	UNDEVELOP
	106	6184	TOTAL
	264		2415
	19499		UNDEVELOP
	264		TOTAL
	19499		5335

Control No. 76674

Application for Renewal  
April 26, 1984

W.S.U.  
(S.I.A)

241	ROOMS MAP, 2nd
(275)	FDD 153, FT



EXISTING BUILDING




REVISION NO.	DATE	TITLE
		WRIGHT STATE UNIVERSITY
TITLE		
FIRST FLOOR		
EDUCATION BLDG. VA. CENTER U-27		
SCHOOL OF MEDICINE VA-507		
DWG NO.	BY	DATE
1670 K.P.		AUG 77

SQUARE FOOT AREAS	NET USABLE	CIRCULAR	CORRIDOR	CIRCULAR	STAIR	MECHANICAL	TOTAL	
							NUMBERED	UNNUMBERED
13,027	8,663	48	2668	815		NUMBERED 535	UNNUMBERED 27	
26,265	17,705	96	4750	2169		TOTAL 633	NUMBERED 153	UNNUMBERED 159

Application for Renewal  
April 26, 1984

241	REC'D MAY 17
(15)	ROOM 150, FT

REVISION NO.	1	12-7-78	DATE	REMOVE OR	145 FT
WRIGHT STATE UNIVERSITY					
TITLE					
GROUND FLOOR,					
EDUCATION BLDG., VA. CENTER U-					
SCHOOL OF MEDICINE					
DWG NO.	1670	BY	K.P.R.	SCALE	1/4" = 1'-0"
APPROVED		DATE	AUG. 77		




SQUARE FOOT AREAS			
GROSS	NET OCCUPYABLE	CURTAIN	MECHANICAL
3242	9042	1912	NUMERICAL TOTAL
66269	17705	4760	NUMERICAL TOTAL

Application for Renewal  
April 26, 1984

Control No. 76674

Radionuclide Laboratory  
Safety Department - Wright State



- ① Class III Fume Hood with HEPA Filter. Exhaust fan #2 - 1150 CFM.
- ② Bench Slot Local Exhaust System with HEPA Filter. Exhaust fan #3 - 900 CFM. This system is designed to provide slight negative pressure to Laboratory.