

Spaulding Rehabilitation Hospital

125 Nashua Street, Boston, Massachusetts 02114 Telephone: 617-720-6400

Dedicated to Patient Care, Teaching and Research

December 21, 1983

Material Licensing Branch Division of Fuel Cycle and Material Safety Region 1 U.S. Regulatory Commission 631 Park Avenue King of Prussia, PA 19406

RE: Amendment of Byproduct Material and License Number 20-20615-01

Dear Sir/Madame:

Enclosed please find two (2) copies of the amendment we wish to file on byproduct material license 20-20615-01. Our check for the \$40.00 amendment fee is also enclosed.

Please contact me if you have any questions regarding this amendment.

Sincerely.

GEORGE A. DEMERITT

leoge a VenerA

Vice President

GAD/m

Enc: as stated

Applicant

8602280419 851226 REG1 LIC30 20-20615-01 PDI PDR

OFFICIAL RECORD COPY"

Formerly called the Massachusetts Rehabilitation Hospital

A private specialized rehabilitation hospital associated with Harvard Medical School and Tufts University School of Medicine offering rehabilitation care in the areas of neurological disease, orthopedics, stroke, arthritis, pediatrics, pain, traumatic head injury, spinal cord disease, encology, alcoholism, cardiology, pulmonary disease and other rehabilitation areas. 02045

Massachusetts Rehabilitation Hospital Expires 7/31/88 License Number 20-20615-01 Please amend name of hospital to: Spaulding Rehabilitation Hospital Please amend item 10 to read: Survey meter calibration will be conducted annually by Robert Johnson, CHP (calibration procedures are on file with the NRC under license number 20-00297-53), or Nuclear Pharmacy, Incorporated (calibration procedures are on file with the NRC under license number 30-19916091), or Health Physics Services, Inc. of Potomac, Maryland (calibration procedures are on filr with the NRC under license number 19-19791-01). Please amend item 10, section 2 to read: Annually, the dose calibrator will be checked for accuracy using several radionuclides such as Cs-137, Co-57, and Ba-133 using appropriate reference standards whose activity is traceable to NBS. Procedure: assay reference standard in the dose calibrator at the appropriate 1) setting and subtract background to obtain a net activity 2) repeat step 1 for a total of 3 determinations and average results 3) the average activity determined in step 2 should agree with the certified activity of the reference source within ± 5% after decay corrections 4) Repeat above procedures for other commonly used radionuclides for which adequate reference standards are available. 5) All calibration accuracy checks will be maintained on file. 6) Calibration checks which do not agree within ± 5% indicate that the instrument should be repaired, recalibrated or adjusted. If this is not possible, a calibration factor should be calculated for use during routine assays of radionuclides. Please amend item 10, number 3 to read: Linearity of the dose calibrator will be determined quarterly, in accordance with the NRC Medical Licensing Guide, Appendix D, Section 2E or by utilizing Calcorp, Inc. calcheck test kit (data attatched), over the full range of technetium activities used.

(item 10, number 3 cont.)

Should the linearity (measured versus calculated) vary by greater than t 5%, appropriate corrective action will be conducted.

Item 18:


In addition to waste disposal procedures listed, we would like to include waste disposal/removal services provided by Nuclear Pharmacy, Incorporated (NPI procedures on file under NRC license number 20-21227-01 MD).

Item 11:

Please note a change in the sink location from the original application. (diagram of facility attached)

FACILITY DIAGRAM
Massachusetts Rehabilitation Hospital
Nuclear Cardiology Department

(OUTSIDE WALL)



(CORRIDOR)

- HOTE: 1. Long and short lived isotope storage, surrounded by 2 inch thick lead bricks, approximately 10"x12"x12".
 - Undercounter radioactive waste and decay-in-storage, 1/8" thick lead, 10"x14"x12".
 - 3. Sink
 - 4. Dose calibrator

Nowawith internew year valoped in activity internew years and prairie activity internew years and prairie activity internew years and prairie activity internew years are produced and prod

Toler Trest

Just four minutes

As simple 41 1 7 3 4 5 6 7 Placered central tube in duse Canterator Prace source in red tube and rake a reading Then sequent, any praise coice curted fuces diserred fuce Addtone resources are lange owned are a consering with a Biran I active and you can see the dedice of the still set until

York Cauchack to be completely a reusable to deriving a finite color of control of the color of

Safe to a contract of a contra


点

. 33

LAN

Carcinose de la companya de la compa





Product Description

Calicheck is a kit designed to perform the activity linearity test on a dose calibrator quickly and accurately. The kit consists of seven tubes, six of which are lead-lined to attenuate gamma radiation from radioactive sources, and a seventh, unlined tube. Each lead-lined tube varies in the thickness of lead so as to simulate various stages of radioactive decay. These tubes are sequentially placed over a source of radioactivity in the dose calibrator and, within minutes, seven successive measurements are acquired representing values that would have been obtained at approximately 0, 6, 12, 20, 30, 40 and 50 hours after the initial assay of To 99m. The need for determining linearity by fractionating eluants, or decaying the elution for several days while data is being collected, is eliminated -- and at greatly reduced radiation exposures to personnel.

