

MAR 04 1986

Mercy Medical Center
Department of Radiology
ATTN: Mr. Paul Halverson
4050 Coon Rapids Boulevard
Coon Rapids, MN 55433

Gentlemen:

Enclosed is Amendment Number 07 to your NRC License Number 22-17307-01 which authorizes the changes you requested in your letter dated February 12, 1986.

Please note that we have not authorized the use of the Minnesota Department of Public Safety (License No. 22-06714-02) as a survey instrument calibration service. This licensee is only authorized to perform survey instrument calibrations for individuals listed as authorized users under Condition 12. of their license.

If they wish to pursue providing a calibration service to others, they will need to submit an application in accordance with the enclosed guide.

If you have any questions or require clarification on any of the information stated above, you may contact us at (312) 790-5625.

Sincerely,

Original Signed By
William J. Adam, Ph.D.
Materials Licensing Section

Enclosures:

1. Amendment No. 07
2. Application Guide for Survey
Instrument Calibration Service

8604170033 860304
REQ3 LIC30
22-17307-01 PDR

RIII

WJA

Adam/cm

3/2/86