



RADIATION PHYSICS SERVICES

3300 Gallows Rd., Patts Church, Vo. 22046 pt. 698-5394

Mr. William E. Cline, Chief Nuclear Materials Safety and Safeguards Branch Division of Radiation Safety and Safeguards U.S. NRC, Region 11 101 Marietta St. N.W. Atlanta, GA 30323 Docket No. 030-20648 License No. 45-17128-03

Dear Mr. Cline

In response to your letter of Aug. 9, 1988, we offer the following additional information relative to the source change which we performed on our Cobalt 60 teletherapy unit on Jan. 15, 1988 (reference our previous submissions dated Jan. 18, 1988 and May 24, 1988).

- As we discussed on the phone two days ago, I am the Radiation Safety Officer for the Fairfax Hospital; therefore my signature of the submitted reports covers the requirement 10 CFR 35.641(c).
- •Enclosed is a revised version of the facility survey which we submitted with the May 24, 1988 material. This revised table includes calculations of the maximum weekly radiation levels which may occur for this facility as required by 10 CFR 35.641(c)
- Again, as we discussed by phone, we check the viewing system for proper operation daily before any treatments are given. Thus we are meeting the requirements of 10 CFR 35.634(d)(4) relative to spot-checks.

I belive this answers the questions which you had raised, but if you need more information please let me know.

Sincerefy:

James A. Deve. Ph.D.

RSO, Fairfax Hospital

OFFICIAL COPY

45-17128-03 PNU

Area Radiation Survey of CGR ALCYON II TELETHERAPY FACILITY The Fairfax Hospital

Source: Co-60, 6507 Ci on Jan. 15, 1988, Mod. no. COT-20 sn 3415

Phantom: 25 cm cube, polystyrene Source to phantom distance: 67.5 cm Field Size (maximum): 32 x 32 cm

Workload: estimated maximum = 80,000 R/wk at Isocenter or approx. 5.7 hrs/wk beam "on" time.

Phantom	Gantry Anale	Location	(un) restricted	Reading(mR/hr)	<u>U</u>	I.	mR/wk
N	270*	A. dressing rm	un	2.8	.25	.25	0.34
N	90.	C. window (max)	79	1.0	.25	1	1.68
N N	90	C. door (max)	re	0.6	.25	1	1.00
N	90,	C. wall (max)	79	<0.2	1	1	<1.3
N	180*	E. outpt. entrance (max)	un	0.2	25	25	0.08
. Y	270	A. dressing rm (max)	un	0.5	1	.25	0.84
	270*	8 control panel	re	< 0.2	1	1	×1.3
. Y	270	C. window	re .	0.4	1	1	2.7
¥	270"	C door	re	0.3	1	1	2.0
Y	90	C. window (max)	re .	3.0	.25	1	5.0max
, , , , , ,	901	C. door (max)	re	2.0	25	1	3.4max
Y	90"	C. wall (max)	re	<0.2	1	1	11.3

Notes:

- 1.0° is toward the floor
- 2.270° is toward the south wall (i.e. dressing rm. wall)
- 3. All measurements made with the Victoreen 470A Survey Meter (sn 1439)
- 4. See Fig 3 and 4 (page 7 and 7a) for location of measurements
- 5. Locations A, B and C are controlled areas and location E is uncontrolled
- 6. U is the use-factor and T is the occupancy-factor.
- 7 The last column contains the calculated estimates of viewby exposures to each area

Conclusion

Based on these measurements, radiation levels in all areas are less than 3 mR/hr. Hence it is not likely that radiation dose levels will exceed those limits set by 10 CFR Part 20.

Dege 9-17-88

NRC FORM 218 (4-76) NRCM 0240 TELEPHONE OR V	TIME 8:40 8 A.M. P.M.				
☐ INCOMING CALL	OUTGOING CALL	□ VISIT			
PERSON CALLING	OFFICE/ADDRESS	PHONE NUMBER EXTENSION			
	RT				
IS Klive	The state of the s	404-331-5084			
PERSON CALLED	OFFICE/ADDRESS	PHONE NUMBER EXTENSION			
JAMES A. Deye, Pl	.D. RSO - The tolkfox Hospital	703-698-3705			
3	CONVERSATION				
SUBJECT NOW Tolox	herry Source Chryse				
SUMMARY	and a source armede				
SERTO CALIBO	Dege New PART 35 which , which sent to NRC POD	nequesto he requesto output ono images requires requires reprocial reseas 4-7-88. Left up call ma.			
ACTION REQUESTED	ADVISE ME OF ACTION TAKEN.				
Deve to send	(A))T/A) C				
Regular feet	INITIALS				
a the Oak E co	DATE				
Called A-3-1	4				
ACTION TAKEN LO GO GO	appropriate copies of	INITIALS			

NRC FORM 218 (4 76)

FARFAY HOSPIAL