

Docket Nos. 030-02489
030-00357

10 NOV 1988

License Nos. 29-06750-01
29-06750-02

The Medical Center at Princeton
ATTN: Anthony Morlino
Assistant Vice President
Professional Services
253 Witherspoon Street
Princeton, New Jersey 08540

Gentlemen:

Subject: Combined Inspection Nos. 030-02489/88-001 and 030-00357/88-001

This refers to your letter dated September 12, 1988, in response to our letter dated August 8, 1988.

Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

Please note for clarification, Item C.a. of the Notice of Violation refers to Item C of Appendix D, Section 2 of Regulatory Guide 10.8 (Rev. 1) Item C details a procedure for performing the dose calibrator constancy test and therefore this item of noncompliance refers to the constancy test and not the accuracy check.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By
John E. Glenn, Ph.D.

John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section A
Division of Radiation Safety
and Safeguards

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
State of New Jersey

8811220198 881110
REG1 LIC30
29-06750-01 PDC

OFFICIAL RECORD COPY

RL MED CTR - 0001.0.0
10/25/88

RETURN ORIGINAL TO
REGION I

IE:07

bcc:
Region I Docket Room (w/concurrences)

RI:DRSS
Joustra/bc
10/26/88

RI:DRSS
Glenn
10/26/88

OFFICIAL RECORD COPY

RL MED CTR - 0002.0.0
10/25/88