American Science and Engineering, Inc. Fort Washington Cambridge, Massachusetts 02139

617 868-1600 Telox 921-458

27 October 1988

ISDR-88-192

US Nuclear Regulatory Commission Region I 475 Allendale Road King of Prussia, PA 19406

RE: Docket #030-04667 License #20-07288-01

## Gentlemen:



In response to your letter dated 29 September 1988 to Mr. Clayton Annis concerning the above referenced notice of violation following your inspection of our facilities conducted on May 25 & 26, 1988, please be advised:

- A. The inventory record that is currently maintained is the record associated with the routine leak testing, and does indicate the location of the source (see attached). However, we agree that this may not be a complete inventory in that it only includes those sources requiring leak testing - sources below the lower leak test requirement limits would not necessarily be inventoried by that system. We therefore agree to institute a six-month inventory of all sealed sources, logging the location of each source at that time.
- B. The radiation survey program has been formalized as of the date of the inspection to include complete records of the survey results in the radiation storage area.

We agree with the inspector's understanding of changes to be made in the program as outlined in your 9/29/88 letter and are preparing an amendment request incorporating these changes into the program.

The above corrective steps are in place at this time and assure ongoing compliance in the future. Full compliance is therefore achieved as of this date. Please don't hesitate to contact us if there are further questions.

Yours truly,

Kluone

F.X. Masse Radiation Safety Officer

8812090274 881207 REG1 LIC30 20-07288-01 PDC CFFICE: 61Z-245-6600 617-283-4888

## F. X. Masse Associates, Inc.

Health Physics Consultants Maple St., P. O. Box 95 Middleton, Mass. 01949

AMERICAN SCIENCE & ENGINEERING Quarterly Wipe-test Report 24 September 1987

Inventoried by: \_\_\_\_G\_\_\_Eallon\_\_\_\_

Source	ldentification Number	Removeble Activity (uCi)	Location
244 Co	394-A 394-B	< 0.005 ( 0.005	Storage Room Storage Room
244 CA 244	400	< 0.005	Storage Room
C.e 244	405-A 405-B	< 0.005 < 0.005	Storage Room Storage Room
CA	435	0,005	Storage Rock

(24 hrs.) 617-283-4888

