

NOTICE OF VIOLATION

City of Kansas City

License No. 24-15241-01

As a result of the inspection conducted on April 25, 1986, and in accordance with 10 CFR Part 2, Appendix C - General Statement of Policy and Procedure for NRC Enforcement Actions (1985), the following violations were identified:

1. License Condition 19 requires that all licensed material be possessed and used in accordance with statements, representations and procedures contained in application dated December 29, 1982.

Application dated December 29, 1982, states that John L. Reddy is the Radiation Protection Officer.

Contrary to the above, an individual other than John L. Reddy was the Radiation Protection Officer. Specifically, John L. Reddy retired in December 1983, and the duties of Radiation Protection Officer were assumed by Stephen Shipp without amending the license.

This is a Severity Level IV violation (Supplement VI)

2. License Condition 13A requires that each sealed source containing licensed material shall be tested for leakage and/or contamination at intervals not to exceed six months.


Contrary to the above, each sealed source containing licensed material was not tested for leakage and/or contamination every 6 months. Specifically, sealed sources serial numbers 3413 and 3414 were not tested for leakage in 1984, 1985, or 1986. The last leak test was November 11, 1983. Sealed sources serial numbers 474897, 405753, 474898, and 405754, were not tested for leakage in 1984 or 1985. The last test for leakage was April 25, 1986.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated

5/13/86

  
D. J. Sreniawski, Chief  
Nuclear Materials Safety  
Section 2