

ENCLOSURE 1

NOTICE OF VIOLATION

Met Lab, Inc.

License No. 45-09663-01

The following violations were identified during an inspection conducted on June 7 and June 27, 1984. The Severity Levels were assigned in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C).

1. 10 CFR 34.24 requires a licensee to calibrate each radiation survey meter, used to conduct physical radiation surveys, at intervals not to exceed three months.

Contrary to the above, the licensee did not calibrate radiation survey meter No. 5855, a meter used to conduct physical radiation surveys, between June 1983 and March 1984, an interval in excess of three months. Radiation survey meter No. 5854, also was not calibrated between March 1983 and December 1983.

This is a Severity Level IV violation (Supplement VI).

2. 10 CFR 34.33(c) requires a licensee to check pocket dosimeters, at least once each year, for correct response to radiation.

Contrary to the above, the licensee did not check pocket dosimeters, worn by individuals conducting radiographic operations between February 1983 and June 1984, for correct response to radiation.

This is a Severity Level V violation (Supplement VI).

Pursuant to 10 CFR 2.201, you are required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved.

Security or safeguards information should be submitted as an enclosure to facilitate withholding it from public disclosure as required by 10 CFR 2.790(d) or 10 CFR 73.21.

Date: 1111 05 1984

8412270465

XA