

NOTICE OF VIOLATION

Mercy Hospital and Medical Center

License No. 12-01257-04

As a result of the inspection conducted on May 7 and 8, 1986, and in accordance with 10 CFR Part 2, Appendix C - General Statement of Policy and Procedure for NRC Enforcement Actions (1985), the following violation was identified:

License Item No. 7.G requires that byproduct material authorized by Item No. 6.G are to be acquired in the form of prepackaged kits.


Contrary to the above, licensed material has not been acquired and used in the form of prepackaged kits. Specifically on nine occasions between February 1984 and April 1986, the licensee has obtained and used iodine -125 in a form other than kits.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated

6/11/86


D. J. Sreniawski, Chief
Nuclear Materials Safety
Section 2

8606180189 860612
REG3 LIC30
12-01257-04 PDR