



RECEIVED

'85 OCT -3 P12:04

*The Christ Hospital*U.S. N.R.C.
LIC. FEE MGMT. BRANCH

5129 Andrew Avenue | Cincinnati, Ohio 45210 | (513) 369-8600

September 17, 1985

Mr. John Madera
U.S.N.R.C. Region III
Materials Licensing Branch
799 Roosevelt Rd.
Glean Ellyn, IL 60137

RE: License # 34-03831-02
Control # 78680

Dear Mr. Madera:

Our license is currently being reviewed for renewal. We wish to add the following information for Item 6.b.; Radioactive Materials for Item 6.a.

Element and Mass #	Chemical Form	Manufacturer Mode	Amount
Iodine 125	ion exchange	AECL C 32	300mCi
Gadolinium 153	GdO ₂	DuPont NEN NER-430	1500mCi

The purpose of use for these sources is the dual photon Bone Mineral Analyzer.
The Gd-153 source manufacturer's NRC certificate number is NR-476-5-153-S.

Thank you for your assistance in this matter.

Oct 3 - 1985

Check ☒ [initials]
 Amount ☒ [initials]
 Date ☒ [initials]
 Date Check Rec'd ☒ [initials]
 Received By: Jacques Cook

Sincerely,

[Signature]
Jack M. Cook
President

RECEIVED
SEP 27 1985
REGION III

JMC/am1

8604090390 851127
REG3 LIC30
34-03831-02 PDR

SEP 27 1985

CONTROL NO. 79842

vm

NOTE TO: License Fee Management Branch, ADM

FROM: Region III

SUBJECT: VOIDED APPLICATION

Control Number 79842

Applicant The Christ Hospital

Date Voided Jan 30, 1986

Reason for Void Withdrawn by
licensee

Signature Chris Murphy

Attachment:
Application

RECEIVED FEB 3 1986

7 Oct 32
no fee
collection

OIC LFMS

CHECKED FEB 3 1986