

RAHWAY RADIOLOGY SERVICES, P.A.
2107 ST. GEORGES AVENUE
RAHWAY, N. J. 07065
—
TELEPHONE (201) 388-1050

DIAGNOSTIC RADIOLOGY
SUPERFICIAL RADIOTHERAPY

SUPER VOLTAGE RADIATION THERAPY
STRONTIUM-90 BETA THERAPY

June 2, 1986

License # 29-18385-01
Docket # 30-14979
Inspection # 030-14979/86-01

John E. Glenn, Chief
Nuclear Materials Safety Section B
Division of Radiation Safety Safeguards
US Nuclear Regulatory Commission Region I
631 Park Avenue
King of Prussia, Pennsylvania 19406

Dear Sir:

In response to your communication of May 13, 1986, please be informed that the cobalt-60 teletherapy room is now a totally restricted area as defined in 10 CFR 20.3 (a)(14) access therinto is totally controlled by means of a physical lock installed on the teletherapy room access door. Therefore, access is now totally controlled totally by appropriately licensed personel.

In addition, cleaning of the teletherapy room is carried out under direct control of duly licensed personel during operational hours only. Access is otherwise completely barred.

Trusting the above is satisfactory and thereby meets full compliance, I remain,

Sincerely yours,


T. E. SHARETT, M. D., Director of Services

TES/rs

c.c. to Michael Caprio, Jr. Radiation Protection Services,
P. O. Box 2395
Darian, Conn. 06820

8607140114 860709
REG1 LIC30
29-18385-01 PDR

RECEIVED-REGION I
JUN -6 PM 2:30