

## Springfield Imaging Center, Ltd.

C.T. • MAMMOGRAPHY  
NUCLEAR MEDICINE • ULTRASOUND

319 E. Madison Street  
Springfield, Illinois 62701  
(217) 753-0235

June 23, 1986

D.J. Sreniawski, Chief  
Nuclear Materials Safety Section II  
U.S. Nuclear Regulatory Commission  
799 Roosevelt Road  
Glen Ellyn, Illinois 60137

reference: 12-24654-01

Dear Mr. Sreniawski:

We are in receipt of your letter dated June 4, 1986, and submit the following information regarding the inspection conducted at our facility on May 6, 1986.


Regarding the Cesium 137 source utilized for the daily function checks of the dose calibrator, we will be filing an amendment request within the next thirty to forty-five days to request that the information be changed in our application regarding the activity of the source used. The amendment request will ask to show a change in the source activity from 200 microcuries to a source of 50 microcuries or greater in activity.

The reason for the thirty to forty-five day delay is we will also be requesting some other changes to our license in this amendment, and we are currently awaiting preceptor statement information for a new individual to add to the license.

We realize that the item of non-compliance will not be considered corrected until the actual amendment is issued. We certainly appreciate your cooperation in this matter.

Should you have any questions or require any additional information in this matter, please contact Mr. Ronald D. Edwards, physicist, Radiation Protection Consultants, Ltd., 4255 Westbrook Drive, Aurora, Illinois 60505, 312-851-5986. He is authorized to answer any questions or provide any additional information you may require regarding our license.

Very truly yours,

  
Nicola Chiaradonna, M.D.  
Radiation Safety Officer

JUN 30 1986

8607140256 860708  
REG3 LIC30  
12-24654-01 PDR