

Enviromental Chem-Test
590 South Street
Raynham, MA 02767

MS=16
K 8

February 21, 1986


United Nuclear Registry
Region 1
631 Park Ave.
King of Prussia, PA 19406

Attention: Lester Tripp

In regards to the wipe-test, it is done every six (6) months under the supervision of Peter Kopcyh, using a kit supplied by Nuclear Sources & Services Inc., from Huston, Texas. The kit is LT-1 or LT-2.

Sincerely,

Peter Kopcyh
Enviromental Chem-Test



8605010361 860402
REG1 LIC30
20-19650-01 PDR

RECEIVED-REGION 1
FEB 25 PM 12:50

184926

FEB 25 1986

"OFFICIAL RECORD COPY"

ML10

2/20/86

4:00

TELEPHONE OR VERBAL CONVERSATION RECORD

☒ INCOMING CALL☐ OUTGOING CALL☐ VISIT

PERSON CALLING

Kopych

OFFICE/ADDRESS

Environmental Chem Test

PHONE NUMBER

617 822 1151

EXTENSION

PERSON CALLED

MTripps

OFFICE/ADDRESS

R 1

PHONE NUMBER

5396

EXTENSION

CONVERSATION

SUBJECT

Telephone Deficiency

SUMMARY

- Kopych to provide documentation of the following
- 1- leak test kit (LT1 or 2, Nuclear Sources & Services, Hopedale)
 - 2- leak test (at 6 mos. intervals)
 - 3- GC used under supervision of (Kopych)

"OFFICIAL RECORD COPY"

ML10

REFERRED TO:

ACTION REQUESTED

ACTION TAKEN

☐ ADVISE ME OF
ACTION TAKEN.

INITIALS

JH

DATE

2/20/86

INITIALS

DATE