



Georgia-Pacific Corporation

P.O. Box 367
Jarratt, Virginia 23867
Telephone (804) 535-8541

06 JUL 28 P 4: 00

July 23, 1986

Nuclear Regulatory Commission
Region II
101 Marietta Street, N.W.
Atlanta, Georgia 30323

SUBJECT: Report NO. 45-18185-01/86-01
Docket No. 030-14787
License No. 45-18185-01


Gentlemen:

This refers to the Nuclear Regulatory Commission (NRC) inspection conducted by C. A. Rowe on June 5, 1986 and our answers to the notice of violation.

1. In reviewing our records we found that the three cesium-137 sources used in our Ohmart gauges had not been tested as required by our licence condition 13. These sources have been wipe tested and sent for testing to Ohmart, covered by our purchase order No. 19026. We should have the test results in our office by August 15, 1986. We have posted a reminder to prevent this from happening again.
2. A notice has been posted at the operator's station stating that the license, all documents, operating procedures and any other information needed is located in the Engineering Office and may be examined at any time. We are getting a form NRC-3 "Notice to Employees" and will post it in the proper locations. This will be completed by Sept. 1, 1986.

If additional information is needed, please call me.

Very truly yours,


James E. Phelps
Plant Engineer

JEP/ds

8608110462 860723
REQ2 LIC30
45-18185-01 PDR

10
IE 07