

NOV 17 1986

Bedford Medical Center  
Radiology Department  
ATTN: Donald D. Donner, M.D.  
Radiation Safety Officer  
2900 West 16th Street  
Bedford, IN 47421

Gentlemen:

Enclosed is Amendment No. 17 to your NRC License No. 13-14823-01  
in accordance with your request.

Please note that Item 10, Page 9. of your application dated May 17, 1982  
indicates that survey instruments will be calibrated at least annually and  
following repair.

Please be aware that we have not authorized you to change the frequency of dose  
calibrator accuracy, constancy, linearity, or geometrical variation tests.  
Your application dated May 17, 1982 indicates that you will follow Section 2  
of Appendix D, Regulatory Guide 10.8, October 1980. Therefore, you are  
required to conduct daily constancy, annual accuracy, and quarterly linearity  
dose calibrator tests.

Please review the enclosed document carefully and be sure that you understand  
all conditions. You must conduct your program involving radioactive materials  
in accordance with the conditions of your NRC license, representations made in  
your license application, and NRC regulations. In particular, note that you  
must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices,  
Instructions and Reports to Workers; Inspections," 10 CFR Part 20,  
"Standards for Protection Against Radiation," and other applicable  
regulations.
2. Possess radioactive material only in the quantity and form indicated in  
your license.
3. Use radioactive material only for the purpose(s) indicated in your  
license.
4. Notify NRC in writing of any change in mailing address.

8703300406 861117  
REG3 LIC30  
13-14823-01 PDR

5. Request and obtain appropriate amendment if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. Any amendment request should be accompanied by the appropriate fee specified in 10 CFR Part 170.
6. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.
7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material prior to your expiration date.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions and representations in your license application will result in enforcement action against you in accordance with the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By  
Kathy J. Graden  
Materials Licensing Section

Enclosure: Amendment No. 17

RIII  
*KJ*  
Graden/pd  
11/20/86