



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
531 PARK AVENUE
KING OF PRUSSIA, PENNSYLVANIA 19406

License No. 08-13619-01
Docket No. 030-28719
Control No. 19011

Department of the Army
U.S. Army Engineers-Baltimore
District
ATTN: Randolph H. Erhart
Chief, Safety and Occupational
Health Office
Washington Aqueduct Division
5900 MacArthur Blvd., N.W.
Washington, DC 20315


Gentlemen:

This is in reference to your letter received May 30, 1985, for renewal of License No. 08-13619-01.

The NRC lists all approved gas chromatograph radioactive sources by detector cell model numbers. It appears the T-0145 Varian number you submitted is a serial number for one of their Ni-63 detector cells. Please submit the model number of the detector cell that contains the 8 millicuries of Ni-63 foil in the Varian gas chromatograph.

We will continue our review of your application upon receipt of the above information. Please reply in duplicate, referencing Control No. 19011.

Sincerely,


John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section B
Division of Radiation Safety and
Safeguards

8511130138 851022
REG1 LIC30
08-13619-02 PDR



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
631 PARK AVENUE
KING OF PRUSSIA, PENNSYLVANIA 19406

AUG 12 1985

License No. 08-13619-01
Docket No. 030-28719
Control No. 19011

Department of the Army
U.S. Army Engineers-Baltimore
District
ATTN: Randolph H. Erhart
Chief, Safety and Occupational
Health Office
Washington Aqueduct Division
5900 MacArthur Blvd., N.W.
Washington, DC 20315

Gentlemen:

This is in reference to your letter received May 30, 1985, for renewal of License No. 08-13619-01.

The NRC lists all approved gas chromatograph radioactive sources by detector cell model numbers. It appears the T-0145 Varian number you submitted is a serial number for one of their Ni-63 detector cells. Please submit the model number of the detector cell that contains the 8 millicuries of Ni-63 foil in the Varian gas chromatograph.

We will continue our review of your application upon receipt of the above information. Please reply in duplicate, referencing Control No. 19011.

Sincerely,

Original Signed By:
Jenny M. Johansen

John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section B
Division of Radiation Safety and
Safeguards

RI:DRSS
Varela
8/9/85

RI:DRSS
Glenn

8/12/85
OFFICIAL RECORD COPY

ML18