

NOV 14 1986

License No. 20-13682-01
Docket No. 030-01969
Control No. 105259

North Adams Regional Hospital
ATTN: Rutledge W. Currie, MD.
Radiation Safety Officer
Hospital Avenue
North Adams, Massachusetts 01247

Gentlemen:

This is in reference to your application dated March 24, 1986 to renew License No. 20-13682-01. In order to continue our review, we need the following additional information:

1. Ancillary personnel (clerical, nursing, housekeeping security, etc.) whose duties may require them to work in the vicinity of radioactive material (whether escorted or not) need to be informed about radiation hazards and appropriate precautions. Outline your method to assure that these employees receive the necessary instruction. Confirm that this instruction will be given both initially and annually thereafter on a refresher basis.
2. Please submit the written instructions that you will provide to personnel who receive shipments of radioactive materials during hours when the Nuclear Medicine is closed. Your instructions should be at least equivalent to those specified in Appendix E of Regulatory Guide 10.8.

We will continue our review upon receipt of this information. Please reply in duplicate to my attention at the Region I office and refer to Mail Control No. 105259 renewal.

In order to continue prompt review of your application, we request that you submit your response to this letter within 30 calendar days from the date of this letter.

Sincerely,
Original Signed By:
John E. Glenn
Jenny M. Johansen, M.S.
Nuclear Materials Safety Section B
Division of Radiation Safety
and Safeguards

Enclosures:

1. 10 CFR Parts 19, 20, 30, 35 and 71
2. Regulatory Guides 10.8

8704010507 861224
REG1 LIC30
20-13682-01 PDR

OFFICIAL RECORD COPY

ML 798 JOHANSEN - 0001.0.0
11/10/86

ML10

NOV 14 1986

RI:DRSS
Johansen/pj
11/13/86

OFFICIAL RECORD COPY

ML 798 JOHANSEN - 0002.0.0
11/10/86