

ATVI
A-TECH/VIRGINIA, INC.

RSA

Roofing Systems Associates, Inc.

ROOFING AND ROOF DECK CONSULTANTS

MOISTURE SURVEYS/THERMAL ANALYSES/SPECIFICATION SUPERVISION/MAINTENANCE PROGRAMS/REPAIR RECOMMENDATIONS/INSPECTION AND TESTING

Box 2411 • Roanoke, Virginia 24010

(703) 982-5607

March 26, 1987

Mr. J. Philip Stohr, Director
Division of Radiation Safety & Safeguards
United States Nuclear Regulatory Commission
Region II
101 Marietta Street, NW
Atlanta, Georgia 30323

Re: Your Letter of 3-18-87
Notice of Violation
Docket No. 030-19267
License No. 45-19806-01

97 MAR 30 AIO: 31

Dear Mr. Stohr:

Pursuant to your "Notice of Violation", the following in sequence, is our response:

A. This is the only violation of the five that we deny.

In our letter to you dated March 3, 1987, we addressed this:

Stephen P. Strauss is no longer employed by us and consequently our license requires an amendment.

Richard Fisher and George W. Gills, Jr., have been to the Troxler School. George W. Gills, Jr., should be added to the amendment.

In no case have we permitted a non-schooled individual to operate the gauge.

Gary Gills has been a member of our two man crew (son of George). His only function has been to participate in marking the roof and as a scribe in recording on our work sheets. In no case has he ever been on a job without one of the two of us present. Nevertheless, we do intend to have him attend the Troxler School in the future.

B. On March 23, 1987 we received from Landauer, the required film badge service on a quarterly basis.

C. On March 19, 1987, Troxler performed the latest leak test on our instrument. This will be done every six months as required.

8704130126 870326
REG2 LIC30
45-19806-01 PDR

OFFICIAL COPY

110 1E07

ATVI:
A-TECH/VIRGINIA, INC.

RSA

Roofing Systems Associates, Inc.

ROOFING AND ROOF DECK CONSULTANTS

MOISTURE SURVEYS/THERMAL ANALYSES/SPECIFICATION SUPERVISION/MAINTENANCE PROGRAMS/REPAIR RECOMMENDATIONS/INSPECTION AND TESTING

Box 2411 • Roanoke, Virginia 24010

(703) 982-5607


Mr. J. Philip Stohr, page 2

D. On March 12, 1987, we received from Troxler, the required certificate of hazardous material for shipping purposes. This and our current leak test will be carried at all times with the instrument.

E. Early in March, 1987, we received from you the documents required for posting, and the posting was accomplished in the office where the gauge is stored.

We regret and apologize for being remiss in these matters, and assure you of our intent to comply in the future.

Sincerely,



Richard H. Fisher

RHF:me